

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Field of measured quantity: **electrical quantities**

Calibration:

Nominal calibration temperature: (23 ± 2) °C

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
1	Direct-current voltage Measurement 0 mV to 20 mV 20 mV to 50 mV 50 mV to 100 mV 100 mV to 200 mV 200 mV to 500 mV 500 mV to 1 V 1 V to 2 V 2 V to 5 V 5 V to 10 V 10 V to 20 V 20 V to 50 V 50 V to 100 V 100 V to 200 V 200 V to 500 V 500 V to 1000 V		0.50 µV 0.0011 % 0.00068 % 0.00057 % 0.00052 % 0.00038 % 0.00034 % 0.00051 % 0.00038 % 0.00034 % 0.00064 % 0.00051 % 0.00047 % 0.00070 % 0.00055 %	TP1, TP21
	Direct-current voltage Generation 10 µV to 20 mV 20 mV to 50 mV 50 mV to 100 mV 100 mV to 200 mV 200 mV to 500 mV 500 mV to 1 V 1 V to 2 V 2 V to 5 V 5 V to 10 V 10 V to 20 V 20 V to 50 V 50 V to 100 V 100 V to 200 V 200 V to 500 V 500 V to 1000 V		0.9 µV 0.0039 % 0.0021 % 0.0015 % 0.0013 % 0.0009 % 0.00078 % 0.00073 % 0.00057 % 0.00052 % 0.0011 % 0.0009 % 0.0008 % 0.0013 % 0.0011 %	TP1, TP21

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
1*	Direct-current voltage Measurement 0 mV to 20 mV 20 mV to 50 mV 50 mV to 100 mV 100 mV to 200 mV 200 mV to 500 mV 500 mV to 2 V 2 V to 10 V 10 V to 20 V 20 V to 50 V 50 V to 100 V 100 V to 300 V 300 V to 600 V 600 V to 800 V 800 V to 1000 V		1 µV 0.0041 % 0.0025 % 0.0019 % 0.0015 % 0.0013 % 0.0011 % 0.0019 % 0.0016 % 0.0014 % 0.0016 % 0.0019 % 0.0022 % 0.0027 %	TP1, TP21
	Direct-current voltage Generation 0.1 mV to 100 mV 100 mV to 150 mV 150 mV to 200 mV 200 mV to 500 mV 500 mV to 1 V 1 V to 2 V 2 V to 5 V 5 V to 10 V 10 V to 20 V 20 V to 50 V 50 V to 100 V 100 V to 200 V 200 V to 500 V 500 V to 1000 V		4 µV 0.0040 % 0.0033 % 0.0026 % 0.0017 % 0.0014 % 0.0025 % 0.0015 % 0.0013 % 0.0030 % 0.0021 % 0.0019 % 0.0031 % 0.0023 %	TP1, TP21
	pH (electric method) (0 to 14) pH (-1500 to +1500) mV		0.01 pH 0.02 mV	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2	Alternating-current voltage Measurement 1 mV to 10 mV	10 Hz to 100 Hz 100 Hz to 305 Hz 305 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	6 µV 5 µV 4 µV 6 µV 12 µV 28 µV	TP2, TP21
	10 mV to 20 mV	10 Hz to 100 Hz 100 Hz to 305 Hz 305 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	8 µV 7 µV 6 µV 8 µV 15 µV 35 µV	
	20 mV to 50 mV	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 305 Hz 305 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.041 % 0.038 % 0.034 % 0.029 % 0.036 % 0.074 % 0.18 %	
	50 mV to 100 mV	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 305 Hz 305 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.027 % 0.025 % 0.024 % 0.021 % 0.023 % 0.049 % 0.12 %	
	100 mV to 200 mV	10 Hz to 40 Hz 40 Hz to 100 Hz	0.018 % 0.015 %	
	100 mV to 200 mV	100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.014 % 0.015 % 0.039 % 0.091 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2	200 mV to 500 mV	10 Hz to 40 Hz	0.022 %	TP2, TP21
		40 Hz to 100 Hz	0.019 %	
		100 Hz to 2 kHz	0.018 %	
		2 kHz to 10 kHz	0.019 %	
		10 kHz to 30 kHz	0.041 %	
		30 kHz to 100 kHz	0.16 %	
		100 kHz to 300 kHz	1.3 %	
	500 mV to 1 V	10 Hz to 40 Hz	0.016 %	
		40 Hz to 100 Hz	0.013 %	
		100 Hz to 2 kHz	0.012 %	
		2 kHz to 10 kHz	0.013 %	
		10 kHz to 30 kHz	0.029 %	
		30 kHz to 100 kHz	0.091 %	
		100 kHz to 300 kHz	0.70 %	
	1 V to 2 V	10 Hz to 40 Hz	0.013 %	
		40 Hz to 100 Hz	0.011 %	
		100 Hz to 2 kHz	0.0084 %	
		2 kHz to 10 kHz	0.011 %	
		10 kHz to 30 kHz	0.025 %	
		30 kHz to 100 kHz	0.071 %	
		100 kHz to 300 kHz	0.50 %	
	2 V to 5 V	300 kHz to 1 MHz	3.0 %	
		10 Hz to 40 Hz	0.021 %	
		40 Hz to 100 Hz	0.019 %	
		100 Hz to 2 kHz	0.017 %	
		2 kHz to 10 kHz	0.019 %	
		10 kHz to 30 kHz	0.041 %	
		30 kHz to 100 kHz	0.16 %	
	5 V to 10 V	100 kHz to 300 kHz	1.3 %	
		10 Hz to 40 Hz	0.015 %	
		40 Hz to 100 Hz	0.013 %	
		100 Hz to 2 kHz	0.011 %	
		2 kHz to 10 kHz	0.013 %	
		10 kHz to 30 kHz	0.029 %	
		30 kHz to 100 kHz	0.091 %	
		100 kHz to 300 kHz	0.70 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2	10 V to 20 V	10 Hz to 40 Hz	0.013 %	TP2, TP21
		40 Hz to 100 Hz	0.011 %	
		100 Hz to 2 kHz	0.0084 %	
		2 kHz to 10 kHz	0.011 %	
		10 kHz to 30 kHz	0.025 %	
		30 kHz to 100 kHz	0.071 %	
		100 kHz to 300 kHz	0.50 %	
		300 kHz to 1 MHz	3.0 %	
	20 V to 50 V	10 Hz to 40 Hz	0.022 %	
		40 Hz to 100 Hz	0.019 %	
		100 Hz to 2 kHz	0.017 %	
		2 kHz to 10 kHz	0.019 %	
		10 kHz to 30 kHz	0.041 %	
		30 kHz to 100 kHz	0.16 %	
	50 V to 100 V	10 Hz to 40 Hz	0.016 %	
		40 Hz to 100 Hz	0.013 %	
		100 Hz to 2 kHz	0.012 %	
		2 kHz to 10 kHz	0.013 %	
		10 kHz to 30 kHz	0.029 %	
		30 kHz to 100 kHz	0.091 %	
	100 V to 200 V	10 Hz to 40 Hz	0.014 %	
		40 Hz to 100 Hz	0.011 %	
		100 Hz to 2 kHz	0.0089 %	
		2 kHz to 10 kHz	0.011 %	
		10 kHz to 30 kHz	0.025 %	
		30 kHz to 100 kHz	0.071 %	
	200 V to 500 V	40 Hz to 10 kHz	0.020 %	
		10 kHz to 30 kHz	0.041 %	
	500 V to 1000 V	40 Hz to 10 kHz	0.014 %	
		10 kHz to 30 kHz	0.029 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2	Alternating current voltage Generation 1 mV to 10 mV	10 Hz to 100 Hz	0.85 %	TP2, TP21
		100 Hz to 2 kHz	0.78 %	
		2 kHz to 10 kHz	0.85 %	
		10 kHz to 30 kHz	1.4 %	
		30 kHz to 100 kHz	2.9 %	
	10 mV to 20 mV	1 Hz to 40 Hz	0.12 %	
		40 Hz to 100 Hz	0.11 %	
		100 Hz to 330 Hz	0.10 %	
		330 Hz to 2 kHz	0.092 %	
		2 kHz to 10 kHz	0.11 %	
		10 kHz to 30 kHz	0.17 %	
		30 kHz to 100 kHz	0.37 %	
		100 kHz to 300 kHz	0.76 %	
		300 kHz to 1 MHz	2.3 %	
	20 mV to 50 mV	1 Hz to 40 Hz	0.077 %	
		40 Hz to 100 Hz	0.072 %	
		100 Hz to 330 Hz	0.069 %	
		330 Hz to 2 kHz	0.066 %	
		2 kHz to 10 kHz	0.070 %	
		10 kHz to 30 kHz	0.096 %	
		30 kHz to 100 kHz	0.20 %	
		100 kHz to 300 kHz	0.55 %	
		300 kHz to 1 MHz	1.7 %	
	50 mV to 100 mV	1 Hz to 40 Hz	0.044 %	
		40 Hz to 100 Hz	0.039 %	
		100 Hz to 330 Hz	0.038 %	
		330 Hz to 2 kHz	0.035 %	
		2 kHz to 10 kHz	0.037 %	
		10 kHz to 30 kHz	0.057 %	
		30 kHz to 100 kHz	0.14 %	
		100 kHz to 300 kHz	0.44 %	
		300 kHz to 1 MHz	1.5 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2	100 mV to 200 mV	1 Hz to 40 Hz	0.031 %	TP2, TP21
		40 Hz to 330 Hz	0.024 %	
		330 Hz to 2 kHz	0.025 %	
		2 kHz to 10 kHz	0.023 %	
		10 kHz to 30 kHz	0.044 %	
		30 kHz to 100 kHz	0.11 %	
		100 kHz to 300 kHz	0.51 %	
		300 kHz to 1 MHz	1.5 %	
	200 mV to 500 mV	10 Hz to 40 Hz	0.036 %	
		40 Hz to 2 kHz	0.026 %	
		2 kHz to 10 kHz	0.022 %	
		10 kHz to 30 kHz	0.043 %	
		30 kHz to 100 kHz	0.17 %	
		100 kHz to 300 kHz	0.44 %	
		300 kHz to 1 MHz	1.5 %	
	500 mV to 1 V	1 Hz to 40 Hz	0.024 %	
		40 Hz to 100 Hz	0.017 %	
		100 Hz to 330 Hz	0.016 %	
		330 Hz to 2 kHz	0.014 %	
		2 kHz to 10 kHz	0.015 %	
		10 kHz to 30 kHz	0.030 %	
		30 kHz to 100 kHz	0.093 %	
		100 kHz to 300 kHz	0.39 %	
		300 kHz to 1 MHz	1.4 %	
	1 V to 2 V	1 Hz to 40 Hz	0.019 %	
		40 Hz to 100 Hz	0.014 %	
		100 Hz to 2 kHz	0.012 %	
		2 kHz to 10 kHz	0.013 %	
		10 kHz to 30 kHz	0.026 %	
		30 kHz to 100 kHz	0.072 %	
		100 kHz to 300 kHz	0.49 %	
		300 kHz to 1 MHz	1.5 %	
	2 V to 5 V	10 Hz to 40 Hz	0.035 %	
		40 Hz to 100 Hz	0.026 %	
		100 Hz to 2 kHz	0.025 %	
		2 kHz to 10 kHz	0.022 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2		10 kHz to 30 kHz	0.043 %	TP2, TP21
		30 kHz to 100 kHz	0.17 %	
		100 kHz to 300 kHz	0.43 %	
		300 kHz to 1 MHz	1.5 %	
	5 V to 10 V	10 Hz to 40 Hz	0.023 %	
		40 Hz to 100 Hz	0.017 %	
		100 Hz to 2 kHz	0.016 %	
		2 kHz to 10 kHz	0.015 %	
		10 kHz to 30 kHz	0.030 %	
		30 kHz to 100 kHz	0.093 %	
		100 kHz to 300 kHz	0.39 %	
		300 kHz to 1 MHz	1.4 %	
	10 V to 20 V	1 Hz to 40 Hz	0.019 %	
		40 Hz to 100 Hz	0.014 %	
		100 Hz to 2 kHz	0.012 %	
		2 kHz to 10 kHz	0.013 %	
		10 kHz to 30 kHz	0.026 %	
		30 kHz to 100 kHz	0.072 %	
		100 kHz to 300 kHz	0.52 %	
		300 kHz to 1 MHz	3.1 %	
	20 V to 50 V	10 Hz to 40 Hz	0.026 %	
		40 Hz to 100 Hz	0.027 %	
		100 Hz to 2 kHz	0.026 %	
		2 kHz to 10 kHz	0.023 %	
		10 kHz to 30 kHz	0.045 %	
		30 kHz to 100 kHz	0.17 %	
	50 V to 100 V	10 Hz to 40 Hz	0.021 %	
		40 Hz to 100 Hz	0.018 %	
		100 Hz to 2 kHz	0.017 %	
		2 kHz to 10 kHz	0.016 %	
		10 kHz to 30 kHz	0.032 %	
		30 kHz to 100 kHz	0.094 %	
	100 V to 200 V	1 Hz to 40 Hz	0.019 %	
		40 Hz to 100 Hz	0.015 %	
		100 Hz to 330 Hz	0.013 %	
		330 Hz to 2 kHz	0.012 %	
		2 kHz to 10 kHz	0.013 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2		10 kHz to 30 kHz	0.027 %	TP2, TP21
		30 kHz to 100 kHz	0.074 %	
	200 V to 500 V	45 Hz to 10 kHz	0.035 %	
		10 kHz to 30 kHz	0.050 %	
	500 V to 1000 V	45 Hz to 10 kHz	0.026 %	
		10 kHz to 30 kHz	0.037 %	
2*	Alternating-current voltage Measurement 1 mV to 5 mV	10 Hz to 40 Hz	8 µV	TP2, TP21
		40 Hz to 10 kHz	5 µV	
		10 kHz to 20 kHz	8 µV	
		20 kHz to 50 kHz	11 µV	
		50 kHz to 100 kHz	0.87 %	
	5 mV to 10 mV	10 Hz to 40 Hz	0.15 %	
		40 Hz to 1 kHz	0.086 %	
		1 kHz to 10 kHz	0.092 %	
		10 kHz to 20 kHz	0.16 %	
		20 kHz to 50 kHz	0.21 %	
		50 kHz to 100 kHz	0.63 %	
	10 mV to 20 mV	10 Hz to 40 Hz	0.086 %	
		40 Hz to 1 kHz	0.058 %	
		1 kHz to 10 kHz	0.063 %	
		10 kHz to 20 kHz	0.11 %	
		20 kHz to 50 kHz	0.12 %	
		50 kHz to 100 kHz	0.16 %	
		100 kHz to 300 kHz	0.72 %	
		300 kHz to 1 MHz	2.1 %	
	20 mV to 50 mV	10 Hz to 40 Hz	0.061 %	
		40 Hz to 1 kHz	0.037 %	
		1 kHz to 10 kHz	0.041 %	
		10 kHz to 20 kHz	0.085 %	
		20 kHz to 50 kHz	0.093 %	
		50 kHz to 100 kHz	0.14 %	
		100 kHz to 300 kHz	0.51 %	
		300 kHz to 1 MHz	1.6 %	
	50 mV to 100 mV	10 Hz to 40 Hz	0.047 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2*		40 Hz to 1 kHz	0.024 %	TP2, TP21
		1 kHz to 10 kHz	0.029 %	
		10 kHz to 20 kHz	0.071 %	
		20 kHz to 50 kHz	0.079 %	
		50 kHz to 100 kHz	0.12 %	
		100 kHz to 300 kHz	0.41 %	
		300 kHz to 1 MHz	1.3 %	
	100 mV to 200 mV	10 Hz to 40 Hz	0.068 %	
		40 Hz to 1 kHz	0.036 %	
		1 kHz to 10 kHz	0.043 %	
		10 kHz to 20 kHz	0.075 %	
		20 kHz to 50 kHz	0.086 %	
		50 kHz to 100 kHz	0.14 %	
		100 kHz to 300 kHz	0.48 %	
		300 kHz to 1 MHz	1.4 %	
	200 mV to 500 mV	10 Hz to 40 Hz	0.046 %	
		40 Hz to 1 kHz	0.023 %	
		1 kHz to 10 kHz	0.030 %	
		10 kHz to 20 kHz	0.048 %	
		20 kHz to 50 kHz	0.061 %	
		50 kHz to 100 kHz	0.12 %	
		100 kHz to 300 kHz	0.42 %	
		300 kHz to 1 MHz	1.3 %	
	500 mV to 1 V	10 Hz to 40 Hz	0.035 %	
		40 Hz to 1 kHz	0.015 %	
		1 kHz to 10 kHz	0.023 %	
		10 kHz to 20 kHz	0.032 %	
		20 kHz to 50 kHz	0.046 %	
		50 kHz to 100 kHz	0.10 %	
		100 kHz to 300 kHz	0.38 %	
		300 kHz to 1 MHz	1.3 %	
	1 V to 2 V	10 Hz to 40 Hz	0.032 %	
		40 Hz to 1 kHz	0.013 %	
		1 kHz to 10 kHz	0.041 %	
		10 kHz to 20 kHz	0.044 %	
		20 kHz to 50 kHz	0.061 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2*		50 kHz to 100 kHz	0.12 %	TP2, TP21
		100 kHz to 300 kHz	0.48 %	
		300 kHz to 1 MHz	1.5 %	
	2 V to 5 V	10 Hz to 40 Hz	0.042 %	
		40 Hz to 1 kHz	0.021 %	
		1 kHz to 10 kHz	0.029 %	
		10 kHz to 20 kHz	0.033 %	
		20 kHz to 50 kHz	0.050 %	
		50 kHz to 100 kHz	0.11 %	
		100 kHz to 300 kHz	0.41 %	
		300 kHz to 1 MHz	1.3 %	
	5 V to 10 V	10 Hz to 40 Hz	0.033 %	
		40 Hz to 1 kHz	0.015 %	
		1 kHz to 10 kHz	0.022 %	
		10 kHz to 20 kHz	0.026 %	
		20 kHz to 50 kHz	0.043 %	
		50 kHz to 100 kHz	0.10 %	
		100 kHz to 300 kHz	0.38 %	
		300 kHz to 1 MHz	1.3 %	
	10 V to 20 V	10 Hz to 40 Hz	0.075 %	
		40 Hz to 10 kHz	0.047 %	
		10 kHz to 20 kHz	0.049 %	
		20 kHz to 50 kHz	0.066 %	
		50 kHz to 100 kHz	0.17 %	
	20 V to 50 V	10 Hz to 40 Hz	0.055 %	
		40 Hz to 10 kHz	0.036 %	
		10 kHz to 20 kHz	0.039 %	
		20 kHz to 50 kHz	0.055 %	
		50 kHz to 100 kHz	0.16 %	
	50 V to 100 V	10 Hz to 40 Hz	0.043 %	
		40 Hz to 10 kHz	0.029 %	
		10 kHz to 20 kHz	0.034 %	
		20 kHz to 50 kHz	0.050 %	
		50 kHz to 100 kHz	0.15 %	
	100 V to 200 V	40 Hz to 1 kHz	0.070 %	
		1 kHz to 10 kHz	0.094 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2*		10 kHz to 20 kHz	0.17 %	TP2, TP21
		20 kHz to 50 kHz	0.21 %	
		50 kHz to 90 kHz	0.38 %	
	200 V to 500 V	40 Hz to 1 kHz	0.059 %	
		1 kHz to 10 kHz	0.083 %	
		10 kHz to 20 kHz	0.24 %	
		20 kHz to 50 kHz	0.27 %	
		50 kHz to 90 kHz	0.39 %	
	500 V to 700 V	40 Hz to 1 kHz	0.053 %	
		1 kHz to 10 kHz	0.077 %	
		10 kHz to 20 kHz	0.25 %	
		20 kHz to 50 kHz	0.28 %	
		50 kHz to 90 kHz	0.42 %	
	Alternating-current voltage Generation			TP2, TP21
	2 mV to 5 mV	10 Hz to 30 kHz	1.6 %	
		30 kHz to 50 kHz	2.0 %	
		50 kHz to 100 kHz	2.8 %	
	5 mV to 10 mV	10 Hz to 45 Hz	0.65 %	
		45 Hz to 10 kHz	0.59 %	
		10 kHz to 20 kHz	0.68 %	
		20 kHz to 30 kHz	0.74 %	
		30 kHz to 50 kHz	0.90 %	
		50 kHz to 100 kHz	1.3 %	
	10 mV to 20 mV	10 Hz to 45 Hz	0.42 %	
		45 Hz to 10 kHz	0.36 %	
		10 kHz to 20 kHz	0.44 %	
		20 kHz to 30 kHz	0.49 %	
		30 kHz to 50 kHz	0.58 %	
		50 kHz to 100 kHz	0.86 %	
	20 mV to 50 mV	10 Hz to 45 Hz	0.17 %	
		45 Hz to 1 kHz	0.11 %	
		1 kHz to 10 kHz	0.18 %	
		10 kHz to 20 kHz	0.20 %	
		20 kHz to 30 kHz	0.35 %	
		30 kHz to 100 kHz	0.39 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2*		100 kHz to 300 kHz	1.1 %	TP2, TP21
		300 kHz to 500 kHz	1.9 %	
	50 mV to 100 mV	10 Hz to 45 Hz	0.11 %	
		45 Hz to 1 kHz	0.053 %	
		1 kHz to 10 kHz	0.083 %	
		10 kHz to 20 kHz	0.093 %	
		20 kHz to 30 kHz	0.21 %	
		30 kHz to 100 kHz	0.24 %	
		100 kHz to 300 kHz	0.71 %	
		300 kHz to 500 kHz	1.5 %	
	100 mV to 200 mV	10 Hz to 45 Hz	0.089 %	
		45 Hz to 1 kHz	0.033 %	
		1 kHz to 10 kHz	0.049 %	
		10 kHz to 20 kHz	0.061 %	
		20 kHz to 30 kHz	0.16 %	
		30 kHz to 100 kHz	0.18 %	
		100 kHz to 300 kHz	0.67 %	
		300 kHz to 500 kHz	1.5 %	
	200 mV to 500 mV	10 Hz to 45 Hz	0.16 %	
		45 Hz to 1 kHz	0.087 %	
		1 kHz to 20 kHz	0.13 %	
		20 kHz to 30 kHz	0.25 %	
		30 kHz to 100 kHz	0.29 %	
		100 kHz to 300 kHz	0.70 %	
		300 kHz to 1 MHz	1.6 %	
	500 mV to 1 V	10 Hz to 45 Hz	0.096 %	
		45 Hz to 1 kHz	0.046 %	
		1 kHz to 10 kHz	0.062 %	
		10 kHz to 20 kHz	0.067 %	
		20 kHz to 30 kHz	0.15 %	
		30 kHz to 100 kHz	0.17 %	
		100 kHz to 300 kHz	0.55 %	
		300 kHz to 1 MHz	1.5 %	
	1 V to 2 V	10 Hz to 45 Hz	0.075 %	
		45 Hz to 1 kHz	0.033 %	
		1 kHz to 10 kHz	0.041 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2*		10 kHz to 20 kHz	0.047 %	TP2, TP21
		20 kHz to 30 kHz	0.11 %	
		30 kHz to 100 kHz	0.13 %	
		100 kHz to 300 kHz	0.60 %	
		300 kHz to 1 MHz	1.7 %	
	2 V to 5 V	10 Hz to 45 Hz	0.15 %	
		45 Hz to 1 kHz	0.075 %	
		1 kHz to 20 kHz	0.12 %	
		20 kHz to 30 kHz	0.25 %	
		30 kHz to 100 kHz	0.29 %	
	5 V to 10 V	10 Hz to 45 Hz	0.088 %	
		45 Hz to 1 kHz	0.041 %	
		1 kHz to 10 kHz	0.057 %	
		10 kHz to 20 kHz	0.063 %	
		20 kHz to 30 kHz	0.14 %	
		30 kHz to 100 kHz	0.16 %	
	10 V to 20 V	10 Hz to 45 Hz	0.070 %	
		45 Hz to 1 kHz	0.029 %	
		1 kHz to 10 kHz	0.039 %	
		10 kHz to 20 kHz	0.045 %	
		20 kHz to 100 kHz	0.12 %	
	20 V to 50 V	30 Hz to 45 Hz	0.17 %	
		45 Hz to 1 kHz	0.086 %	
		1 kHz to 10 kHz	0.12 %	
		10 kHz to 30 kHz	0.21 %	
		30 kHz to 40 kHz	0.26 %	
	50 V to 100 V	30 Hz to 45 Hz	0.098 %	
		45 Hz to 1 kHz	0.045 %	
		1 kHz to 10 kHz	0.058 %	
		10 kHz to 30 kHz	0.11 %	
		30 kHz to 40 kHz	0.14 %	
	100 V to 200 V	30 Hz to 45 Hz	0.075 %	
		45 Hz to 1 kHz	0.031 %	
		1 kHz to 10 kHz	0.039 %	
		10 kHz to 30 kHz	0.068 %	
		30 kHz to 40 kHz	0.095 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2*	200 V to 500 V	45 Hz to 1 kHz	0.056 %	TP2, TP21
		1 kHz to 10 kHz	0.095 %	
	500 V to 1000 V	46 Hz to 1 kHz	0.035 %	
		1 kHz to 10 kHz	0.054 %	
3	Direct current Measurement 0 nA to 20 µA 20 µA to 50 µA 50 µA to 100 µA 100 µA to 200 µA 200 µA to 500 µA 500 µA to 1 mA 1 mA to 2 mA 2 mA to 5 mA 5 mA to 10 mA 10 mA to 20 mA 20 mA to 50 mA 50 mA to 100 mA 100 mA to 200 mA 200 mA to 500 mA 500 mA to 1 A 1 A to 100 A 200 A 300 A 400 A 500 A 600 A		0.6 nA 0.0029 % 0.0017 % 0.0013 % 0.0028 % 0.0017 % 0.0013 % 0.0029 % 0.0017 % 0.0013 % 0.0074 % 0.0050 % 0.0042 % 0.024 % 0.019 % 0.020 % 0.020 % 0.020 % 0.020 % 0.020 % 0.020 %	TP3, TP21
	Direct current Generation 0.2 µA to 20 µA 20 µA to 50 µA 50 µA to 100 µA 100 µA to 200 µA 200 µA to 500 µA 500 µA to 1 mA 1 mA to 2 mA 2 mA to 5 mA		5 nA 0.024 % 0.017 % 0.015 % 0.011 % 0.0072 % 0.0060 % 0.011 %	TP3, TP21

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
3	5 mA to 10 mA 10 mA to 20 mA 20 mA to 50 mA 50 mA to 100 mA 100 mA to 200 mA 200 mA to 500 mA 500 mA to 1 A 1 A to 2 A 2 A to 11 A 11 A to 1000 A		0.0072 % 0.0060 % 0.013 % 0.0086 % 0.0072 % 0.035 % 0.027 % 0.031 % 0.029 % 0.50 %	TP3, TP21
3*	Direct current Measurement 0 nA to 10 µA 10 µA to 20 µA 20 µA to 50 µA 50 µA to 100 µA 100 µA to 200 µA 200 µA to 500 µA 500 µA to 1 mA 1 mA to 2 mA 2 mA to 5 mA 5 mA to 10 mA 10 mA to 20 mA 20 mA to 50 mA 50 mA to 100 mA 100 mA to 200 mA 200 mA to 500 mA 500 mA to 1 A 1 A to 100 A 200 A 300 A 400 A 500 A 600 A		7 nA 0.066 % 0.035 % 0.017 % 0.014 % 0.0088 % 0.0061 % 0.012 % 0.0076 % 0.0056 % 0.016 % 0.011 % 0.0079 % 0.032 % 0.024 % 0.019 % 0.020 % 0.020 % 0.020 % 0.020 % 0.020 % 0.020 %	TP3, TP21

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
3*	Direct current			TP3, TP21
	Generation			
	1 µA to 20 µA		14 nA	
	20 µA to 50 µA		0.068 %	
	50 µA to 100 µA		0.033 %	
	100 µA to 200 µA		0.021 %	
	200 µA to 500 µA		0.023 %	
	500 µA to 1 mA		0.012 %	
	1 mA to 2 mA		0.0082 %	
	2 mA to 5 mA		0.017 %	
	5 mA to 10 mA		0.0094 %	
	10 mA to 20 mA		0.0071 %	
	20 mA to 50 mA		0.018 %	
	50 mA to 100 mA		0.011 %	
	100 mA to 200 mA		0.0082 %	
	200 mA to 500 mA		0.037 %	
	500 mA to 1 A		0.026 %	
	1 A to 2 A		0.023 %	
	2 A to 5 A		0.046 %	
	5 A to 10 A		0.037 %	
	10 A to 20 A		0.034 %	
	20 A to 30 A		0.042 %	
	30 A to 1000 A		0.50 %	
4	Alternating current			TP4, TP21
	Measurement			
	10 µA to 20 µA	10 Hz to 5 kHz	29 nA	
	20 µA to 50 µA	10 Hz to 5 kHz	0.15 %	
	50 µA to 100 µA	10 Hz to 1 kHz	0.071 %	
		1 kHz to 5 kHz	0.082 %	
	100 µA to 200 µA	10 Hz to 1 kHz	0.046 %	
		1 kHz to 5 kHz	0.048 %	
	200 µA to 500 µA	10 Hz to 10 kHz	0.14 %	
	500 µA to 1 mA	10 Hz to 5 kHz	0.067 %	
		5 kHz to 10 kHz	0.071 %	
	1 mA to 2 mA	10 Hz to 10 kHz	0.047 %	
	2 mA to 5 mA	10 Hz to 10 kHz	0.13 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
4	5 mA to 10 mA	10 Hz to 10 kHz	0.066 %	TP4, TP21
	10 mA to 20 mA	10 Hz to 10 kHz	0.046 %	
	20 mA to 50 mA	10 Hz to 10 kHz	0.13 %	
	50 mA to 100 mA	10 Hz to 10 kHz	0.066 %	
	100 mA to 200 mA	10 Hz to 10 kHz	0.046 %	
	200 mA to 500 mA	10 Hz to 10 kHz	0.18 %	
	500 mA to 1 A	10 Hz to 10 kHz	0.12 %	
	1 A to 2 A	10 Hz to 2 kHz	0.081 %	
		2 kHz to 10 kHz	0.093 %	
	2 A to 5 A	10 Hz to 2 kHz	0.19 %	
		2 kHz to 10 kHz	0.36 %	
	5 A to 10 A	10 Hz to 2 kHz	0.13 %	
		2 kHz to 10 kHz	0.30 %	
	10 A to 20 A	10 Hz to 2 kHz	0.11 %	
		2 kHz to 10 kHz	0.28 %	
	20 A to 25 A	50 Hz	0.44 %	
	25 A to 50 A	50 Hz	0.27 %	
	50 A to 60 A	50 Hz	0.20 %	
	60 A to 70 A	50 Hz	0.33 %	
	70 A to 80 A	50 Hz	0.30 %	
	80 A to 120 A	50 Hz	0.28 %	
	120 A to 150 A	50 Hz	0.24 %	
	150 A to 300 A	50 Hz	0.21 %	
	300 A to 600 A	50 Hz	0.18 %	
	600 A to 1200 A	50 Hz	0.23 %	
	Alternating current Generation			TP4, TP21
	10 µA to 20 µA	10 Hz to 1 kHz	34 nA	
		1 kHz to 5 kHz	39 nA	
	20 µA to 50 µA	10 Hz to 1 kHz	0.17 %	
		1 kHz to 5 kHz	0.19 %	
	50 µA to 100 µA	10 Hz to 1 kHz	0.082 %	
		1 kHz to 5 kHz	0.11 %	
	100 µA to 200 µA	10 Hz to 1 kHz	0.055 %	
		1 kHz to 5 kHz	0.070 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
4	200 µA to 500 µA	10 Hz to 5 kHz	0.16 %	TP4, TP21
	500 µA to 1 mA	10 Hz to 1 kHz	0.076 %	
		1 kHz to 5 kHz	0.082 %	
	1 mA to 2 mA	10 Hz to 40 Hz	0.052 %	
		40 Hz to 1 kHz	0.058 %	
	2 mA to 5 mA	10 Hz to 5 kHz	0.16 %	
	5 mA to 10 mA	10 Hz to 1 kHz	0.075 %	
		1 kHz to 5 kHz	0.081 %	
	10 mA to 20 mA	10 Hz to 1 kHz	0.052 %	
		1 kHz to 5 kHz	0.058 %	
	20 mA to 50 mA	10 Hz to 1 kHz	0.16 %	
	50 mA to 100 mA	10 Hz to 1 kHz	0.075 %	
		1 kHz to 5 kHz	0.081 %	
	100 mA to 200 mA	10 Hz to 1 kHz	0.052 %	
		1 kHz to 5 kHz	0.058 %	
	200 mA to 500 mA	10 Hz to 1 kHz	0.21 %	
		1 kHz to 5 kHz	0.23 %	
	500 mA to 1 A	10 Hz to 5 kHz	0.15 %	
	1 A to 2 A	10 Hz to 1 kHz	0.22 %	
		1 kHz to 5 kHz	0.30 %	
	2 A to 5 A	10 Hz to 45 kHz	0.23 %	
		45 Hz to 1 kHz	0.17 %	
		1 kHz to 2 kHz	0.27 %	
		2 kHz to 5 kHz	0.41 %	
	5 A to 11 A	10 Hz to 45 kHz	0.10 %	
		45 Hz to 1 kHz	0.13 %	
		1 kHz to 2 kHz	0.19 %	
		2 kHz to 5 kHz	0.33 %	
	11 A to 20 A	10 Hz to 45 Hz	0.24 %	
		45 Hz to 100 Hz	0.13 %	
		100 Hz to 1 kHz	0.52 %	
	20 A to 25 A	50 Hz	0.45 %	
	25 A to 30 A	50 Hz	0.28 %	
	30 A to 1000 A	50 Hz	0.70 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
4*	Alternating current Measurement			TP4, TP21
	10 µA to 20 µA	20 Hz to 45 Hz 45 Hz to 1 kHz	71 nA 50 nA	
	20 µA to 50 µA	20 Hz to 45 Hz 45 Hz to 1 kHz	0.36 % 0.25 %	
	50 µA to 100 µA	20 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz	0.25 % 0.15 % 0.073 %	
	100 µA to 200 µA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz	0.41 % 0.31 % 0.27 % 0.31 %	
	200 µA to 500 µA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz	0.30 % 0.19 % 0.16 % 0.18 %	
	500 µA to 1 mA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz	0.22 % 0.12 % 0.083 % 0.093 %	
	1 mA to 2 mA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.41 % 0.31 % 0.27 %	
	2 mA to 5 mA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.29 % 0.19 % 0.16 %	
	5 mA to 10 mA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz	0.23 % 0.12 % 0.083 % 0.085 %	
	10 mA to 20 mA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.41 % 0.31 % 0.27 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
4*	20 mA to 50 mA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.29 % 0.19 % 0.16 %	TP4, TP21
	50 mA to 100 mA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz	0.23 % 0.12 % 0.083 % 0.085 %	
	100 mA to 200 mA	20 Hz to 45 Hz 45 Hz to 1 kHz 100 Hz to 5 kHz	0.42 % 0.32 % 0.35 %	
	200 mA to 500 mA	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.31 % 0.21 % 0.24 %	
	500 mA to 1 A	20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.24 % 0.15 % 0.17 %	
	1 A to 15 A	45 Hz to 1 kHz	0.10 %	
	15 A to 25 A	50 Hz	0.44 %	
	25 A to 50 A	50 Hz	0.27 %	
	50 A to 60 A	50 Hz	0.20 %	
	60 A to 70 A	50 Hz	0.33 %	
	70 A to 80 A	50 Hz	0.30 %	
	80 A to 120 A	50 Hz	0.28 %	
	120 A to 150 A	50 Hz	0.24 %	
	150 A to 300 A	50 Hz	0.21 %	
	300 A to 600 A	50 Hz	0.18 %	
	600 A to 1200 A	50 Hz	0.23 %	
	Alternating current Generation			TP4, TP21
	20 µA to 50 µA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	1.7 % 0.94 % 2.1 %	
	50 µA to 100 µA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.77 % 0.42 % 1.4 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
4*	100 µA to 200 µA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.48 % 0.24 % 1.2 %	TP4, TP21
	200 µA to 500 µA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.35 % 0.21 % 0.66 %	
	500 µA to 1 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.25 % 0.12 % 0.54 %	
	1 mA to 2 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.21 % 0.084 % 0.51 %	
	2 mA to 5 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.38 % 0.21 % 0.43 %	
	5 mA to 10 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.26 % 0.11 % 0.31 %	
	10 mA to 20 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.22 % 0.079 % 0.28 %	
	20 mA to 50 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.38 % 0.21 % 0.71 %	
	50 mA to 100 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.26 % 0.11 % 0.56 %	
	100 mA to 200 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.22 % 0.079 % 0.52 %	
	200 mA to 500 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz	0.40 % 0.25 % 0.72 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
4*	500 mA to 1 A	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz	0.28 % 0.16 % 0.57 %	TP4, TP21
	1 A to 2 A	10 Hz to 45 Hz 45 Hz to 1 kHz	0.23 % 0.11 %	
		1 kHz to 5 kHz	0.52 %	
	2 A to 5 A	10 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz	0.40 % 0.20 % 0.71 %	
	5 A to 10 A	10 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz	0.28 % 0.14 % 0.57 %	
	10 A to 20 A	10 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz	0.24 % 0.13 % 0.52 %	
	20 A to 25 A	50 Hz	0.45 %	
	25 A to 30 A	50 Hz	0.28 %	
	30 A to 1000 A	50 Hz	0.70 %	
5	DC resistance Measurement 0 Ω to 1 Ω 1 Ω to 5 Ω 5 Ω to 10 Ω 10 Ω to 20 Ω 20 Ω to 50 Ω 50 Ω to 100 Ω 100 Ω to 200 Ω 200 Ω to 500 Ω 500 Ω to 1 kΩ 1 kΩ to 2 kΩ 2 kΩ to 5 kΩ 5 kΩ to 10 kΩ 10 kΩ to 20 kΩ 20 kΩ to 50 kΩ 50 kΩ to 100 kΩ 100 kΩ to 200 kΩ		15 μΩ 0.0015 % 0.0011 % 0.00089 % 0.0011 % 0.00085 % 0.00079 % 0.0010 % 0.00084 % 0.00079 % 0.0010 % 0.00084 % 0.00079 % 0.0011 % 0.00087 % 0.00081 %	TP5, TP21

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
5	200 kΩ to 500 kΩ 500 kΩ to 1 MΩ 1 MΩ to 2 MΩ 2 MΩ to 5 MΩ 5 MΩ to 20 MΩ 20 MΩ to 50 MΩ 50 MΩ to 100 MΩ 100 MΩ to 200 MΩ 200 MΩ to 500 MΩ 500 MΩ to 1 GΩ 1 GΩ to 2 GΩ 2 GΩ to 5 GΩ 5 GΩ to 10 GΩ 10 GΩ to 20 GΩ		0.0014 % 0.0011 % 0.0010 % 0.0021 % 0.0018 % 0.0079 % 0.0055 % 0.0049 % 0.063 % 0.030 % 0.020 % 0.64 % 0.30 % 0.20 %	TP5, TP21
	DC resistance Generation 0.1 mΩ 1 mΩ 10 mΩ, 100 mΩ 1 Ω, 10 Ω 100 Ω 1 kΩ, 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ 1 GΩ 10 GΩ 100 GΩ		0.01 % 0.008 % 0.002 % 0.001 % 0.0007 % 0.0005 % 0.0006 % 0.0042 % 0.0066 % 0.009 % 0.014 % 0.1 % 0.54 %	
5*	DC resistance Measurement 0 Ω to 0.5 Ω 0.5 Ω to 1 Ω 1 Ω to 2 Ω 2 Ω to 5 Ω 5 Ω to 10 Ω 10 Ω to 20 Ω		0.10 mΩ 0.018 % 0.011 % 0.0018 % 0.0041 % 0.0075 %	TP5, TP21

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
5*	20 Ω to 50 Ω 50 Ω to 100 Ω 100 Ω to 200 Ω 200 Ω to 500 Ω 500 Ω to 1 kΩ 1 kΩ to 2 kΩ 2 kΩ to 5 kΩ 5 kΩ to 10 kΩ 10 kΩ to 20 kΩ 20 kΩ to 50 kΩ 50 kΩ to 100 kΩ 100 kΩ to 200 kΩ 200 kΩ to 500 kΩ 500 kΩ to 1 MΩ 1 MΩ to 2 MΩ 2 MΩ to 5 MΩ 5 MΩ to 10 MΩ 10 MΩ to 20 MΩ 20 MΩ to 50 MΩ 50 MΩ to 100 MΩ 100 MΩ to 1 GΩ		0.0046 % 0.0028 % 0.0021 % 0.0018 % 0.0016 % 0.0020 % 0.0017 % 0.0016 % 0.0022 % 0.0019 % 0.0017 % 0.0046 % 0.0036 % 0.0030 % 0.018 % 0.013 % 0.0091 % 0.071 % 0.065 % 0.061 % 0.59 %	TP5, TP21
	DC resistance Generation 0.1 mΩ 1 mΩ, 10 mΩ 100 mΩ 1 Ω 10 Ω, 100 Ω 1 kΩ, 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ 1 GΩ 10 GΩ 100 GΩ		0.02 % 0.012 % 0.008 % 0.005 % 0.004 % 0.003 % 0.0035 % 0.0042 % 0.0066 % 0.009 % 0.014 % 0.1 % 0.54 %	TP5, TP21

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
5*	DC resistance Generation 0 Ω to 1 Ω 1 Ω to 5 Ω 5 Ω to 10 Ω 10 Ω to 11 Ω 11 Ω to 20 Ω 20 Ω to 33 Ω 33 Ω to 110 Ω 110 Ω to 330 Ω 330 Ω to 1.1 kΩ 1.1 kΩ to 3.3 kΩ 3.3 kΩ to 11 kΩ 11 kΩ to 33 kΩ 33 kΩ to 110 kΩ 110 kΩ to 330 kΩ 330 kΩ to 1.1 MΩ 1.1 MΩ to 3.3 MΩ 3.3 MΩ to 11 MΩ 11 MΩ to 33 MΩ 33 MΩ to 110 MΩ 110 MΩ to 330 MΩ 330 MΩ to 1.1 GΩ		1.5 mΩ 0.13 % 0.021 % 0.026 % 0.030 % 0.023 % 0.016 % 0.016 % 0.012 % 0.013 % 0.012 % 0.013 % 0.014 % 0.016 % 0.019 % 0.021 % 0.072 % 0.15 % 0.59 % 0.69 % 2.0 %	TP5, TP21
6	A.C. resistance Measurement 0.1 Ω 1 Ω 10 Ω, 100 Ω, 1 kΩ, 10 kΩ, 100 kΩ 1 MΩ 10 MΩ	50 Hz to 10 kHz 20 kHz 50 Hz to 10 kHz 20 kHz 50 Hz to 10 kHz 20 kHz 50 Hz to 20 kHz 50 Hz to 4 kHz 10 kHz 20 kHz	0.5 % 1 % 0.3 % 0.5 % 0.1 % 0.2 % 0.2 % 2 % 3 % 5 %	TP6, TP24

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
6*	A.C. resistance Generation 0,1 Ω, 1 Ω	1 kHz	0.02 %	TP6, TP24
	10 Ω, 100 Ω, 1 kΩ, 10 kΩ, 100 kΩ	1 kHz 1 kHz	0.01 % 0.01 %	
	A.C. resistance Generation, real part of impedance 0.1 Ω	10 Hz to 10 kHz 10 kHz to 0.3 MHz 0.3 MHz to 0.5 MHz 0.5 MHz to 1 MHz	0.03 % 0.45 % 0.55 % 0.8 %	
	1 Ω	10 Hz to 10 kHz 10 kHz to 0.3 MHz 0.3 MHz to 0.5 MHz 0.5 MHz to 1 MHz 1 MHz to 3 MHz 3 MHz to 13 MHz	0.02 % 0.15 % 0.20 % 0.30 % 1.5 % 2.5 %	
	10 Ω	10 Hz to 10 kHz 10 kHz to 0.3 MHz 0.3 MHz to 1 MHz 1 MHz to 5 MHz 5 MHz to 13 MHz	0.02 % 0.05 % 0.08 % 0.30 % 1.0 %	
	100 Ω, 1 kΩ	10 Hz to 10 kHz 10 kHz to 0.3 MHz 0.3 MHz to 1 MHz 1 MHz to 5 MHz 5 MHz to 13 MHz	0.02 % 0.05 % 0.09 % 0.30 % 0.8 %	
	10 kΩ	10 Hz to 10 kHz 10 kHz to 0.3 MHz 0.3 MHz to 0.5 MHz 0.5 MHz to 1 MHz	0.02 % 0.05 % 0.07 % 0.09 %	
	100 kΩ	10 Hz to 10 kHz 10 kHz to 0.5 MHz 0.5 MHz to 1 MHz	0.02 % 0.15 % 0.25 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
6*	A.C. resistance Measurement 0.1 Ω to 1 Ω	42 Hz to 100 Hz 100 Hz to 100 kHz	6.5 % 4.7 %	TP6, TP24
	1 Ω to 10 Ω	42 Hz to 100 Hz 100 Hz to 100 kHz 100 kHz to 1 MHz	1.2 % 0.71 % 2.0 %	
	10 Ω to 100 Ω	42 Hz to 100 Hz 100 Hz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 5 MHz	0.80 % 0.39 % 0.57 % 0.86 % 5.0 %	
	100 Ω to 10 kΩ	42 Hz to 100 Hz 100 Hz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 5 MHz	0.66 % 0.26 % 0.57 % 0.86 % 5.0 %	
	10 kΩ to 100 kΩ	42 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 5 MHz	0.66 % 0.26 % 0.36 % 0.92 % 4.7 % 9.8 %	
	100 kΩ to 1 MΩ	42 Hz to 100 Hz 100 Hz to 10 kHz 10 kHz to 100 kHz	1.3 % 0.90 % 1.6 %	
	1 MΩ to 10 MΩ	42 Hz to 100 Hz 100 Hz to 10 kHz 10 kHz to 100 kHz	6.6 % 3.3 % 8.3 %	
7	Capacity Measurement 2 pF to 20 pF	10 kHz 20 kHz	0.3 % 1.5 %	TP7, TP21, TP24
	20 pF to 200 pF	1 kHz to 4 kHz 10 kHz	0.3 % 0.2 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
7	0.20 nF to 2 nF	100 Hz to 400 Hz 1 kHz to 4 kHz 10 kHz 20 kHz	0.3 % 0.2 % 0.1 % 0.5 %	TP7, TP21, TP24
	2 nF to 20 nF	100 Hz to 400 Hz 1 kHz to 10 kHz 20 kHz	0.2 % 0.1 % 0.5 %	
	20 nF to 200 nF	100 Hz to 10 kHz 20 kHz	0.1 % 0.5 %	
	0.2 µF to 2 µF	100 Hz to 4 kHz 10 kHz	0.1 % 0.5 %	
	2 µF to 20 µF	100 Hz to 400 Hz 1 kHz to 4 kHz 10 kHz 20 kHz	0.1 % 0.5 % 1 % 2 %	
	20 µF to 200 µF	100 Hz to 400 Hz 1 kHz to 4 kHz 10 kHz 20 kHz	0.5 % 1 % 2 % 5 %	
	0.2 mF to 2 mF	100 Hz to 400 Hz 1 kHz to 4 kHz 10 kHz 20 kHz	1 % 2 % 5 % 10 %	
	2 mF to 20 mF	100 Hz to 400 Hz 1 kHz to 4 kHz	2 % 5 %	
	20 mF to 200 mF 1 pF, 10 pF, 100 pF, 1000 pF, 10 nF	100 Hz to 400 Hz 1 kHz	5 % 0.003 %	
	100 nF, 1 µF	1 kHz	0.005 %	
	10 µF	1 kHz	0.007 %	
	100 µF	1 kHz	0.02 %	
7*	Capacity Measurement 160 pF to 380 pF 1.6 nF to 16 nF 16 nF to 16 µF	42 Hz to 100 Hz	6.6 % 1.3 % 0.66 %	TP7, TP21, TP24

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
7*	16 µF to 160 µF		0.53 %	TP7, TP21, TP24
	160 µF to 1.6 mF		0.60 %	
	1.6 mF to 16 mF		1.1 %	
	16 pF to 160 pF	100 Hz to 1000 Hz	3.3 %	
	160 pF to 1.6 nF		0.90 %	
	1.6 nF to 1.6 µF		0.26 %	
	1.6 µF to 16 µF		0.12 %	
	16 µF to 160 µF		0.30 %	
	160 µF to 1.6 mF		0.60 %	
	1.6 pF to 16 pF	1 kHz to 10 kHz	3.3 %	
	16 pF to 160 pF		0.90 %	
	160 pF to 1.6 nF		0.36 %	
	1.6 nF to 160 nF		0.26 %	
	160 nF to 1.6 µF		0.12 %	
	1.6 µF to 16 µF		0.30 %	
	16 µF to 160 µF		0.60 %	
	1.6 pF to 16 pF	10 kHz to 100 kHz	1.6 %	
	16 pF to 160 pF		1.0 %	
	160 pF to 16 nF		0.57 %	
	16 nF to 1.6 µF		0.30 %	
	1.6 µF to 16 µF		0.60 %	
	1.6 pF to 16 pF	100 kHz to 1 MHz	4.7 %	
	16 pF to 1.6 nF		0.86 %	
	1.6 nF to 16 nF		0.45 %	
	16 nF to 160 nF		0.60 %	
	160 nF to 1.6 µF		1.5 %	
	1.6 pF to 160 pF	1 MHz to 5 MHz	5.0 %	
	160 pF to 1.6 nF		2.3 %	
	3.2 nF to 16 nF		3.0 %	
	Capacity			TP7, TP21, TP24
	Generation			
	1 pF, 10 pF, 100 pF, 1000 pF	1 kHz	0.003 %	
	10 nF	1 kHz	0.005 %	
	100 nF, 1 µF, 10 µF	1 kHz	0.01 %	
	100 µF	1 kHz	0.1 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
7*	10 pF	10 kHz, 100 kHz	0.02 %	TP7, TP21, TP24
	100 pF, 1000 pF	100 Hz, 10 kHz, 100 kHz	0.02 %	
	10 nF, 100 nF, 1 µF,	100 kHz	0.1 %	
	10 µF	100 kHz	0.1 %	
	100 µF	100 Hz, 10 kHz	0.1 %	
	10 pF, 10 nF	100 Hz to 20 kHz	0.03 %	
	100 nF	100 Hz to 20 kHz	0.05 %	
	1 µF, 10 µF	100 Hz to 20 kHz	0.1 %	
	1 pF	20 kHz to 1 MHz 1 MHz to 10 MHz	0.3 % 1.3 %	
	10 pF	20 kHz to 1 MHz 1 MHz to 10 MHz	0.1 % 0.25 %	
	100 pF	100 Hz to 1 MHz 1 MHz to 10 MHz	0.03 % 0.2 %	
	1 nF	100 Hz to 1 MHz 1 MHz to 10 MHz	0.03 % 0.86 %	
	Capacity			TP7, TP21, TP24
	Generation			
	220 pF to 1.1 nF	10 Hz to 10 kHz	0.5 % + 0.01 nF	
	1.1 nF to 3.3 nF	10 Hz to 3 kHz	0.5 % + 0.01 nF	
	3.3 nF to 11 nF	10 Hz to 1 kHz	0.25 % + 0.01 nF	
	11 nF to 110 nF	10 Hz to 1 kHz	0.25 % + 0.1 nF	
	110 nF to 330 nF	10 Hz to 1 kHz	0.25 % + 0.3 nF	
	0.33 µF to 1.1 µF	10 Hz to 600 Hz	0.25 % + 1 nF	
	1.1 µF to 3.3 µF	10 Hz to 300 kHz	0.25 % + 3 nF	
	3.3 µF to 11 µF	10 Hz to 150 Hz	0.25 % + 10 nF	
	11 µF to 33 µF	10 Hz to 120 Hz	0.40 % + 30 nF	
	33 µF to 110 µF	10 Hz to 80 Hz	0.45 % + 100 nF	
	110 µF to 330 µF	0 Hz to 50 Hz	0.45 % + 300 nF	
	0.33 mF to 1.1 mF	0 Hz to 20 Hz	0.45 % + 1 µF	
	1.1 mF to 3.3 mF	0 Hz to 6 Hz	0.45 % + 3 µF	
	3.3 mF to 11 mF	0 Hz to 2 Hz	0.45 % + 10 µF	
	11 mF to 33 mF	0 Hz to 0.6 Hz	0.75 % + 30 µF	
	33 mF to 110 mF	0 Hz to 0.2 Hz	1 % + 100 µF	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
8*	Loss factor of capacitor D Generation for (10, 100, 1000 and 10000) pF D < 0.001	1 kHz	Expressed as absolute value 0.00001	TP30, TP24
	0.001 ≤ D < 0.1	1 kHz	0.00002	
	0.1 ≤ D < 1	1 kHz	0.002	
	100 nF D < 0.1	1 kHz	0.00005	
	0.1 ≤ D < 1	1 kHz	0.0025	
	1000 nF D < 0.1	1 kHz	0.004	
9	Inductance Measurement 0.2 μH to 2 μH	10 kHz 20 kHz	1 % 2 %	TP8, TP21, TP24
	2 μH to 20 μH	10 kHz 20 kHz	0.5 % 1 %	
	20 μH to 200 μH	100 Hz to 400 Hz 1 kHz to 20 kHz	1 % 0.5 %	
	0.2 mH to 2 mH	100 Hz to 400 Hz 1 kHz to 10 kHz 20 kHz	0.5 % 0.1 % 0.5 %	
	2 mH to 20 mH	100 Hz to 10 kHz 20 kHz	0.1 % 0.5 %	
	20 mH to 200 mH	100 Hz to 4 kHz 10 kHz 20 kHz	0.1 % 0.5 % 1 %	
	0.2 H to 2 H	100 Hz to 400 Hz 1 kHz to 4 kHz	0.1 % 0.5 %	
	2 H to 20 H	100 Hz to 400 Hz	0.5 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
9	Inductance Measurement 1 µH	1 kHz	0.2 %	TP8, TP21, TP24
	10 µH	1 kHz	0.08 %	
	100 µH	1 kHz	0.03 %	
	1 mH, 10 mH, 100 mH 1 H, 10 H	1 kHz	0.03 %	
	100 H	1 kHz	0.05 %	
	1000 H	1 kHz	0.07 %	
	10 kH	1 kHz	0.1 %	
9*	Inductance Generation 1 µH	1 kHz	0.3 %	TP8, TP21, TP24
	10 µH	1 kHz	0.12 %	
	100 µH, 1 mH, 10 mH 100 mH	100 Hz, 1 kHz, 10 kHz	0.03 %	
	1 H, 10 H	100 Hz, 1 kHz	0.05 %	
	Simulated inductor 100 H	1 kHz	0.1 %	TP8, TP21, TP24
	1000 H, 2 kH	1 kHz	1.0 %	
	Inductance Measurement 160 µH to 1.6 mH 1.6 mH to 16 mH 16 mH to 160 mH 160 mH to 1.6 H 1.6 H to 1.6 kH 1.6 kH to 16 kH 16 kH to 38 kH	42 Hz to 100 Hz	6.0 % 1.1 % 0.60 % 0.53 % 0.66 % 1.3 % 6.6 %	TP8, TP21, TP24
	160 µH to 1.6 mH 1.6 mH to 16 mH 16 mH to 160 H 160 H to 160 H 160 H to 1.6 kH 1.6 kH to 16 kH	100 Hz to 1000 Hz	0.60 % 0.30 % 0.12 % 0.26 % 0.90 % 3.3 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
9*	1.6 µH to 16 µH	1 kHz to 10 kHz	4.5 %	TP8, TP21, TP24
	16 µH to 160 µH		0.60 %	
	160 µH to 1.6 mH		0.30 %	
	1.6 mH to 16 mH		0.12 %	
	16 mH to 1.6 H		0.26 %	
	1.6 H to 16 H		0.36 %	
	16 H to 160 H		0.90 %	
	160 H to 1.6 kHz		3.3 %	
	160 nH to 1.6 µH	10 kHz to 100 kHz	4.5 %	
	1.6 µH to 16 µH		0.60 %	
	16 µH to 1.6 mH		0.30 %	
	1.6 mH to 160 mH		0.57 %	
	160 mH to 1.6 H		1.0 %	
	1.6 H to 16 H		1.6 %	
	160 nH to 1.6 µH	100 kHz to 1 MHz	1.5 %	
	1.6 µH to 16 µH		0.60 %	
	16 µH to 160 µH		0.45 %	
	160 µH to 16 mH		0.86 %	
	16 mH to 160 mH		4.7 %	
10*	Non-linear distortion Generation U ≥ 200 mV pure harmonic signal	20 Hz to 20 kHz	0.003 %	TP10
		20 kHz to 50 kHz	0.03 %	
		50 kHz to 100 kHz	0.05 %	
	Non-linear distortion Measurement 0.01 % to 100 %			TP10
	BW 80 kHz	20 Hz to 20 kHz	1 dB + 0.01 % abs.	
	BW 500 kHz	20 kHz to 50 kHz	2 dB + 0.04 % abs.	
	BW 500 kHz	50 kHz to 100 kHz	2 dB + 0.06 % abs.	
11*	HF power Measurement 50 Ω -50 dBm to -20 dBm	10 MHz to 30 MHz	7.8 %	TP11, TP23
		30 MHz to 18 GHz	5.4 %	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
12*	-20 dBm to +17 dBm	10 MHz to 30 MHz	6.5 %	
		30 MHz to 18 GHz	4.4 %	
	+17 dBm to +40 dBm	10 MHz to 18 GHz	5.7 %	
	HF voltage Generation 50 Ω 5 mV to 5.5 V (peak-to-peak value)	50 kHz to 1 MHz	4.7 %	TP12, TP23
		1 MHz to 10 MHz	6.6 %	
		10 MHz to 30 MHz	9.0 %	
		30 MHz to 250 MHz	6.8 %	
	5 mV to 3.0 V (peak-to-peak value)	250 MHz to 300 MHz	6.8 %	TP12, TP23
	HF voltage Measurement 50 Ω 20 mV to 1.5 V	20 MHz to 30 MHz	6.9 %	
		30 MHz to 100 MHz	4.7 %	
		100 MHz to 1 GHz	4.3 %	
	40 mV to 100 mV	1 MHz to 10 MHz	3.2 %	
		10 MHz to 20 MHz	3.5 %	
	250 mV to 1 V	1 MHz to 10 MHz	2.5 %	
		10 MHz to 20 MHz	4.0 %	
	1.5 V to 3 V	1 MHz to 10 MHz	3.3 %	
13	HF attenuation Measurement comparative method 0 dB to 60 dB	1.2 GHz to 3.8 GHz	0.30 dB	TP13
		3.8 GHz to 8.2 GHz	0.40 dB	
		8.2 GHz to 18 GHz	0.70 dB	
	60 dB to 80 dB	1.2 GHz to 3.8 GHz	0.80 dB	
		3.8 GHz to 8.2 GHz	1.20 dB	
		8.2 GHz to 18 GHz	1.50 dB	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
13*	HF attenuation Measurement power method 0 dB to 30 dB	10 MHz to 30 MHz	0.50 dB	TP13
		30 MHz to 2 GHz	0.15 dB	
		2 GHz to 18 GHz	0.35 dB	
	30 dB to 50 dB	10 MHz to 30 MHz	0.70 dB	
		30 MHz to 2 GHz	0.35 dB	
		2 GHz to 18 GHz	0.50 dB	
14*	Depth of AM Measurement 5 % to 99 %	f_c : 150 kHz to 10 MHz f_{mod} : 50 Hz to 10 kHz	2.2 % rel.	TP14
	5 % to 99 %	f_c : 150 kHz to 10 MHz f_{mod} : 20 Hz to 50 Hz	3.2 % rel.	
	5 % to 99 %	f_c : 10 MHz to 1300 MHz f_{mod} : 50 Hz to 50 kHz	1.2 % rel.	
	5 % to 99 %	f_c : 10 MHz to 1300 MHz f_{mod} : 20 Hz to 50 Hz f_{mod} : 50 kHz to 100 kHz	3.2 % rel.	
15*	FM frequency deviation Measurement 90 Hz to 40 kHz	f_c : 250 kHz to 10 MHz f_{mod} : 20 Hz to 10 kHz	2.1 % + 1 Hz	TP15
	90 Hz to 400 kHz	f_c : 10 MHz to 1300 MHz f_{mod} : 50 Hz to 100 kHz	1.1 % + 1 Hz	
	90 Hz to 400 kHz	f_c : 10 MHz to 1300 MHz f_{mod} : 20 Hz to 50 Hz f_{mod} : 100 kHz to 200 kHz	5.1 % + 1 Hz	
16*	DC power Generation U: 0.33 V to 1 kV I: 0.33 mA to 11 A 1mW to 330 mW 330 mW to 11 kW 11 kW to 550 kW		0.04 % 0.12 % 0.6 %	TP16

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
16*	U: 25 V to 100 V I: 50 mA to 100 mA 1.25 W to 10 W		0.013 %	TP16
17*	A.C. power Generation U: 0.33 V to 1 kV I: 0.33 mA to 11 A PF = 1 1 mW to 11 kW 11 kW to 550 kW	45 Hz to 65 Hz 45 Hz to 65 Hz	0.25 % 0.7 %	TP16
	U: 0.33 V to 1 kV I: 0.33 mA to 11 A			
	PF = 0.5 inductive and capacitive 1 mW to 11 kW	45 Hz to 65 Hz	0.8 %	
	U: 0.33 V to 1 kV I: 0.33 mA to 11 A PF = 1 1 mW to 11 kW	65 Hz to 500 Hz	1 %	
	U: 0.33 V to 1 kV I: 0.33 mA to 11 A PF = 0.5 inductive and capacitive 1 mW to 11 kW	65 Hz to 500 Hz	2.8 %	
18*	Direct current electrical work Generation U: 0.33 V to 1 kV I: 0.33 mA to 11 A t ≥ 600 s 0.9 Ws to 300 Ws 300 Ws to 9.9 MWs 9.9 MWs to 500 MWs		0.09 % 0.14 % 0.61 %	TP16

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
18*	U: 25 V to 100 V I: 50 mA to 100 mA t ≥ 600 s 0.9 Ws to 9 MWs		0.53 %	TP16
	I ₁ : 50mA to 100 mA I ₂ : 1 mA to 1 A t ≥ 600 s 9 MWs to 18 GWs		0.53 %	
19*	A.C. electrical work Generation U: 0.33 V to 1 kV I: 0.33 mA to 11 A t ≥ 600 s, PF = 1 0.9 Ws to 9.9 MWs	45 Hz to 65 Hz	0.09 %	TP16
	U: 0.33 V to 1 kV I: 0.33 mA to 11 A t ≥ 600 s, PF = 1 0.9 Ws to 9.9 MWs	45 Hz to 65 Hz	0.14 %	
	U: 0.33 V to 1 kV I: 0.33 mA to 11 A t ≥ 600 s, PF = 0.5 inductive and capacitive 0.9 Ws to 9.9 MWs	45 Hz to 65 Hz	0.61 %	
	U: 0.33 V to 1 kV I: 0.33 mA to 11 A t ≥ 600 s, PF = 1; PF = 0.5 inductive and capacitive 0.9 Ws to 9.9 MWs	65 Hz to 500 Hz	0.53 %	
20*	Reflection coefficient Measurement N connector 50 Ω 0.01 to 0.10 0.10 to 0.20 0.20 to 0.30 0.05 to 0.15	10 MHz to 2 GHz 10 MHz to 2 GHz 10 MHz to 2 GHz 2 GHz to 18 GHz	0.020 0.030 0.045 0.070	TP17

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
	0.15 to 0.20	2 GHz to 18 GHz	0.080	
	0.20 to 0.30	2 GHz to 18 GHz	0.10	
21*	Voltage above 1000 V Measurement 1000 V to 30000 V		0.21 %	TP18
	Generation > 1000 V to 30000 V		0.25 %	
	Measurement 1000 V to 25000 V	50 Hz	0.3 %	
	Generation > 1000 V to 4000 V	50 Hz	0.5 %	
	AC HV measurement 1000 V to 10000 V	Max. 75 MHz	3 dB	
22	Mains impedance Generation 0.16 Ω	50 Hz	0.015 Ω	TP6, TP27
	Generation 0.17 Ω to 1 Ω	50 Hz	0.5 % + 0.015 Ω	
	1 Ω to 10 Ω	50 Hz	0.3 % + 0.015 Ω	
	10 Ω to 2 kΩ	50 Hz	0.1 % + 0.015 Ω	
23*	Phase shift Measurement For $U_1 = U_2$ or for $10 \text{ mV} \leq U_1 \leq 30 \text{ V}$ and $1 \text{ V} \leq U_2 \leq 250 \text{ V}$ 0° to 360°	2 Hz to 200 kHz	1°	TP32
	Generation For $10 \text{ mV} \leq U_1 \leq 3 \text{ V}$ For $10 \text{ mV} \leq U_2 \leq 3 \text{ V}$ 0° to 360°	2 Hz to 200 kHz	1°	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
24*	Calibration of electrical part of temperature indicators and simulators Measurement and generation of equivalent direct current voltage for thermocouples, type: B 600 °C to 800 °C 800 °C to 1000 °C 1000 °C to 1550 °C 1550 °C to 1820 °C		0.44 °C 0.34 °C 0.30 °C 0.33 °C	TP1,TP21
	C 0 °C to 150 °C 150 °C to 650 °C 650 °C to 1000 °C 1000 °C to 1800 °C 1800 °C to 2316 °C		0.30 °C 0.26 °C 0.31 °C 0.50 °C 0.84 °C	
	E -250 °C to -100 °C -100 °C to -25 °C -25 °C to 350 °C 350 °C to 650 °C 650 °C to 1000 °C		0.50 °C 0.16 °C 0.14 °C 0.16 °C 0.21 °C	
	J -210 °C to -100 °C -100 °C to -30 °C -30 °C to 150 °C 150 °C to 760 °C 760 °C to 1200 °C		0.27 °C 0.16 °C 0.14 °C 0.17 °C 0.23 °C	
	K -200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C		0.33 °C 0.18 °C 0.16 °C 0.26 °C 0.40 °C	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
25*	0 °C to 100 °C		0.07 °C	TP5,TP21
	100 °C to 300 °C		0.09 °C	
	300 °C to 400 °C		0.10 °C	
	400 °C to 630 °C		0.12 °C	
	630 °C to 800 °C		0.23 °C	
	Pt 3926, 100 Ω			
	-200 °C to 0 °C		0.05 °C	
	0 °C to 100 °C		0.07 °C	
	100 °C to 300 °C		0.09 °C	
	300 °C to 400 °C		0.10 °C	
	400 °C to 630 °C		0.12 °C	
	Pt 3916, 100 Ω			
	-200 °C to -190 °C		0.25 °C	
	-190 °C to -80 °C		0.04 °C	
	-80 °C to 0 °C		0.05 °C	
	0 °C to 100 °C		0.06 °C	
	100 °C to 260 °C		0.07 °C	
	260 °C to 300 °C		0.08 °C	
	300 °C to 400 °C		0.09 °C	
	400 °C to 600 °C		0.10 °C	
	600 °C to 630 °C		0.23 °C	
	Pt 385, 200 Ω			
	-200 °C to 100 °C		0.04 °C	
	100 °C to 260 °C		0.05 °C	
	260 °C to 300 °C		0.12 °C	
	300 °C to 400 °C		0.13 °C	
	400 °C to 600 °C		0.14 °C	
	600 °C to 630 °C		0.16 °C	
	Pt 385, 500 Ω			
	-200 °C to -80 °C		0.04 °C	
	-80 °C to 100 °C		0.05 °C	
	100 °C to 260 °C		0.06 °C	
	260 °C to 400 °C		0.08 °C	
	400 °C to 600 °C		0.09 °C	
	600 °C to 630 °C		0.11 °C	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
25*	Pt 385, 1000 Ω -200 °C to 0 °C 0 °C to 100 °C 100 °C to 260 °C 260 °C to 300 °C 300 °C to 600 °C 600 °C to 630 °C		0.03 °C 0.04 °C 0.05 °C 0.06 °C 0.07 °C 0.23 °C	TP5, TP21
	Cu 427, 10 Ω -100 °C to 260 °C		0.30 °C	
	Calibration of electrical part of temperature simulation systems using resistance temperature sensors Measurement of equivalent resistance for RTD, type: PT25 -200 °C to 630 °C		0.02 °C	
	PT100 -200 °C to 630 °C		0.02 °C	

¹⁾ Asterisk at the ordinal number identifies the calibrations performed outside/also outside the laboratory premises.

²⁾ Expressed like uncertainty in accordance with the requirements of the document EA 4/02 at $k = 2$.

Measured instruments or devices:

(In accordance with the above list of measured quantities and the ranges of measurement the following types of instruments or devices can be measured.)

Ordinal number	Measured instrument/device type
1	D.C. analogue and digital voltmeters up to 6.5 digits, D.C. voltage calibrators and sources, electrical part of pH meters and pH simulators
2	A.C. analogue and digital voltmeters up to 6,5 digits, A.C. voltage calibrators and sources
3	D.C. analogue and digital ammeters up to 6.5 digits, D.C. calibrators, tong-test ammeters, D.C. current sources and transducers
4	A.C. analogue and digital ammeters up to 6.5 digits, A.C. calibrators, tong-test ammeters and A.C. current sources

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number	Measured instrument/device type
5	Ohmmeters, resistance bridges, compensators, decade resistance boxes, resistance standards, resistance calibrators, insulation and transition resistance testers
6	A.C. resistance and conductivity meters up to 13 MHz frequency, A.C. resistance and conductivity standards
7	Capacity meters, bridges, standards, decades
8	Inductance meters, standards, decades
9	Non-linear distortion meters and sine wave signal generators
10	HF meters and power sources and spectrum analyzers, radio communication testers (COM, CMS,CTS,...) and radio navigation testers, simulators and imitators (ATC, DMF, TACAN, VOR – ILS, TCAS), testing and inspection equipment and their parts
11	HF voltage meters and sources
12	HF resistive attenuators and attenuation meters, radio communication testers (COM, CMS,CTS,...) and radio navigation testers, simulators and imitators (ATC, DMF, TACAN, VOR – ILS, TCAS), testing and inspection equipment and their parts
13	Amplitude modulated signal generators, amplitude modulation meters, radio communication testers (COM, CMS,CTS,...) and radio navigation testers, simulators and imitators (ATC, DMF, TACAN, VOR – ILS, TCAS), testing and inspection equipment and their parts
14	Frequency modulated signal generators, frequency modulation meters, radio communication testers (COM, CMS,CTS,...) and radio navigation testers, simulators and imitators (ATC, DMF, TACAN, VOR – ILS, TCAS), testing and inspection equipment and their parts
15	Tong type, D.C. and A.C. wattmeters,
16	Measurement of impedance matching to 50 Ω impedance
17	D.C. and A.C. high voltage power sources, D.C. high voltage meters and high voltage voltage/current transducers
18	Time pulse meter
19	Oscilloscopes, transient recorders
20	Instruments for the inspection of electrical installations and resistance-based impedance standards
21	Instruments for the measurement of capacity loss factor D
22	Phase shifted voltage signal sources, phase shifted voltage signal meters
23	Electrical parts of temperature meters and simulators using thermocouples
24	Electrical parts of temperature meters and simulators using RTD

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Field of measured quantity: **frequency, time**

Calibration: Nominal calibration temperature: $(23 \pm 2) ^\circ\text{C}$

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [$\pm$] ²⁾	Calibration procedure identification
1*	Frequency Measurement 0.005 Hz to 400 kHz 400 kHz to 1.5 GHz 1.5 GHz to 18 GHz		$3.3 \cdot 10^{-4} / f^{3)}$ $1.2 \cdot 10^{-9}$ $3 \cdot 10^{-9}$	TP9, TP22
	Frequency 1 MHz, 5 MHz, 10 MHz $1 \text{ s} \leq \tau \leq 2\,000 \text{ s}$ $2\,000 \text{ s} \leq \tau \leq 15\,000 \text{ s}$ $\tau \leq 15\,000 \text{ s}$		$3 \cdot 10^{-10}$ $1 \cdot 10^{-11}$ $5 \cdot 10^{-12}$	
	Frequency Generation of reference signal 10 MHz 1 Hz to 18 GHz		$1 \cdot 10^{-11}$ $1 \cdot 10^{-9}$	
2*	Time stamps Generation 1.8 ns to 2.2 ns 4.5 ns to 11 ns 18 ns to 22 ns 45 ns to 60 s		0.005 % 0.005 % 0.005 % 0.005 %	TP23
	Time stamps Generation 1 s		$1 \cdot 10^{-9}$	
	Rise time $\leq 1 \text{ ns}$ Generation 4.5 mV to 5.5 mV	1 kHz to 1 MHz	0.3 ns	
	9 mV to 11 mV 22.5 mV to 27.5 mV 45 mV to 55 mV (peak-to-peak value)	1 kHz to 1 MHz	0.2 ns	

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Ordinal number ¹⁾	Measured quantity and range of measurement	Frequency	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
2*	90 mV to 110 mV 225 mV to 275 mV 450 mV to 550 mV 0.9 V to 1.1 V 2.25 V to 2.75 V (peak-to-peak value)	1 kHz to 1 MHz	0.2 ns	TP23
3*	Time – chronometers Measurement – electronically switched 1 s to 100,000 s Measurement – (manually switched) 1 s to 90,000 s		0.01 s 0.20 s	TP33

¹⁾ Asterisk at the ordinal number identifies the calibrations performed outside/also outside the laboratory premises.

²⁾ Expressed like uncertainty in accordance with the requirements of the document EA 4/02 at $k = 2$.

³⁾ measured frequency in Hz

Measured instruments or devices:

(In accordance with the above list of measured quantities and the ranges of measurement the following types of instruments or devices can be measured.)

Ordinal number	Measured instrument/device type
1	LF and HF counters, LF and HF generators, variable speed drives, frequency standards, frequency comparators and pulse generators
2	Oscilloscopes, transient recorders
3	Stopwatches, timers and chronometers

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Field of measured quantity: pressure

Calibration:

Nominal calibration temperature: $(20 \pm 2) ^\circ\text{C}$

Ordinal number ¹⁾	Measured quantity	Measured quantity range	Calibration and Measurement Capability [$\pm$] ²⁾	Calibration procedure identification
1*	Absolute pressure	(0 to 131) kPa (131 to 350) kPa	16 Pa 42 Pa	TP40, TP41, TP42, TP43
	Vacuum – gas	(0 to 2.5) kPa (2.5 to 35) kPa (35 to 100) kPa	2.4 Pa 32 Pa 45 Pa	TP40, TP41 TP42
	Overpressure – gas	(0 to 20) kPa (20 to 200) kPa (0.2 to 3.5) MPa (3.5 to 30) MPa	2.4 Pa 0.011 % 2.1 kPa 21 kPa	TP40, TP41, TP42
	Overpressure – liquids (water, alcohol, oil)	(0 to 2.5) kPa (2.5 to 35) kPa (35 to 100) kPa (0.1 to 1.2) MPa (1.2 to 12) MPa (12 to 35) MPa (35 to 70) MPa	2.4 Pa 32 Pa 2.1 kPa 0.22 kPa 0.018 % 21 kPa 40 kPa	TP40, TP41, TP42

¹⁾ Asterisk at the ordinal number identifies the calibrations performed outside/also outside the laboratory premises.

²⁾ Expressed like uncertainty in accordance with the requirements of the document EA 4/02 at $k = 2$.

Measured instruments or devices:

(In accordance with the above list of measured quantities and the ranges of measurement the following types of instruments or devices can be measured.)

Ordinal number	Measured instrument/device type
1	Deformation manometers, pointer manometers, pressure converters, dial manometers, digital manometers, pressure calibrators, barometers, aircraft pressure altimeters, pressure airspeed indicators and aircraft pressure calibrators

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Field of measured quantity: temperature

Calibration: Nominal calibration temperature – direct measurement: (20 ± 5) °C

Ordinal number ¹⁾	Measured quantity	Measured quantity range	Calibration and Measurement Capability [±] ²⁾	Calibration procedure identification
1*	Direct indicating thermometers and temperature measuring chains	-196 °C (-80 to -30) °C (-30 to 0) °C (0 to 50) °C (50 to 140) °C (140 to 300) °C (300 to 600) °C (600 to 1000) °C	0.32 °C 0.20 °C 0.07 °C 0.04 °C 0.06 °C 0.34 °C 0.62 °C 3.5 °C	TP44.1
2*	Resistance temperature sensors	-196 °C (-80 to -30) °C (-30 to 0) °C (0 to 50) °C (50 to 140) °C (140 to 300) °C (300 to 600) °C (600 to 1000) °C	0.3 °C 0.15 °C 0.07 °C 0.03 °C 0.06 °C 0.26 °C 0.5 °C 3.5 °C	TP44.2
3*	Thermoelectric temperature sensors	-196 °C (-80 to -30) °C (-30 to 0) °C 0 °C (0 to 140) °C (140 to 300) °C (300 to 600) °C (600 to 1000) °C	4.0 °C 1.5 °C 0.9 °C 0.7 °C 0.8 °C 1.5 °C 1.7 °C 3.5 °C	TP44.3
4*	Contactless thermometers and thermal cameras	(-25 °C to 0) °C (0 to 50) °C (50 to 100) °C (100 to 300) °C (300 to 500) °C	1.9 °C 1.3 °C 1.6 °C 2.0 °C 2.5 °C	TP44.4 TP44.5

¹⁾ Asterisk at the ordinal number identifies the calibrations performed outside/also outside the laboratory premises.

²⁾ Expressed like uncertainty in accordance with the requirements of the document EA 4/02 at k = 2.

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2018 of 10/07/2018**

Accredited entity according to ČSN EN ISO/IEC 17025:2005:

HES, s.r.o.
Calibration Laboratory
U Dráhy 11, 664 49 Ostopovice

Measured instruments or devices:

(In accordance with the above list of measured quantities and the ranges of measurement the following types of instruments or devices can be measured.)

Ordinal number	Measured instrument/device type
1	Temperature measuring chains, electronic thermometers, analogue thermometers, thermoelectric thermometers, thermocouples, resistance thermometers
2	Resistance temperature sensors
3	Thermoelectric temperature sensors
4	Contactless thermometers and contactless thermometer measuring chains, infrared thermometers, thermal cameras, thermal imaging camera

Field of measured quantity: **humidity**

Calibration: Nominal calibration temperature: $(20 \pm 5) ^\circ\text{C}$

Ordinal number ¹⁾	Measured quantity	Measured quantity range	Calibration and Measurement Capability [$\pm$] ²⁾	Calibration procedure identification
1*	Direct indicating hygrometers and humidity measuring chains Measured at temperatures $(15 \text{ to } 30) ^\circ\text{C}$ ⁴⁾	$(5 \text{ to } 10) \% \text{ RH}$ ³⁾ $(10 \text{ to } 50) \% \text{ RH}$ $(50 \text{ to } 70) \% \text{ RH}$ $(70 \text{ to } 90) \% \text{ RH}$	2.2 % RH 1.4 % RH 1.6 % RH 1.8 % RH	TP45

¹⁾ Asterisk at the ordinal number identifies the calibrations performed outside/also outside the laboratory premises.

²⁾ Expressed like uncertainty in accordance with the requirements of the document EA 4/02 at $k = 2$.

³⁾ RH – Relative Humidity

⁴⁾ The specified calibration values apply to the specified temperature range

Measured instruments or devices:

(In accordance with the above list of measured quantities and the ranges of measurement the following types of instruments or devices can be measured.)

Ordinal number	Measured instrument/device type
1	Relative humidity measuring chains, electronic analogue relative humidity hygrometers, electronic digital relative humidity hygrometers