

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

CMC for the field of measured quantity: Pressure

Ord. number ₁	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
1*	Deformation manometers, pointer manometers, digital manometers, pressure converters, pressure measuring chains, pressure calibrators, barometers, aircraft pressure altimeters, pressure airspeed indicators and aircraft pressure calibrators	1 kPa	to	30 kPa	gas	absolute pressure	7 Pa	Comparison with a standard digital manometer	TP40, TP41, TP42, TP43	
		30 kPa	to	107 kPa			8 Pa			
		107 kPa	to	130 kPa			12 Pa			
		130 kPa	to	173 kPa			16 Pa			
		173 kPa	to	225 kPa		20 Pa				
		225 kPa	to	270 kPa		24 Pa				
		270 kPa	to	350 kPa		32 Pa				
		0 kPa	to	2.5 kPa	gas	negative gauge pressure	2.4 Pa	Comparison with a standard digital manometer	TP40, TP41, TP42, TP43	
		2.5 kPa	to	35 kPa			32 Pa			
		35 kPa	to	100 kPa			45 Pa			
		0 kPa	to	20 kPa	gas	positive gauge pressure	2.4 Pa	Comparison with a standard piston manometer	TP40, TP41, TP42, TP43	
		20 kPa	to	200 kPa			0.011 %			
		0.2 MPa	to	0.7 MPa	gas	positive gauge pressure	0.17 kPa	Comparison with a standard digital manometer	TP40, TP41, TP42, TP43	
		0.7 MPa	to	3.5 MPa			0.43 kPa			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Workplace
		min	unit	max	unit					
		3.5 MPa	to	20 MPa			4.7 kPa			
		20 MPa	to	30 MPa			21 kPa			
		0 kPa	to	2.5 kPa	liquid (water, alcohol, oil)	positive gauge pressure	2.4 Pa	Comparison with a standard digital manometer	TP40, TP41, TP42, TP43	
		2.5 kPa	to	35 kPa			32 Pa			
		35 kPa	to	100 kPa			2.1 kPa			
		0.1 MPa	to	1.2 MPa	liquid (water, alcohol, oil)	positive gauge pressure	0.22 kPa	Comparison with a standard piston manometer	TP40, TP41, TP42, TP43	
		1.2 MPa	to	12 MPa			0.018 %			
		12 MPa	to	35 MPa	liquid (water, alcohol, oil)	positive gauge pressure	21 kPa	Comparison with a standard digital manometer	TP40, TP41, TP42, TP43	
		35 MPa	to	70 MPa			40 kPa			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

CMC for the field of measured quantity: Temperature

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work place
		min	unit	max	unit					
1*	Direct indicating thermometers and temperature measuring chains, electronic and analogue thermometers, resistance thermometers, thermoelectric thermometers and thermocouples	-196 °C					0.32 °C	Comparison with a standard resistance sensor in a nitrogen vessel	TP44.1	
		-80 °C	to	-30 °C		0.20 °C	Comparison with a standard resistance sensor in liquid bath.	TP44.1		
		-30 °C	to	0 °C		0.07 °C				
0 °C	to	50 °C		0.04 °C						
50 °C	to	140 °C		0.06 °C						
2*	Resistance temperature sensors, resistance thermometers	-196 °C					0.3 °C	Comparison with a standard resistance sensor in a nitrogen vessel	TP44.2	
		-80 °C	to	-30 °C		0.15 °C	Comparison with a standard resistance sensor in liquid bath.	TP44.2		
		-30 °C	to	0 °C		0.07 °C				
0 °C	to	50 °C		0.03 °C						
50 °C	to	140 °C		0.06 °C						
		140 °C	to	300 °C		0.26 °C	Comparison with a standard resistance sensor in a calibrating oven	TP44.2		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work place
		min	unit	max	unit					
		300 °C	to	600 °C		0.5 °C				
		600 °C	to	1,000 °C		3.5 °C				
3*	Thermoelectric cells, thermocouples			-196 °C		4.0 °C	Comparison with a standard resistance sensor in a nitrogen vessel	TP44.3		
		-80 °C	to	-30 °C		1.5 °C	Comparison with a standard resistance sensor in liquid bath.	TP44.3		
		-30 °C	to	0 °C		0.9 °C				
		0 °C	to	140 °C		0.7 °C				
		140 °C	to	300 °C		1.5 °C	Comparison with a standard resistance sensor in a calibrating oven	TP44.3		
		300 °C	to	600 °C		1.7 °C				
		600 °C	to	1,000 °C		3.5 °C				
4*	Non-contact thermometers and measuring chains of non-contact thermometers, thermal cameras, infrared thermometers	-25 °C	to	0 °C		1.9 °C	Comparison with a reference non-contact thermometer on a black body	TP44.4, TP44.5		
		0 °C	to	50 °C		1.3 °C				
		50 °C	to	100 °C		1.6 °C				
		100 °C	to	300 °C		2.0 °C				
		300 °C	to	500 °C		2.5 °C				
5*	Temperature / Calibration of electrical part of temperature simulators, electrical parts of temperature gauges using thermocouples: Type B	600 °C	to	800 °C		0.44 °C	Direct generation and measurement by a calibrator, of equivalent DC voltage	TP1, TP21		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work place
		min	unit	max	unit					
		800 °C	to	1,000 °C		0.34 °C				
		1,000 °C	to	1,550 °C		0.30 °C				
		1,550 °C	to	1,820 °C		0.33 °C				
		Type C	0 °C	to	150 °C					0.30 °C
			150 °C	to	650 °C					0.26 °C
			650 °C	to	1,000 °C					0.31 °C
			1,000 °C	to	1,800 °C					0.50 °C
			1,800 °C	to	2,316 °C					0.84 °C
		Type E	-250 °C	to	-100 °C					0.50 °C
			-100 °C	to	-25 °C					0.16 °C
			-25 °C	to	350 °C					0.14 °C
			350 °C	to	650 °C					0.16 °C
			650 °C	to	1,000 °C					0.21 °C
		Type J	-210 °C	to	-100 °C					0.27 °C
			-100 °C	to	-30 °C					0.16 °C
			-30 °C	to	150 °C					0.14 °C
			150 °C	to	760 °C					0.17 °C
			760 °C	to	1,200 °C					0.23 °C
		Type K	-200 °C	to	-100 °C					0.33 °C
			-100 °C	to	-25 °C					0.18 °C
-25 °C	to		120 °C		0.16 °C					
120 °C	to		1,000 °C		0.26 °C					
1,000 °C	to		1,372 °C		0.40 °C					
Type L	-200 °C	to	-100 °C		0.37 °C					
	-100 °C	to	800 °C		0.26 °C					
	800 °C	to	900 °C		0.17 °C					

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work place
		min	unit	max	unit					
	Type N	-200 °C	to	-100 °C		0.40 °C				
		-100 °C	to	-25 °C		0.22 °C				
		-25 °C	to	120 °C		0.19 °C				
		120 °C	to	410 °C		0.18 °C				
		410 °C	to	1,300 °C		0.27 °C				
	Type R	0 °C	to	250 °C		0.57 °C				
		250 °C	to	400 °C		0.35 °C				
		400 °C	to	1,000 °C		0.33 °C				
		1,000 °C	to	1,767 °C		0.40 °C				
	Type S	0 °C	to	250 °C		0.47 °C				
		250 °C	to	1,000 °C		0.36 °C				
		1,000 °C	to	1,400 °C		0.37 °C				
		1,400 °C	to	1,767 °C		0.46 °C				
	Type T	-250 °C	to	-150 °C		0.63 °C				
		-150 °C	to	0 °C		0.24 °C				
		0 °C	to	120 °C		0.16 °C				
		120 °C	to	400 °C		0.14 °C				
	Type U	-200 °C	to	0 °C		0.57 °C				
		0 °C	to	600 °C		0.27 °C				
6*	Temperature / Calibration of electric part of thermometers using RTD sensors RTD type Pt 385, 100 Ω	-200 °C	to	0 °C		0.05 °C		Direct generation by a calibrator, equivalent resistance generation	TP5, TP21	
		0 °C	to	100 °C		0.07 °C				
		100 °C	to	300 °C		0.09 °C				
		300 °C	to	400 °C		0.10 °C				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work place
		min	unit	max	unit					
	RTD type Pt 3926, 100 Ω	400 °C	to	630 °C		0.12 °C				
		630 °C	to	800 °C		0.23 °C				
		-200 °C	to	0 °C		0.05 °C				
		0 °C	to	100 °C		0.07 °C				
		100 °C	to	300 °C		0.09 °C				
		300 °C	to	400 °C		0.10 °C				
	RTD type Pt 3916, 100 Ω	400 °C	to	630 °C		0.12 °C				
		-200 °C	to	-190 °C		0.25 °C				
		-190 °C	to	-80 °C		0.04 °C				
		-80 °C	to	0 °C		0.05 °C				
		0 °C	to	100 °C		0.06 °C				
		100 °C	to	260 °C		0.07 °C				
	RTD type Pt 385, 200 Ω	260 °C	to	300 °C		0.08 °C				
		300 °C	to	400 °C		0.09 °C				
		400 °C	to	600 °C		0.10 °C				
		600 °C	to	630 °C		0.23 °C				
		-200 °C	to	100 °C		0.04 °C				
		100 °C	to	260 °C		0.05 °C				
RTD type Pt 385, 500 Ω	260 °C	to	300 °C		0.12 °C					
	300 °C	to	400 °C		0.13 °C					
	400 °C	to	600 °C		0.14 °C					
	600 °C	to	630 °C		0.16 °C					
	-200 °C	to	-80 °C		0.04 °C					
	-80 °C	to	100 °C		0.05 °C					
		100 °C	to	260 °C		0.06 °C				
		260 °C	to	400 °C		0.08 °C				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work place	
		min	unit	max	unit						
	RTD type Pt 385, 1,000 Ω	400 °C	to	600 °C		0.09 °C					
		600 °C	to	630 °C		0.11 °C					
		-200 °C	to	0 °C		0.03 °C					
		0 °C	to	100 °C		0.04 °C					
		100 °C	to	260 °C		0.05 °C					
		260 °C	to	300 °C		0.06 °C					
		300 °C	to	600 °C		0.07 °C					
	RTD type PtNi 385 120 Ω	600 °C	to	630 °C		0.23 °C					
		-80 °C	to	100 °C		0.08 °C					
		100 °C	to	260 °C		0.14 °C					
	Cu 427, 10 Ω	-100 °C	to	260 °C		0.30 °C					
	Temperature / Calibration of electric part of temperature simulators using RTD sensors									Direct measurement by a multimeter, equivalent resistance measurement	TP5, TP21
		PRT 25 Ω	-200 °C	to	0 °C						
			0 °C	to	660 °C		0.010 °C				
		PRT 100 Ω	-200 °C	to	0 °C		0.003 °C				
			0 °C	to	232 °C		0.005 °C				
		232 °C	to	400 °C		0.007 °C					
		400 °C	to	660 °C		0.50 °C					

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies..

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

CMC for the field of measured quantity: Electrical quantities

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Workplace
		min	unit	max	unit					
1*	DC voltage / DC voltage sources	0 mV	to	20 mV		0.20 μV	Direct measurement using a multimeter	TP1, TP21		
		20 mV	to	30 mV		0.00096 %				
		30 mV	to	40 mV		0.00079 %				
		40 mV	to	50 mV		0.00071 %				
		50 mV	to	90 mV		0.00066 %				
		90 mV	to	190 mV		0.00057 %				
		190 mV	to	300 mV		0.00051 %				
		300 mV	to	500 mV		0.00044 %				
		500 mV	to	1 V		0.00039 %				
		1 V	to	2 V		0.00035 %				
		2 V	to	3 V		0.00051 %				
		3 V	to	5 V		0.00044 %				
		5 V	to	10 V		0.00039 %				
		10 V	to	20 V		0.00035 %				
		20 V	to	30 V		0.00066 %				
		30 V	to	50 V		0.00059 %				
		50 V	to	100 V		0.00054 %				
100 V	to	200 V	0.00050 %							
200 V	to	250 V	0.00071 %							
250 V	to	400 V	0.00066 %							
400 V	to	600 V	0.00059 %							
600 V	to	1,050 V	0.00054 %							

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
	DC Voltage / DC voltage meters, electrical parts of pH meters	0 mV	to	20 mV		0.8 μV	Direct generation with a calibrator	TP1, TP21		
		20 mV	to	25 mV		0.0039 %				
		25 mV	to	30 mV		0.0033 %				
		30 mV	to	35 mV		0.0029 %				
		35 mV	to	45 mV		0.0026 %				
		45 mV	to	70 mV		0.0022 %				
		70 mV	to	100 mV		0.0017 %				
		100 mV	to	150 mV		0.0015 %				
		150 mV	to	300 mV		0.0013 %				
		300 mV	to	400 mV		0.0011 %				
		400 mV	to	500 mV		0.00097 %				
		500 mV	to	1 V		0.00092 %				
		1 V	to	1.5 V		0.00082 %				
		1.5 V	to	2.2 V		0.00079 %				
		2.2 V	to	3 V		0.00087 %				
		3 V	to	4.5 V		0.00083 %				
		4.5 V	to	9 V		0.00079 %				
		9 V	to	11 V		0.00075 %				
		11 V	to	22 V		0.00077 %				
		22 V	to	45 V		0.0012 %				
		45 V	to	60 V		0.0010 %				
		60 V	to	100 V		0.00095 %				
		100 V	to	150 V		0.00090 %				
		150 V	to	220 V		0.00087 %				
		220 V	to	350 V		0.0012 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place	
		min	unit						max
		350 V		to	1,100 V				
2*	AC voltage / AC voltage sources	1.2 mV		to	2 mV	10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	4.4 μV 2.1 μV 2.3 μV 3.8 μV 13 μV	Direct measurement using a multimeter	TP2, TP21
		2 mV		to	4 mV	10 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 50 kHz 30 kHz to 100 kHz	4.6 μV 2.5 μV 4.5 μV 6.3 μV 23 μV		
		4 mV		to	10 mV	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 50 kHz to 100 kHz	5.4 μV 5.2 μV 3.1 μV 5.1 μV 12 μV 28 μV 35 μV		
		10 mV		to	15 mV	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	6.0 μV 5.7 μV 3.6 μV 5.6 μV 13 μV 31 μV		
		15 mV		to	20 mV	10 Hz to 40 Hz 40 Hz to 100 Hz	6.7 μV 6.3 μV		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	4.2 μV 6.2 μV 11 μV 23 μV 82 μV 0.19 mV			
		20 mV	to	30 mV		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.034 % 0.032 % 0.021 % 0.031 % 0.053 % 0.12 % 0.41 % 1.3 %			
		30 mV	to	50 mV		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.027 % 0.025 % 0.018 % 0.024 % 0.047 % 0.11 % 0.39 % 1.3 %			
		50 mV	to	100 mV		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz	0.022 % 0.020 % 0.015 % 0.019 % 0.044 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.11 % 0.38 % 1.3 %			
		100 mV	to	150 mV		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.018 % 0.016 % 0.013 % 0.015 % 0.039 % 0.091 % 0.45 % 1.3 %			
		150 mV	to	200 mV		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.016 % 0.014 % 0.012 % 0.014 % 0.036 % 0.084 % 0.43 % 1.3 %			
		200 mV	to	300 mV		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.021 % 0.019 % 0.017 % 0.019 % 0.041 % 0.11 % 0.41 % 1.3 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		300 mV		to	500 mV	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.018 % 0.016 % 0.014 % 0.016 % 0.034 % 0.11 % 0.39 % 1.3 %			
		500 mV		to	1 V	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.015 % 0.013 % 0.011 % 0.013 % 0.029 % 0.091 % 0.38 % 1.2 %			
		1 V		to	1.5 V	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.013 % 0.011 % 0.0086 % 0.011 % 0.025 % 0.071 % 0.45 % 1.3 %			
		1.5 V		to	2 V	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz	0.012 % 0.0099 % 0.0079 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.010 % 0.024 % 0.065 % 0.43 % 1.3 %			
		2 V	to	3 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.021 % 0.019 % 0.017 % 0.019 % 0.041 % 0.11 % 0.41 % 1.3 %			
		3 V	to	5 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.018 % 0.016 % 0.014 % 0.016 % 0.034 % 0.11 % 0.39 % 1.3 %			
		5 V	to	10 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.015 % 0.013 % 0.011 % 0.013 % 0.029 % 0.091 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						100 kHz to 300 kHz 300 kHz to 1 MHz	0.38 % 1.2 %			
		10 V	to	15 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.013 % 0.011 % 0.0086 % 0.011 % 0.025 % 0.071 % 0.51 % 3.1 %			
		15 V	to	20 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.012 % 0.0099 % 0.0079 % 0.010 % 0.024 % 0.065 % 0.44 % 2.4 %			
		20 V	to	30 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.021 % 0.019 % 0.017 % 0.019 % 0.041 % 0.16 %			
		30 V	to	50 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz	0.018 % 0.016 % 0.014 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.016 % 0.034 % 0.12 %			
		50 V	to	100 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.015 % 0.013 % 0.011 % 0.013 % 0.029 % 0.091 %			
		100 V	to	150 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.013 % 0.011 % 0.0086 % 0.011 % 0.025 % 0.071 %			
		150 V	to	200 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.012 % 0.0099 % 0.0079 % 0.010 % 0.024 % 0.065 %			
		200 V	to	300 V		40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.020 % 0.041 % 0.11 %			
		300 V	to	400 V		40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.017 % 0.034 % 0.077 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit					
		400 V	to	500 V	40 Hz to 10 kHz	0.016 %		
					10 kHz to 20 kHz	0.035 %		
					20 kHz to 30 kHz	0.039 %		
					30 kHz to 100 kHz	0.063 %		
		500 V	to	600 V	40 Hz to 10 kHz	0.017 %		
					10 kHz to 20 kHz	0.040 %		
					20 kHz to 30 kHz	0.049 %		
			30 kHz to 100 kHz	0.065 %				
		600 V	to	700 V	40 Hz to 10 kHz	0.019 %		
					10 kHz to 20 kHz	0.049 %		
					20 kHz to 30 kHz	0.065 %		
					30 kHz to 100 kHz	0.077 %		
		700 V	to	800 V	40 Hz to 10 kHz	0.023 %		
					10 kHz to 20 kHz	0.061 %		
					20 kHz to 30 kHz	0.086 %		
					30 kHz to 100 kHz	0.095 %		
		800 V	to	900 V	40 Hz to 10 kHz	0.027 %		
					10 kHz to 20 kHz	0.076 %		
					20 kHz to 100 kHz	0.12 %		
		900 V	to	1,000 V	40 Hz to 10 kHz	0.032 %		
					10 kHz to 20 kHz	0.094 %		
					20 kHz to 100 kHz	0.15 %		
		1,000 V	to	1,050 V	40 Hz to 10 kHz	0.034 %		
					10 kHz to 20 kHz	0.11 %		
					20 kHz to 100 kHz	0.16 %		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	12 μV 18 μV 35 μV 43 μV			
		5 mV	to	10 mV		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	12 μV 8.0 μV 7.1 μV 9.4 μV 17 μV 25 μV 44 μV 61 μV			
		10 mV	to	15 mV		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	14 μV 8.8 μV 7.3 μV 12 μV 21 μV 31 μV 54 μV 78 μV			
		15 mV	to	22 mV		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz	18 μV 10 μV 7.7 μV 14 μV 27 μV 39 μV 67 μV			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						500 kHz to 1 MHz	0.11 mV			
		22 mV	to	30 mV		10 Hz to 20 Hz	0.12 %			
						20 Hz to 40 Hz	0.059 %			
						40 Hz to 20 kHz	0.048 %			
						20 kHz to 50 kHz	0.072 %			
						50 kHz to 100 kHz	0.21 %			
						100 kHz to 300 kHz	0.24 %			
						300 kHz to 500 kHz	0.35 %			
						500 kHz to 1 MHz	0.72 %			
		30 mV	to	50 mV		10 Hz to 20 Hz	0.099 %			
						20 Hz to 40 Hz	0.049 %			
						40 Hz to 20 kHz	0.039 %			
						20 kHz to 50 kHz	0.062 %			
						50 kHz to 100 kHz	0.18 %			
						100 kHz to 300 kHz	0.21 %			
						300 kHz to 500 kHz	0.31 %			
						500 kHz to 1 MHz	0.62 %			
		50 mV	to	100 mV		10 Hz to 20 Hz	0.082 %			
						20 Hz to 40 Hz	0.039 %			
						40 Hz to 20 kHz	0.029 %			
						20 kHz to 50 kHz	0.052 %			
						50 kHz to 100 kHz	0.15 %			
						100 kHz to 300 kHz	0.18 %			
						300 kHz to 500 kHz	0.27 %			
						500 kHz to 1 MHz	0.51 %			
		100 mV	to	150 mV		10 Hz to 20 Hz	0.069 %			
						20 Hz to 40 Hz	0.031 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.021 % 0.045 % 0.12 % 0.15 % 0.23 % 0.44 %			
		150 mV	to	220 mV		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.065 % 0.028 % 0.017 % 0.039 % 0.11 % 0.13 % 0.21 % 0.41 %			
		220 mV	to	300 mV		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.087 % 0.028 % 0.011 % 0.020 % 0.058 % 0.11 % 0.28 % 0.61 %			
		300 mV	to	500 mV		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.078 % 0.025 % 0.010 % 0.018 % 0.049 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.088 % 0.23 % 0.51 %			
		500 mV	to	1 V		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.067 % 0.022 % 0.0092 % 0.016 % 0.040 % 0.071 % 0.19 % 0.40 %			
		1 V	to	1.5 V		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.059 % 0.019 % 0.0086 % 0.015 % 0.033 % 0.058 % 0.16 % 0.32 %			
		1.5 V	to	2.2 V		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.056 % 0.019 % 0.0084 % 0.014 % 0.030 % 0.053 % 0.15 % 0.29 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		2.2 V to 3 V				10 Hz to 20 Hz	0.087 %			
						20 Hz to 40 Hz	0.028 %			
						40 Hz to 20 kHz	0.011 %			
						20 kHz to 50 kHz	0.020 %			
						50 kHz to 100 kHz	0.042 %			
						100 kHz to 300 kHz	0.12 %			
						300 kHz to 500 kHz	0.33 %			
						500 kHz to 1 MHz	0.67 %			
		3 V to 5 V				10 Hz to 20 Hz	0.078 %			
						20 Hz to 40 Hz	0.025 %			
						40 Hz to 20 kHz	0.099 %			
						20 kHz to 50 kHz	0.018 %			
						50 kHz to 100 kHz	0.030 %			
						100 kHz to 300 kHz	0.11 %			
						300 kHz to 500 kHz	0.28 %			
						500 kHz to 1 MHz	0.56 %			
5 V to 10 V				10 Hz to 20 Hz	0.067 %					
				20 Hz to 40 Hz	0.022 %					
				40 Hz to 20 kHz	0.0091 %					
				20 kHz to 50 kHz	0.016 %					
				50 kHz to 100 kHz	0.033 %					
				100 kHz to 300 kHz	0.081 %					
				300 kHz to 500 kHz	0.23 %					
				500 kHz to 1 MHz	0.45 %					
10 V to 15 V				10 Hz to 20 Hz	0.059 %					
				20 Hz to 40 Hz	0.019 %					
				40 Hz to 20 kHz	0.0086 %					

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.014 % 0.029 % 0.066 % 0.19 % 0.37 %			
		15 V	to	22 V		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.056 % 0.019 % 0.087 % 0.014 % 0.029 % 0.061 % 0.18 % 0.34 %			
		22 V	to	30 V		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.087 % 0.028 % 0.013 % 0.039 % 0.087 %			
		30 V	to	50 V		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.078 % 0.025 % 0.012 % 0.034 % 0.078 %			
		50 V	to	100 V		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz	0.067 % 0.022 % 0.011 % 0.030 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						50 kHz to 100 kHz	0.067 %			
		100 V	to	150 V		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.059 % 0.019 % 0.010 % 0.026 % 0.059 %			
		150 V	to	220 V		10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.057 % 0.019 % 0.011 % 0.027 % 0.065 %			
		220 V	to	300 V		10 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	0.030 % 0.012 % 0.026 % 0.042 %			
		300 V	to	400 V		10 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	0.030 % 0.012 % 0.026 % 0.042 %			
		400 V	to	500 V		10 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	0.028 % 0.012 % 0.025 % 0.045 %			
		500 V	to	600 V		10 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	0.027 % 0.011 % 0.023 % 0.053 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		600 V	to	700 V		10 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	0.028 % 0.012 % 0.025 % 0.068 %			
		700 V	to	800 V		10 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	0.030 % 0.011 % 0.028 % 0.089 %			
		800 V	to	1,050 V		10 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	0.040 % 0.010 % 0.038 % 0.17 %			
3*	DC current / DC current sources	0 nA	to	1 μA			0.42 nA	Direct measurement with a multimeter or indirect measurement with a current shunt	TP3, TP21	
		1 μA	to	5 μA			0.47 nA			
		5 μA	to	10 μA			0.53 nA			
		10 μA	to	15 μA			0.59 nA			
		15 μA	to	20 μA			0.65 nA			
		20 μA	to	30 μA			0.0033 %			
		30 μA	to	50 μA			0.0026 %			
		50 μA	to	60 μA			0.0021 %			
		60 μA	to	90 μA			0.0019 %			
		90 μA	to	140 μA			0.0017 %			
		140 μA	to	200 μA			0.0015 %			
		200 μA	to	250 μA			0.0033 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		250 µA	to	300 µA		0.0029 %				
		300 µA	to	400 µA		0.0026 %				
		400 µA	to	500 µA		0.0023 %				
		500 µA	to	600 µA		0.0021 %				
		600 µA	to	900 µA		0.0019 %				
		900 µA	to	1.1 mA		0.0017 %				
		1.1 mA	to	2 mA		0.0016 %				
		2 mA	to	2.5 mA		0.0034 %				
		2.5 mA	to	3 mA		0.0030 %				
		3 mA	to	4 mA		0.0027 %				
		4 mA	to	5 mA		0.0024 %				
		5 mA	to	6 mA		0.0022 %				
		6 mA	to	9 mA		0.0020 %				
		9 mA	to	11 mA		0.0018 %				
		11 mA	to	20 mA		0.0017 %				
		20 mA	to	25 mA		0.0077 %				
		25 mA	to	30 mA		0.0069 %				
		30 mA	to	40 mA		0.0064 %				
		40 mA	to	50 mA		0.0057 %				
		50 mA	to	60 mA		0.0053 %				
		60 mA	to	70 mA		0.0050 %				
		70 mA	to	90 mA		0.0048 %				
		90 mA	to	100 mA		0.0046 %				
		100 mA	to	130 mA		0.0045 %				
		130 mA	to	190 mA		0.0043 %				
		190 mA	to	200 mA		0.0041 %				
		200 mA	to	250 mA		0.026 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		250 mA	to	300 mA		0.024 %				
		300 mA	to	400 mA		0.023 %				
		400 mA	to	600 mA		0.022 %				
		600 mA	to	900 mA		0.020 %				
		900 mA	to	1.7 A		0.019 %				
		1.7 A	to	2 A		0.018 %				
		2 A	to	60 A		0.020 %				
		60 A	to	200 A		0.040 %				
		200 A	to	300 A		0.2 %				
		300 A	to	600 A		0.3 %				
	DC current / DC current meters	0 μA	to	20 μA		10 nA	Direct generation with a calibrator	TP3, TP21		
		20 μA	to	30 μA		0.046 %				
		30 μA	to	40 μA		0.032 %				
		40 μA	to	50 μA		0.026 %				
		50 μA	to	60 μA		0.022 %				
		60 μA	to	70 μA		0.019 %				
		70 μA	to	95 μA		0.017 %				
		95 μA	to	120 μA		0.014 %				
		120 μA	to	170 μA		0.012 %				
		170 μA	to	200 μA		0.0099 %				
		200 μA	to	220 μA		0.0092 %				
		220 μA	to	300 μA		0.0088 %				
		300 μA	to	400 μA		0.0078 %				
		400 μA	to	500 μA		0.0071 %				
		500 μA	to	600 μA		0.0067 %				
		600 μA	to	700 μA		0.0065 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		700 μA	to	800 μA		0.0063 %				
		800 μA	to	1 mA		0.0061 %				
		1 mA	to	1.3 mA		0.0059 %				
		1.3 mA	to	1.9 mA		0.0057 %				
		1.9 mA	to	2.2 mA		0.0055 %				
		2.2 mA	to	3 mA		0.0088 %				
		3 mA	to	4 mA		0.0078 %				
		4 mA	to	5 mA		0.0071 %				
		5 mA	to	6 mA		0.0067 %				
		6 mA	to	7 mA		0.0065 %				
		7 mA	to	8 mA		0.0063 %				
		8 mA	to	10 mA		0.0061 %				
		10 mA	to	13 mA		0.0059 %				
		13 mA	to	19 mA		0.0057 %				
		19 mA	to	22 mA		0.0055 %				
		22 mA	to	30 mA		0.0098 %				
		30 mA	to	40 mA		0.0088 %				
		40 mA	to	50 mA		0.0082 %				
		50 mA	to	60 mA		0.0077 %				
		60 mA	to	70 mA		0.0075 %				
		70 mA	to	80 mA		0.0073 %				
		80 mA	to	180 mA		0.0072 %				
		180 mA	to	220 mA		0.0075 %				
		220 mA	to	300 mA		0.020 %				
		300 mA	to	400 mA		0.017 %				
		400 mA	to	600 mA		0.015 %				
		600 mA	to	900 mA		0.013 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		900 mA	to	1 A		0.011 %				
		1 A	to	1.4 A		0.012 %				
		1.4 A	to	1.8 A		0.013 %				
		1.8 A	to	2 A		0.014 %				
		2 A	to	2.2 A		0.015 %				
		2.2 A	to	3 A		0.044 %				
		3 A	to	4 A		0.041 %				
		4 A	to	5 A		0.038 %				
		5 A	to	6 A		0.037 %				
		6 A	to	7 A		0.036 %				
		7 A	to	8 A		0.035 %				
		8 A	to	11 A		0.034 %				
		11 A	to	20.2 A		0.033 %				
		20.2 A	to	23 A		0.043 %				
		23 A	to	30 A		0.042 %				
		30 A	to	1,000 A		0.4 %	Indirect generation with a current coil calibrator	TP3, TP21		
4*	AC current / AC current sources	10 μA	to	20 μA	10 Hz to 10 kHz	30 nA	Direct measurement with a multimeter or indirect measurement with a current shunt, current transformer	TP4, TP21		
		20 μA	to	30 μA	10 Hz to 10 kHz	0.15 %				
		30 μA	to	40 μA	10 Hz to 10 kHz	0.12 %				
		40 μA	to	50 μA	10 Hz to 10 kHz	0.098 %				
		50 μA	to	60 μA	10 Hz to 10 kHz	0.088 %				
		60 μA	to	70 μA	10 Hz to 10 kHz	0.081 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		70 µA	to	80 µA	10 Hz to 10 kHz	0.077 %				
		80 µA	to	90 mA	10 Hz to 10 kHz	0.073 %				
		90 µA	to	100 µA	10 Hz to 10 kHz	0.070 %				
		100 µA	to	110 µA	10 Hz to 10 kHz	0.068 %				
		110 µA	to	130 µA	10 Hz to 10 kHz	0.066 %				
		130 µA	to	150 µA	10 Hz to 10 kHz	0.063 %				
		150 µA	to	180 µA	10 Hz to 10 kHz	0.061 %				
		180 µA	to	200 µA	10 Hz to 10 kHz	0.059 %				
		200 µA	to	250 µA	10 Hz to 10 kHz	0.13 %				
		250 µA	to	300 µA	10 Hz to 10 kHz	0.11 %				
		300 µA	to	400 µA	10 Hz to 10 kHz	0.095 %				
		400 µA	to	500 µA	10 Hz to 10 kHz	0.079 %				
		500 µA	to	600 µA	10 Hz to 10 kHz	0.069 %				
		600 µA	to	700 µA	10 Hz to 10 kHz	0.062 %				
		700 µA	to	800 µA	10 Hz to 10 kHz	0.057 %				
		800 µA	to	900 µA	10 Hz to 10 kHz	0.054 %				
		900 µA	to	1 mA	10 Hz to 10 kHz	0.051 %				
		1 mA	to	1.1 mA	10 Hz to 10 kHz	0.049 %				
		1.1 mA	to	1.2 mA	10 Hz to 10 kHz	0.047 %				
		1.2 mA	to	1.4 mA	10 Hz to 10 kHz	0.045 %				
		1.4 mA	to	1.6 mA	10 Hz to 10 kHz	0.043 %				
		1.6 mA	to	1.9 mA	10 Hz to 10 kHz	0.041 %				
		1.9 mA	to	2 mA	10 Hz to 10 kHz	0.039 %				
		2 mA	to	2.5 mA	10 Hz to 10 kHz	0.13 %				
		2.5 mA	to	3 mA	10 Hz to 10 kHz	0.11 %				
		3 mA	to	4 mA	10 Hz to 10 kHz	0.095 %				
		4 mA	to	5 mA	10 Hz to 10 kHz	0.079 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		5 mA	to	6 mA		10 Hz to 10 kHz	0.069 %			
		6 mA	to	7 mA		10 Hz to 10 kHz	0.062 %			
		7 mA	to	8 mA		10 Hz to 10 kHz	0.057 %			
		8 mA	to	9 mA		10 Hz to 10 kHz	0.054 %			
		9 mA	to	10 mA		10 Hz to 10 kHz	0.051 %			
		10 mA	to	11 mA		10 Hz to 10 kHz	0.049 %			
		11 mA	to	12 mA		10 Hz to 10 kHz	0.047 %			
		12 mA	to	14 mA		10 Hz to 10 kHz	0.045 %			
		14 mA	to	16 mA		10 Hz to 10 kHz	0.043 %			
		16 mA	to	19 mA		10 Hz to 10 kHz	0.041 %			
		19 mA	to	20 mA		10 Hz to 10 kHz	0.039 %			
		20 mA	to	25 mA		10 Hz to 10 kHz	0.13 %			
		25 mA	to	30 mA		10 Hz to 10 kHz	0.11 %			
		30 mA	to	40 mA		10 Hz to 10 kHz	0.092 %			
		40 mA	to	50 mA		10 Hz to 10 kHz	0.076 %			
		50 mA	to	60 mA		10 Hz to 10 kHz	0.066 %			
		60 mA	to	70 mA		10 Hz to 10 kHz	0.059 %			
		70 mA	to	80 mA		10 Hz to 10 kHz	0.054 %			
		80 mA	to	90 mA		10 Hz to 10 kHz	0.051 %			
		90 mA	to	100 mA		10 Hz to 10 kHz	0.048 %			
		100 mA	to	110 mA		10 Hz to 10 kHz	0.046 %			
		110 mA	to	120 mA		10 Hz to 10 kHz	0.044 %			
		120 mA	to	140 mA		10 Hz to 10 kHz	0.042 %			
		140 mA	to	160 mA		10 Hz to 10 kHz	0.040 %			
		160 mA	to	190 mA		10 Hz to 10 kHz	0.038 %			
		190 mA	to	200 mA		10 Hz to 10 kHz	0.036 %			
		200 mA	to	250 mA		10 Hz to 2 kHz	0.17 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		250 mA	to	300 mA		2 kHz to 10 kHz	0.18 %			
						10 Hz to 2 kHz	0.15 %			
		300 mA	to	400 mA		2 kHz to 10 kHz	0.16 %			
						10 Hz to 2 kHz	0.13 %			
		400 mA	to	500 mA		2 kHz to 10 kHz	0.14 %			
						10 Hz to 2 kHz	0.12 %			
		500 mA	to	600 mA		2 kHz to 10 kHz	0.13 %			
						10 Hz to 2 kHz	0.11 %			
		600 mA	to	700 mA		2 kHz to 10 kHz	0.12 %			
						10 Hz to 2 kHz	0.094 %			
		700 mA	to	800 mA		2 kHz to 10 kHz	0.11 %			
						10 Hz to 2 kHz	0.089 %			
		800 mA	to	900 mA		2 kHz to 10 kHz	0.099 %			
						10 Hz to 2 kHz	0.086 %			
		900 mA	to	1 A		2 kHz to 10 kHz	0.096 %			
						10 Hz to 2 kHz	0.083 %			
		1 A	to	1.1 A		2 kHz to 10 kHz	0.093 %			
						10 Hz to 2 kHz	0.081 %			
		1.1 A	to	1.2 A		2 kHz to 10 kHz	0.091 %			
						10 Hz to 2 kHz	0.079 %			
		1.2 A	to	1.3 A		2 kHz to 10 kHz	0.089 %			
						10 Hz to 2 kHz	0.077 %			
		1.3 A	to	1.4 A		2 kHz to 10 kHz	0.087 %			
						10 Hz to 2 kHz	0.076 %			
		1.4 A	to	1.5 A		2 kHz to 10 kHz	0.087 %			
						10 Hz to 2 kHz	0.076 %			
						2 kHz to 10 kHz	0.085 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		1.5 A	to	1.6 A		10 Hz to 2 kHz	0.074 %			
						2 kHz to 10 kHz	0.085 %			
		1.6 A	to	1.7 A		10 Hz to 2 kHz	0.074 %			
						2 kHz to 10 kHz	0.083 %			
		1.7 A	to	2 A		10 Hz to 2 kHz	0.072 %			
						2 kHz to 10 kHz	0.083 %			
		2 A	to	2.5 A		10 Hz to 2 kHz	0.19 %			
						2 kHz to 10 kHz	0.36 %			
		2.5 A	to	3 A		10 Hz to 2 kHz	0.16 %			
						2 kHz to 10 kHz	0.34 %			
		3 A	to	4 A		10 Hz to 2 kHz	0.15 %			
						2 kHz to 10 kHz	0.32 %			
		4 A	to	5 A		10 Hz to 2 kHz	0.14 %			
						2 kHz to 10 kHz	0.31 %			
		5 A	to	6 A		10 Hz to 2 kHz	0.13 %			
						2 kHz to 10 kHz	0.30 %			
		6 A	to	7 A		10 Hz to 2 kHz	0.12 %			
						2 kHz to 10 kHz	0.29 %			
		7 A	to	11 A		10 Hz to 2 kHz	0.11 %			
						2 kHz to 10 kHz	0.28 %			
		11 A	to	20 A		10 Hz to 2 kHz	0.10 %			
						2 kHz to 10 kHz	0.27 %			
		2 A	to	4 A		50 Hz	0.12 %			
		4 A	to	10 A		50 Hz	0.05 %			
		10 A	to	1,200 A		50 Hz	0.04 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
	AC current / AC current meters	10 μA	to	20 μA	10 Hz to 20 Hz	40 nA	Direct generation with a calibrator	TP4, TP21		
20 Hz to 40 Hz					28 nA					
40 Hz to 1 kHz					18 nA					
1 kHz to 5 kHz					47 nA					
5 kHz to 10 kHz					97 nA					
20 μA		to	30 μA	10 Hz to 20 Hz	0.20 %					
				20 Hz to 40 Hz	0.14 %					
				40 Hz to 1 kHz	0.095 %					
				1 kHz to 5 kHz	0.26 %					
				5 kHz to 10 kHz	0.57 %					
30 μA		to	40 μA	10 Hz to 20 Hz	0.16 %					
				20 Hz to 40 Hz	0.11 %					
				40 Hz to 1 kHz	0.069 %					
				1 kHz to 5 kHz	0.20 %					
				5 kHz to 10 kHz	0.43 %					
40 μA		to	50 μA	10 Hz to 20 Hz	0.14 %					
				20 Hz to 40 Hz	0.086 %					
				40 Hz to 1 kHz	0.055 %					
				1 kHz to 5 kHz	0.17 %					
				5 kHz to 10 kHz	0.37 %					
50 μA	to	60 μA	10 Hz to 20 Hz	0.13 %						
			20 Hz to 40 Hz	0.076 %						
			40 Hz to 1 kHz	0.048 %						
			1 kHz to 5 kHz	0.15 %						
			5 kHz to 10 kHz	0.33 %						
60 μA	to	80 μA	10 Hz to 20 Hz	0.12 %						

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						20 Hz to 40 Hz	0.070 %			
						40 Hz to 1 kHz	0.042 %			
						1 kHz to 5 kHz	0.13 %			
						5 kHz to 10 kHz	0.30 %			
		80 μA	to	100 μA		10 Hz to 20 Hz	0.099 %			
						20 Hz to 40 Hz	0.061 %			
						40 Hz to 1 kHz	0.036 %			
						1 kHz to 5 kHz	0.12 %			
						5 kHz to 10 kHz	0.27 %			
		100 μA	to	130 μA		10 Hz to 20 Hz	0.096 %			
						20 Hz to 40 Hz	0.056 %			
						40 Hz to 1 kHz	0.032 %			
						1 kHz to 5 kHz	0.11 %			
						5 kHz to 10 kHz	0.25 %			
		130 μA	to	180 μA		10 Hz to 20 Hz	0.090 %			
						20 Hz to 40 Hz	0.052 %			
						40 Hz to 1 kHz	0.029 %			
						1 kHz to 5 kHz	0.092 %			
						5 kHz to 10 kHz	0.23 %			
		180 μA	to	220 μA		10 Hz to 20 Hz	0.085 %			
						20 Hz to 40 Hz	0.048 %			
						40 Hz to 1 kHz	0.025 %			
						1 kHz to 5 kHz	0.084 %			
						5 kHz to 10 kHz	0.21 %			
		220 μA	to	300 μA		10 Hz to 20 Hz	0.089 %			
						20 Hz to 40 Hz	0.052 %			
						40 Hz to 1 kHz	0.032 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		300 μA	to	400 μA	1 kHz to 5 kHz	0.25 %				
					5 kHz to 10 kHz	0.53 %				
					10 Hz to 20 Hz	0.084 %				
					20 Hz to 40 Hz	0.048 %				
					40 Hz to 1 kHz	0.028 %				
					1 kHz to 5 kHz	0.20 %				
		400 μA	to	500 μA	5 kHz to 10 kHz	0.43 %				
					10 Hz to 20 Hz	0.081 %				
					20 Hz to 40 Hz	0.045 %				
					40 Hz to 1 kHz	0.025 %				
					1 kHz to 5 kHz	0.17 %				
					5 kHz to 10 kHz	0.37 %				
		500 μA	to	600 μA	10 Hz to 20 Hz	0.079 %				
					20 Hz to 40 Hz	0.044 %				
					40 Hz to 1 kHz	0.023 %				
					1 kHz to 5 kHz	0.15 %				
					5 kHz to 10 kHz	0.33 %				
		600 μA	to	800 μA	10 Hz to 20 Hz	0.078 %				
					20 Hz to 40 Hz	0.043 %				
					40 Hz to 1 kHz	0.022 %				
					1 kHz to 5 kHz	0.13 %				
					5 kHz to 10 kHz	0.30 %				
		800 μA	to	1 mA	10 Hz to 20 Hz	0.076 %				
					20 Hz to 40 Hz	0.041 %				
					40 Hz to 1 kHz	0.021 %				
					1 kHz to 5 kHz	0.12 %				
					5 kHz to 10 kHz	0.27 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit					
		1 mA	to	1.3 mA	10 Hz to 20 Hz	0.075 %		
					20 Hz to 40 Hz	0.040 %		
					40 Hz to 1 kHz	0.020 %		
					1 kHz to 5 kHz	0.11 %		
					5 kHz to 10 kHz	0.25 %		
		1.3 mA	to	1.8 mA	10 Hz to 20 Hz	0.074 %		
					20 Hz to 40 Hz	0.039 %		
					40 Hz to 1 kHz	0.019 %		
					1 kHz to 5 kHz	0.092 %		
					5 kHz to 10 kHz	0.23 %		
		1.8 mA	to	2.2 mA	10 Hz to 20 Hz	0.073 %		
					20 Hz to 40 Hz	0.074 %		
					40 Hz to 1 kHz	0.019 %		
					1 kHz to 5 kHz	0.084 %		
					5 kHz to 10 kHz	0.21 %		
		2.2 mA	to	3 mA	10 Hz to 20 Hz	0.089 %		
					20 Hz to 40 Hz	0.052 %		
					40 Hz to 1 kHz	0.032 %		
					1 kHz to 5 kHz	0.25 %		
					5 kHz to 10 kHz	0.53 %		
		3 mA	to	4 mA	10 Hz to 20 Hz	0.084 %		
					20 Hz to 40 Hz	0.048 %		
					40 Hz to 1 kHz	0.028 %		
					1 kHz to 5 kHz	0.20 %		
					5 kHz to 10 kHz	0.43 %		
		4 mA	to	5 mA	10 Hz to 20 Hz	0.081 %		
					20 Hz to 40 Hz	0.045 %		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						40 Hz to 1 kHz	0.025 %			
						1 kHz to 5 kHz	0.17 %			
						5 kHz to 10 kHz	0.37 %			
		5 mA	to	6 mA		10 Hz to 20 Hz	0.079 %			
						20 Hz to 40 Hz	0.043 %			
						40 Hz to 1 kHz	0.023 %			
						1 kHz to 5 kHz	0.15 %			
						5 kHz to 10 kHz	0.33 %			
		6 mA	to	8 mA		10 Hz to 20 Hz	0.078 %			
						20 Hz to 40 Hz	0.042 %			
						40 Hz to 1 kHz	0.022 %			
						1 kHz to 5 kHz	0.13 %			
						5 kHz to 10 kHz	0.30 %			
		8 mA	to	10 mA		10 Hz to 20 Hz	0.076 %			
						20 Hz to 40 Hz	0.041 %			
						40 Hz to 1 kHz	0.021 %			
						1 kHz to 5 kHz	0.12 %			
						5 kHz to 10 kHz	0.27 %			
		10 mA	to	13 mA		10 Hz to 20 Hz	0.075 %			
						20 Hz to 40 Hz	0.040 %			
						40 Hz to 1 kHz	0.020 %			
						1 kHz to 5 kHz	0.11 %			
						5 kHz to 10 kHz	0.25 %			
		13 mA	to	18 mA		10 Hz to 20 Hz	0.074 %			
						20 Hz to 40 Hz	0.039 %			
						40 Hz to 1 kHz	0.019 %			
						1 kHz to 5 kHz	0.092 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		18 mA	to	22 mA		5 kHz to 10 kHz	0.23 %			
						10 Hz to 20 Hz	0.073 %			
						20 Hz to 40 Hz	0.039 %			
						40 Hz to 1 kHz	0.019 %			
						1 kHz to 5 kHz	0.084 %			
		22 mA	to	30 mA		5 kHz to 10 kHz	0.21 %			
						10 Hz to 20 Hz	0.089 %			
						20 Hz to 40 Hz	0.052 %			
						40 Hz to 1 kHz	0.032 %			
						1 kHz to 5 kHz	0.25 %			
		30 mA	to	40 mA		5 kHz to 10 kHz	0.53 %			
						10 Hz to 20 Hz	0.084 %			
						20 Hz to 40 Hz	0.048 %			
						40 Hz to 1 kHz	0.028 %			
						1 kHz to 5 kHz	0.20 %			
		40 mA	to	50 mA		5 kHz to 10 kHz	0.43 %			
						10 Hz to 20 Hz	0.081 %			
						20 Hz to 40 Hz	0.045 %			
						40 Hz to 1 kHz	0.025 %			
						1 kHz to 5 kHz	0.17 %			
		50 mA	to	60 mA		5 kHz to 10 kHz	0.37 %			
						10 Hz to 20 Hz	0.079 %			
						20 Hz to 40 Hz	0.043 %			
						40 Hz to 1 kHz	0.023 %			
						1 kHz to 5 kHz	0.15 %			
		60 mA	to	80 mA		5 kHz to 10 kHz	0.33 %			
						10 Hz to 20 Hz	0.078 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						20 Hz to 40 Hz	0.042 %			
						40 Hz to 1 kHz	0.022 %			
						1 kHz to 5 kHz	0.13 %			
						5 kHz to 10 kHz	0.30 %			
		80 mA	to	100 mA		10 Hz to 20 Hz	0.076 %			
						20 Hz to 40 Hz	0.041 %			
						40 Hz to 1 kHz	0.021 %			
						1 kHz to 5 kHz	0.12 %			
						5 kHz to 10 kHz	0.27 %			
		100 mA	to	130 mA		10 Hz to 20 Hz	0.075 %			
						20 Hz to 40 Hz	0.040 %			
						40 Hz to 1 kHz	0.020 %			
						1 kHz to 5 kHz	0.11 %			
						5 kHz to 10 kHz	0.25 %			
		130 mA	to	180 mA		10 Hz to 20 Hz	0.074 %			
						20 Hz to 40 Hz	0.039 %			
						40 Hz to 1 kHz	0.019 %			
						1 kHz to 5 kHz	0.092 %			
						5 kHz to 10 kHz	0.24 %			
		180 mA	to	220 mA		10 Hz to 20 Hz	0.073 %			
						20 Hz to 40 Hz	0.038 %			
						40 Hz to 1 kHz	0.019 %			
						1 kHz to 5 kHz	0.084 %			
						5 kHz to 10 kHz	0.21 %			
		220 mA	to	300 mA		20 Hz to 1 kHz	0.082 %			
						1 kHz to 5 kHz	0.12 %			
						5 kHz to 10 kHz	0.93 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		300 mA	to	400 mA		20 Hz to 1 kHz	0.078 %			
						1 kHz to 5 kHz	0.11 %			
						5 kHz to 10 kHz	0.91 %			
		400 mA	to	500 mA		20 Hz to 1 kHz	0.075 %			
						1 kHz to 5 kHz	0.098 %			
						5 kHz to 10 kHz	0.90 %			
		500 mA	to	600 mA		20 Hz to 1 kHz	0.074 %			
						1 kHz to 5 kHz	0.094 %			
						5 kHz to 10 kHz	0.89 %			
		600 mA	to	1 A		20 Hz to 1 kHz	0.072 %			
						1 kHz to 5 kHz	0.092 %			
						5 kHz to 10 kHz	0.89 %			
		1 A	to	2.2 A		20 Hz to 1 kHz	0.070 %			
						1 kHz to 5 kHz	0.086 %			
						5 kHz to 10 kHz	0.88 %			
		2.2 A	to	3 A		10 Hz to 1 kHz	0.16 %			
						1 kHz to 5 kHz	0.22 %			
						5 kHz to 10 kHz	0.60 %			
		3 A	to	5 A		10 Hz to 1 kHz	0.14 %			
						1 kHz to 5 kHz	0.20 %			
						5 kHz to 10 kHz	0.51 %			
		5 A	to	11 A		10 Hz to 1 kHz	0.13 %			
						1 kHz to 5 kHz	0.18 %			
						5 kHz to 10 kHz	0.42 %			
		2.2 A	to	3 A		50 Hz	0.12 %			
		3 A	to	4 A		50 Hz	0.095 %			
		4 A	to	5 A		50 Hz	0.084 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		5 A	to	6 A		50 Hz	0.077 %			
		6 A	to	7 A		50 Hz	0.073 %			
		7 A	to	8 A		50 Hz	0.070 %			
		8 A	to	10 A		50 Hz	0.067 %			
		10 A	to	11 A		50 Hz	0.064 %			
		11 A	to	15 A		30 Hz to 45 Hz	0.23 %			
						45 Hz to 100 Hz	0.12 %			
						100 Hz to 1 kHz	0.52 %			
		15 A	to	25 A		30 Hz to 45 Hz	0.22 %			
						45 Hz to 100 Hz	0.12 %			
						100 Hz to 1 kHz	0.51 %			
		25 A	to	30 A		30 Hz to 45 Hz	0.21 %			
						45 Hz to 100 Hz	0.11 %			
						100 Hz to 1 kHz	0.49 %			
		30 A	to	1,000 A		50 Hz	0.5 %	Indirect generation with a current coil calibrator	TP4, TP21	
5*	DC resistance / DC resistance standards	0 Ω	to	0.1 Ω			6 μΩ	Direct measurement using a multimeter	TP5, TP21	
		0.1 Ω	to	0.5 Ω			12 μΩ			
		0.5 Ω	to	1 Ω			20 μΩ			
		1 Ω	to	2 Ω			0.0020 %			
		2 Ω	to	3 Ω			0.0017 %			
		3 Ω	to	5 Ω			0.0014 %			
		5 Ω	to	15 Ω			0.0012 %			
		15 Ω	to	19 Ω			0.0010 %			
		19 Ω	to	20 Ω			0.00098 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		20 Ω	to	25 Ω		0.0011 %				
		25 Ω	to	30 Ω		0.00096 %				
		30 Ω	to	50 Ω		0.00092 %				
		50 Ω	to	100 Ω		0.00086 %				
		100 Ω	to	190 Ω		0.00081 %				
		190 Ω	to	200 Ω		0.00078 %				
		200 Ω	to	250 Ω		0.0011 %				
		250 Ω	to	300 Ω		0.00096 %				
		300 Ω	to	500 Ω		0.00092 %				
		500 Ω	to	1 kΩ		0.00086 %				
		1 kΩ	to	1.9 kΩ		0.00081 %				
		1.9 kΩ	to	2 kΩ		0.00078 %				
		2 kΩ	to	2.5 kΩ		0.0011 %				
		2.5 kΩ	to	3 kΩ		0.00096 %				
		3 kΩ	to	5 kΩ		0.00092 %				
		5 kΩ	to	10 kΩ		0.00086 %				
		10 kΩ	to	19 kΩ		0.00081 %				
		19 kΩ	to	20 kΩ		0.00078 %				
		20 kΩ	to	25 kΩ		0.0011 %				
		25 kΩ	to	30 kΩ		0.00096 %				
		30 kΩ	to	50 kΩ		0.00092 %				
		50 kΩ	to	100 kΩ		0.00086 %				
		100 kΩ	to	190 kΩ		0.00081 %				
		190 kΩ	to	200 kΩ		0.00078 %				
		200 kΩ	to	300 kΩ		0.0014 %				
		300 kΩ	to	1 MΩ		0.0012 %				
		1 MΩ	to	1.9 MΩ		0.00096 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit					
		1.9 MΩ	to	2 MΩ	0.00091 %			
		2 MΩ	to	5 MΩ	0.0021 %			
		5 MΩ	to	19 MΩ	0.0018 %			
		19 MΩ	to	20 MΩ	0.0016 %			
		20 MΩ	to	30 MΩ	0.012 %			
		30 MΩ	to	50 MΩ	0.0094 %			
		50 MΩ	to	100 MΩ	0.0081 %			
		100 MΩ	to	190 MΩ	0.0071 %			
		190 MΩ	to	200 MΩ	0.0066 %			
		200 MΩ	to	250 MΩ	0.066 %			
		250 MΩ	to	300 MΩ	0.056 %			
		300 MΩ	to	400 MΩ	0.049 %			
		400 MΩ	to	500 MΩ	0.041 %			
		500 MΩ	to	700 MΩ	0.036 %			
		700 MΩ	to	1 GΩ	0.030 %			
		1 GΩ	to	1.9 GΩ	0.026 %			
		1.9 GΩ	to	2 GΩ	0.021 %			
		2 GΩ	to	2.5 GΩ	0.56 %			
		2.5 GΩ	to	3 GΩ	0.46 %			
		3 GΩ	to	4 GΩ	0.39 %			
		4 GΩ	to	5 GΩ	0.31 %			
		5 GΩ	to	6 GΩ	0.26 %			
		6 GΩ	to	7 GΩ	0.23 %			
		7 GΩ	to	10 GΩ	0.20 %			
		10 GΩ	to	15 GΩ	0.16 %			
		15 GΩ	to	19 GΩ	0.13 %			
		19 GΩ	to	20 GΩ	0.11 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
	DC resistance / DC resistance meters			0.1 mΩ		0.005 %	Direct generation with a calibrator	TP5, TP21		
				1 mΩ		0.005 %				
				10 mΩ		0.002 %				
				100 mΩ		0.001 %				
				1 Ω		0.0005 %				
				1.9 Ω		0.02 %				
				10 Ω		0.0005 %				
				19 Ω		0.004 %				
				100 Ω		0.0007 %				
				190 Ω		0.003 %				
				1 kΩ		0.0004 %				
				1.9 kΩ		0.002 %				
				10 kΩ		0.0004 %				
				19 kΩ		0.002 %				
				100 kΩ		0.0006 %				
				190 kΩ		0.003 %				
				1 MΩ		0.004 %				
				1.9 MΩ		0.004 %				
				10 MΩ		0.006 %				
				19 MΩ		0.008 %				
			100 MΩ		0.006 %					
			1 GΩ		0.014 %					
			10 GΩ		0.23 %					
			100 GΩ		0.80 %					
			200 GΩ		1.0 %					

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		0 Ω	to	1 Ω		1.5 mΩ				
		1 Ω	to	5 Ω		0.13 %				
		5 Ω	to	10 Ω		0.021%				
		10 Ω	to	11 Ω		0.026 %				
		11 Ω	to	20 Ω		0.030 %				
		20 Ω	to	33 Ω		0.023 %				
		33 Ω	to	330 Ω		0.016 %				
		330 Ω	to	1.1 kΩ		0.012 %				
		1.1 kΩ	to	3.3 kΩ		0.013 %				
		3.3 kΩ	to	11 kΩ		0.012 %				
		11 kΩ	to	33 kΩ		0.013 %				
		33 kΩ	to	110 kΩ		0.014 %				
		110 kΩ	to	330 kΩ		0.016 %				
		330 kΩ	to	1.1 MΩ		0.019 %				
		1.1 MΩ	to	3.3 MΩ		0.021%				
		3.3 MΩ	to	11 MΩ		0.072 %				
		11 MΩ	to	33 MΩ		0.15 %				
		33 MΩ	to	110 MΩ		0.59 %				
		110 MΩ	to	330 MΩ		0.69 %				
		330 MΩ	to	1.1 GΩ		2.0 %				
6*	AC resistance / AC resistance standards for frequencies from 20 Hz to 1 MHz	50 mΩ	to	0.1 Ω	100 Hz to 100 kHz	3.1 %	Direct measurement by a LCR meter	TP6, TP24		
					100 kHz to 300 kHz	5.3 %				
		0.1 Ω	to	0.2 Ω	20 Hz to 50 Hz	4.9 %				
					50 Hz to 100 Hz	3.5 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						100 Hz to 100 kHz	1.6 %			
						100 kHz to 300 kHz	2.7 %			
						300 kHz to 1 MHz	3.9 %			
		0.2 Ω	to	0.5 Ω		20 Hz to 50 Hz	2.6 %			
						50 Hz to 100 Hz	1.9 %			
						100 Hz to 100 kHz	0.85 %			
						100 kHz to 300 kHz	1.5 %			
						300 kHz to 1 MHz	2.1 %			
		0.5 Ω	to	1 Ω		20 Hz to 50 Hz	1.3 %			
						50 Hz to 100 Hz	0.83 %			
						100 Hz to 100 kHz	0.44 %			
						100 kHz to 300 kHz	0.67 %			
						300 kHz to 1 MHz	0.90 %			
		1 Ω	to	5 Ω		20 Hz to 50 Hz	0.78 %			
						50 Hz to 100 Hz	0.49 %			
						100 Hz to 100 kHz	0.30 %			
						100 kHz to 300 kHz	0.41 %			
						300 kHz to 1 MHz	0.53 %			
		5 Ω	to	10 Ω		20 Hz to 50 Hz	0.42 %			
						50 Hz to 100 Hz	0.23 %			
						100 Hz to 100 kHz	0.19 %			
						100 kHz to 300 kHz	0.21 %			
						300 kHz to 1 MHz	0.23 %			
		10 Ω	to	15 Ω		20 Hz to 50 Hz	0.38 %			
						50 Hz to 100 Hz	0.20 %			
						100 Hz to 100 kHz	0.18 %			
						100 kHz to 1 MHz	0.20 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit					
		15 Ω	to	50 Ω	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 1 MHz	0.36 % 0.18 % 0.13 %		
		50 Ω	to	2 kΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 1 MHz	0.34 % 0.17 % 0.12 %		
		2 kΩ	to	4 kΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 300 kHz 300 kHz to 1 MHz	0.33 % 0.16 % 0.11 % 0.17 %		
		4 kΩ	to	20 kΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.34 % 0.16 % 0.11 % 0.17 % 0.19 %		
		20 kΩ	to	100 kΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.37 % 0.19 % 0.13 % 0.17 % 0.20 % 0.28 %		
		100 kΩ	to	320 kΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.46 % 0.26 % 0.15 % 0.20 % 0.28 % 0.55 %		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		320 kΩ		to	400 kΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.49 % 0.28 % 0.21 % 0.31 % 0.65 %			
		400 kΩ		to	500 kΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.53 % 0.31 % 0.22 % 0.34 % 0.77 %			
		500 kΩ		to	700 kΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.61 % 0.37 % 0.25 % 0.42 % 1.1 %			
		700 kΩ		to	900 kΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.69 % 0.43 % 0.27 % 0.49 % 1.3 %			
		900 kΩ		to	1 MΩ	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.73 % 0.46 % 0.28 % 0.53 % 1.4 %			
		1 MΩ		to	2 MΩ	20 Hz to 50 Hz 50 Hz to 100 Hz	1.2 % 0.75 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						100 Hz to 100 kHz	0.41 %			
						100 kHz to 300 kHz	0.90 %			
						300 kHz to 1 MHz	2.7 %			
		2 MΩ	to	4 MΩ		20 Hz to 50 Hz	2.0 %			
						50 Hz to 100 Hz	1.4 %			
						100 Hz to 100 kHz	0.65 %			
						100 kHz to 300 kHz	1.7 %			
						300 kHz to 1 MHz	5.1 %			
		4 MΩ	to	6 MΩ		20 Hz to 50 Hz	2.8 %			
						50 Hz to 100 Hz	2.0 %			
						100 Hz to 100 kHz	0.90 %			
						100 kHz to 300 kHz	2.4 %			
						300 kHz to 1 MHz	7.6 %			
		6 MΩ	to	8 MΩ		20 Hz to 50 Hz	3.6 %			
						50 Hz to 100 Hz	2.6 %			
						100 Hz to 100 kHz	1.2 %			
						100 kHz to 300 kHz	3.2 %			
		8 MΩ	to	10 MΩ		20 Hz to 50 Hz	4.4 %			
						50 Hz to 100 Hz	3.2 %			
						100 Hz to 100 kHz	1.4 %			
						100 kHz to 300 kHz	3.9 %			
		10 MΩ	to	15 MΩ		20 Hz to 50 Hz	6.4 %			
						50 Hz to 100 Hz	4.8 %			
						100 Hz to 100 kHz	2.2 %			
						100 kHz to 300 kHz	5.8 %			
		15 MΩ	to	20 MΩ		20 Hz to 50 Hz	8.4 %			
						50 Hz to 100 Hz	6.3 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						3 MHz to 4 MHz 4 MHz to 5 MHz	8.6 % 9.9 %			
		5 Ω	to	6 Ω		1 MHz to 2 MHz 2 MHz to 3 MHz 3 MHz to 4 MHz 4 MHz to 5 MHz	5.1 % 6.1 % 7.1 % 8.1 %			
		6 Ω	to	7 Ω		1 MHz to 2 MHz 2 MHz to 3 MHz 3 MHz to 4 MHz 4 MHz to 5 MHz	4.4 % 5.2 % 6.1 % 7.0 %			
		7 Ω	to	8 Ω		1 MHz to 2 MHz 2 MHz to 3 MHz 3 MHz to 4 MHz 4 MHz to 5 MHz 5 MHz to 6 MHz	3.9 % 4.6 % 5.4 % 6.2 % 9.2 %			
		8 Ω	to	9 Ω		1 MHz to 2 MHz 2 MHz to 3 MHz 3 MHz to 4 MHz 4 MHz to 5 MHz 5 MHz to 6 MHz 6 MHz to 7 MHz	3.5 % 4.2 % 4.8 % 5.5 % 8.3 % 9.6 %			
		9 Ω	to	10 Ω		1 MHz to 2 MHz 2 MHz to 3 MHz 3 MHz to 4 MHz 4 MHz to 5 MHz 5 MHz to 6 MHz 6 MHz to 7 MHz	3.2 % 3.8 % 4.4 % 5.5 % 7.6 % 8.8 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit					
		10 Ω	to	20 Ω	1 MHz to 2 MHz	4.8 %		
					2 MHz to 3 MHz	5.8 %		
					3 MHz to 4 MHz	6.7 %		
					4 MHz to 5 MHz	7.7 %		
		20 Ω	to	30 Ω	1 MHz to 2 MHz	3.4 %		
					2 MHz to 3 MHz	4.0 %		
					3 MHz to 4 MHz	4.7 %		
					4 MHz to 5 MHz	5.4 %		
					5 MHz to 6 MHz	8.0 %		
					6 MHz to 7 MHz	9.3 %		
		30 Ω	to	50 Ω	1 MHz to 2 MHz	2.9 %		
					2 MHz to 3 MHz	3.5 %		
					3 MHz to 4 MHz	4.0 %		
					4 MHz to 5 MHz	4.6 %		
					5 MHz to 6 MHz	6.9 %		
					6 MHz to 7 MHz	8.0 %		
					7 MHz to 8 MHz	9.1 %		
		50 Ω	to	100 Ω	1 MHz to 2 MHz	2.5 %		
					2 MHz to 3 MHz	3.0 %		
					3 MHz to 4 MHz	3.5 %		
4 MHz to 5 MHz	4.0 %							
5 MHz to 6 MHz	5.9 %							
6 MHz to 7 MHz	6.9 %							
7 MHz to 8 MHz	7.9 %							
100 Ω	to	300 Ω	1 MHz to 2 MHz	2.8 %				
			2 MHz to 3 MHz	3.3 %				
			3 MHz to 4 MHz	3.9 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						4 MHz to 5 MHz	4.4 %			
						5 MHz to 6 MHz	6.8 %			
						6 MHz to 7 MHz	7.7 %			
						7 MHz to 8 MHz	8.8 %			
		300 Ω	to	500 Ω		1 MHz to 2 MHz	3.4 %			
						2 MHz to 3 MHz	4.0 %			
						3 MHz to 4 MHz	4.7 %			
						4 MHz to 5 MHz	5.4 %			
						5 MHz to 6 MHz	8.0 %			
						6 MHz to 7 MHz	9.3 %			
		500 Ω	to	700 Ω		1 MHz to 2 MHz	3.9 %			
						2 MHz to 3 MHz	4.6 %			
						3 MHz to 4 MHz	5.5 %			
						4 MHz to 5 MHz	6.3 %			
						5 MHz to 6 MHz	9.4 %			
		700 Ω	to	1 kΩ		1 MHz to 2 MHz	4.8 %			
						2 MHz to 3 MHz	5.7 %			
						3 MHz to 4 MHz	6.7 %			
						4 MHz to 5 MHz	7.6 %			
		1 kΩ	to	2 kΩ		1 MHz to 2 MHz	3.7 %			
						2 MHz to 3 MHz	4.4 %			
						3 MHz to 4 MHz	5.1 %			
						4 MHz to 5 MHz	5.8 %			
						5 MHz to 6 MHz	8.7 %			
		2 kΩ	to	3 kΩ		1 MHz to 2 MHz	4.4 %			
						2 MHz to 3 MHz	5.2 %			
						3 MHz to 4 MHz	6.1 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place	
		min	unit	max	unit						
						4 MHz to 5 MHz	7.0 %				
		3 kΩ	to	4 kΩ		1 MHz to 2 MHz	5.1 %				
						2 MHz to 3 MHz	6.1 %				
						3 MHz to 4 MHz	7.1 %				
						4 MHz to 5 MHz	8.1 %				
		4 kΩ	to	5 kΩ		1 MHz to 2 MHz	5.8 %				
						2 MHz to 3 MHz	7.0 %				
						3 MHz to 4 MHz	8.1 %				
						4 MHz to 5 MHz	9.3 %				
		5 kΩ	to	7 kΩ		1 MHz to 2 MHz	7.3 %				
						2 MHz to 3 MHz	8.7 %				
		7 kΩ	to	9 kΩ		1 MHz to 2 MHz	8.7 %				
		9 kΩ	to	10 kΩ		1 MHz to 2 MHz	9.4 %				
	AC resistance / AC resistance meters			0.1 Ω		10 Hz to 10 kHz	0.14 %	Direct measurement of a resistance standard, real impedance component	TP6, TP24		
						10 kHz to 100 kHz	0.18 %				
						100 kHz to 300 kHz	0.34 %				
						300 kHz to 500 kHz	0.47 %				
						500 kHz to 1 MHz	0.54 %				
				1 Ω		10 Hz to 10 kHz	0.038 %				
						10 kHz to 100 kHz	0.065 %				
						100 kHz to 300 kHz	0.074 %				
						300 kHz to 500 kHz	0.076 %				
						500 kHz to 1 MHz	0.098 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						1 MHz to 5 MHz 0.24 %				
						5 MHz to 10 MHz 0.47 %				
			10 Ω			10 Hz to 10 kHz 0.043 %				
						10 kHz to 100 kHz 0.055 %				
						100 kHz to 300 kHz 0.050 %				
						300 kHz to 500 kHz 0.045 %				
						500 kHz to 1 MHz 0.063 %				
						1 MHz to 5 MHz 0.21 %				
						5 MHz to 10 MHz 0.38 %				
			100 Ω			10 Hz to 10 kHz 0.005 %				
						10 kHz to 300 kHz 0.021 %				
						300 kHz to 1 MHz 0.022 %				
						1 MHz to 5 MHz 0.20 %				
						5 MHz to 10 MHz 0.21 %				
			1 kΩ			10 Hz to 10 kHz 0.005 %				
						10 kHz to 1 MHz 0.025 %				
						1 MHz to 5 MHz 0.21 %				
						5 MHz to 10 MHz 0.22 %				
			10 kΩ			10 Hz to 1 MHz 0.025 %				
			100 kΩ			10 Hz to 10 kHz 0.010 %				
						10 kHz to 500 kHz 0.11 %				
						500 kHz to 1 MHz 0.12 %				
			1 MΩ			10 Hz to 1 kHz 0.010 %				
						1 kHz to 1 MHz 0.11 %				
			10 MΩ			10 Hz to 10 kHz 0.016 %				
						10 kHz to 100 kHz 0.38 %				
						100 kHz to 300 kHz 0.52 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						300 kHz to 1 MHz	0.58 %			
7*	Capacity / Capacity standards	1 pF		5 pF		10 kHz to 15 kHz	2.2 %	Direct measurement by a RLC meter	TP7, TP24	
						15 kHz to 20 kHz	1.5 %			
						20 kHz to 25 kHz	1.2 %			
						25 kHz to 30 kHz	0.94 %			
						30 kHz to 40 kHz	0.81 %			
						40 kHz to 50 kHz	0.65 %			
						50 kHz to 60 kHz	0.55 %			
						60 kHz to 70 kHz	0.49 %			
						70 kHz to 80 kHz	0.44 %			
						80 kHz to 100 kHz	0.40 %			
						100 kHz to 110 kHz	0.75 %			
						110 kHz to 120 kHz	0.69 %			
						120 kHz to 130 kHz	0.65 %			
						130 kHz to 140 kHz	0.61 %			
						140 kHz to 150 kHz	0.58 %			
						150 kHz to 170 kHz	0.55 %			
						170 kHz to 200 kHz	0.50 %			
						200 kHz to 250 kHz	0.45 %			
						250 kHz to 300 kHz	0.39 %			
						300 kHz to 350 kHz	0.81 %			
						350 kHz to 400 kHz	0.72 %			
						400 kHz to 450 kHz	0.65 %			
						450 kHz to 500 kHz	0.59 %			
						500 kHz to 600 kHz	0.55 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						600 kHz to 700 kHz	0.49 %			
						700 kHz to 800 kHz	0.44 %			
						800 kHz to 1 MHz	0.40 %			
		5 pF	to	10 pF		1 kHz to 2 kHz	4.1 %			
						2 kHz to 3 kHz	2.2 %			
						3 kHz to 4 kHz	1.5 %			
						4 kHz to 5 kHz	1.2 %			
						5 kHz to 6 kHz	0.94 %			
						6 kHz to 7 kHz	0.81 %			
						7 kHz to 8 kHz	0.72 %			
						8 kHz to 9 kHz	0.65 %			
						9 kHz to 10 kHz	0.59 %			
						10 kHz to 15 kHz	0.55 %			
						15 kHz to 20 kHz	0.42 %			
						20 kHz to 25 kHz	0.36 %			
						25 kHz to 30 kHz	0.32 %			
						30 kHz to 1 MHz	0.29 %			
		10 pF	to	50 pF		500 Hz to 600 Hz	4.3 %			
						600 Hz to 700 Hz	3.7 %			
						700 Hz to 800 Hz	3.2 %			
						800 Hz to 900 Hz	2.8 %			
						900 Hz to 1 kHz	2.6 %			
						1 kHz to 2 kHz	2.2 %			
						2 kHz to 3 kHz	1.2 %			
						3 kHz to 4 kHz	0.81 %			
						4 kHz to 5 kHz	0.65 %			
						5 kHz to 6 kHz	0.55 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						6 kHz to 7 kHz	0.49 %			
						7 kHz to 8 kHz	0.44 %			
						8 kHz to 9 kHz	0.40 %			
						9 kHz to 10 kHz	0.38 %			
						10 kHz to 15 kHz	0.36 %			
						15 kHz to 20 kHz	0.29 %			
						20 kHz to 1 MHz	0.26 %			
		50 pF	to	100 pF		100 Hz to 200 Hz	4.3 %			
						200 Hz to 300 Hz	2.3 %			
						300 Hz to 400 Hz	1.7 %			
						400 Hz to 500 Hz	1.4 %			
						500 Hz to 600 Hz	1.2 %			
						600 Hz to 700 Hz	0.99 %			
						700 Hz to 800 Hz	0.89 %			
						800 Hz to 900 Hz	0.82 %			
						900 Hz to 1 kHz	0.77 %			
						1 kHz to 2 kHz	0.55 %			
						2 kHz to 3 kHz	0.36 %			
						3 kHz to 4 kHz	0.29 %			
						4 kHz to 5 kHz	0.26 %			
						5 kHz to 1 MHz	0.24 %			
		100 pF	to	500 pF		50 Hz to 60 Hz	9.9 %			
						60 Hz to 70 Hz	7.9 %			
						70 Hz to 80 Hz	6.5 %			
						80 Hz to 90 Hz	5.6 %			
						90 Hz to 100 Hz	4.9 %			
						100 Hz to 160 Hz	2.3 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						160 Hz to 200 Hz	1.4 %			
						200 Hz to 300 Hz	1.2 %			
						300 Hz to 400 Hz	0.81 %			
						400 Hz to 500 Hz	0.65 %			
						500 Hz to 600 Hz	0.55 %			
						600 Hz to 700 Hz	0.49 %			
						700 Hz to 800 Hz	0.44 %			
						800 Hz to 900 Hz	0.40 %			
						900 Hz to 1 kHz	0.38 %			
						1 kHz to 2 kHz	0.36 %			
						2 kHz to 3 kHz	0.26 %			
						3 kHz to 4 kHz	0.23 %			
						4 kHz to 5 kHz	0.21 %			
						5 kHz to 1 MHz	0.17 %			
		500 pF	to	1 nF		50 Hz to 60 Hz	2.3 %			
						60 Hz to 70 Hz	1.9 %			
						70 Hz to 80 Hz	1.6 %			
						80 Hz to 100 Hz	1.4 %			
						100 Hz to 160 Hz	0.72 %			
						160 Hz to 200 Hz	0.40 %			
						200 Hz to 300 Hz	0.36 %			
						300 Hz to 400 Hz	0.29 %			
						400 Hz to 500 Hz	0.26 %			
						500 Hz to 700 Hz	0.24 %			
						700 Hz to 1 kHz	0.22 %			
						1 kHz to 2 kHz	0.36 %			
						2 kHz to 3 kHz	0.26 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						3 kHz to 4 kHz	0.23 %			
						4 kHz to 5 kHz	0.21 %			
						5 kHz to 1 MHz	0.17 %			
		1 nF	to	5 nF		20 Hz to 30 Hz	3.6 %			
						30 Hz to 40 Hz	2.2 %			
						40 Hz to 50 Hz	1.6 %			
						50 Hz to 60 Hz	1.2 %			
						60 Hz to 70 Hz	0.91 %			
						70 Hz to 80 Hz	0.77 %			
						80 Hz to 90 Hz	0.68 %			
						90 Hz to 100 Hz	0.61 %			
						100 Hz to 200 Hz	0.36 %			
						200 Hz to 300 Hz	0.26 %			
						300 Hz to 500 Hz	0.23 %			
						500 Hz to 1 MHz	0.15 %			
		5 nF	to	10 nF		20 Hz to 30 Hz	0.97 %			
						30 Hz to 40 Hz	0.70 %			
						40 Hz to 50 Hz	0.58 %			
						50 Hz to 60 Hz	0.35 %			
						60 Hz to 70 Hz	0.31 %			
						70 Hz to 100 Hz	0.28 %			
						100 Hz to 400 Hz	0.20 %			
						400 Hz to 500 Hz	0.17 %			
						500 Hz to 1 MHz	0.13 %			
		10 nF	to	50 nF		20 Hz to 30 Hz	0.65 %			
						30 Hz to 40 Hz	0.52 %			
						40 Hz to 50 Hz	0.46 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						50 Hz to 70 Hz 70 Hz to 100 Hz 100 Hz to 103 kHz 103 kHz to 150 kHz 150 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.26 % 0.22 % 0.13 % 0.19 % 0.21 % 0.28 % 0.39 %			
		50 nF	to	100 nF		20 Hz to 40 Hz 40 Hz to 50 Hz 50 Hz to 70 Hz 70 Hz to 100 Hz 100 Hz to 103 kHz 103 kHz to 150 kHz 150 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.40 % 0.36 % 0.26 % 0.22 % 0.13 % 0.19 % 0.21 % 0.28 % 0.39 %			
		100 nF	to	500 nF		20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 10 kHz 10 kHz to 70 kHz 70 kHz to 100 kHz 100 kHz to 140 kHz 140 kHz to 230 kHz 230 kHz to 280 kHz 280 kHz to 300 kHz 300 kHz to 350 kHz 350 kHz to 400 kHz	0.36 % 0.18 % 0.12 % 0.22 % 0.25 % 0.27 % 0.34 % 0.38 % 0.40 % 0.56 % 0.62 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						400 kHz to 450 kHz	0.68 %			
						450 kHz to 500 kHz	0.74 %			
						500 kHz to 600 kHz	0.86 %			
						600 kHz to 700 kHz	0.97 %			
						700 kHz to 800 kHz	1.1 %			
						800 kHz to 1 MHz	1.4 %			
		500 nF	to	1 μF		20 Hz to 50 Hz	0.36 %			
						50 Hz to 100 Hz	0.18 %			
						100 Hz to 10 kHz	0.12 %			
						10 kHz to 70 kHz	0.22 %			
						70 kHz to 100 kHz	0.25 %			
						100 kHz to 120 kHz	0.35 %			
						120 kHz to 150 kHz	0.40 %			
						150 kHz to 180 kHz	0.45 %			
						180 kHz to 220 kHz	0.51 %			
						220 kHz to 250 kHz	0.56 %			
						250 kHz to 300 kHz	0.64 %			
						300 kHz to 350 kHz	0.97 %			
						350 kHz to 400 kHz	1.1 %			
						400 kHz to 500 kHz	1.4 %			
						500 kHz to 600 kHz	1.6 %			
						600 kHz to 700 kHz	1.8 %			
						700 kHz to 1 MHz	2.5 %			
		1 μF	to	5 μF		20 Hz to 50 Hz	0.34 %			
						50 Hz to 5 kHz	0.20 %			
						5 kHz to 10 kHz	0.25 %			
						10 kHz to 15 kHz	0.29 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						15 kHz to 20 kHz	0.33 %			
						20 kHz to 25 kHz	0.38 %			
						25 kHz to 30 kHz	0.42 %			
						30 kHz to 40 kHz	0.33 %			
						40 kHz to 50 kHz	0.38 %			
						50 kHz to 60 kHz	0.42 %			
						60 kHz to 70 kHz	0.46 %			
						70 kHz to 80 kHz	0.51 %			
						80 kHz to 90 kHz	0.55 %			
						90 kHz to 100 kHz	0.59 %			
						100 kHz to 110 kHz	1.1 %			
						110 kHz to 140 kHz	1.3 %			
						140 kHz to 160 kHz	1.5 %			
						160 kHz to 190 kHz	1.7 %			
						190 kHz to 210 kHz	1.9 %			
						210 kHz to 250 kHz	2.2 %			
						250 kHz to 300 kHz	2.6 %			
						300 kHz to 350 kHz	4.3 %			
						350 kHz to 400 kHz	4.9 %			
						400 kHz to 450 kHz	5.4 %			
						450 kHz to 500 kHz	6.0 %			
		5 μF	to	10 μF		20 Hz to 50 Hz	0.34 %			
						50 Hz to 5 kHz	0.20 %			
						5 kHz to 10 kHz	0.25 %			
						10 kHz to 15 kHz	0.29 %			
						15 kHz to 20 kHz	0.33 %			
						20 kHz to 25 kHz	0.38 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						25 kHz to 30 kHz	0.42 %			
						30 kHz to 40 kHz	0.51 %			
						40 kHz to 50 kHz	0.59 %			
						50 kHz to 60 kHz	0.68 %			
						60 kHz to 70 kHz	0.77 %			
						70 kHz to 80 kHz	0.85 %			
						80 kHz to 90 kHz	0.94 %			
						90 kHz to 100 kHz	1.1 %			
						100 kHz to 110 kHz	2.0 %			
						110 kHz to 140 kHz	2.4 %			
						140 kHz to 160 kHz	2.8 %			
						160 kHz to 190 kHz	3.2 %			
						190 kHz to 210 kHz	3.6 %			
						210 kHz to 250 kHz	4.2 %			
						250 kHz to 300 kHz	5.0 %			
		10 μF	to	20 μF		20 Hz to 50 Hz	0.34 %			
						50 Hz to 500 Hz	0.20 %			
						500 Hz to 700 Hz	0.22 %			
						700 Hz to 1 kHz	0.25 %			
						1 kHz to 2 kHz	0.33 %			
						2 kHz to 3 kHz	0.42 %			
						3 kHz to 4 kHz	0.51 %			
						4 kHz to 5 kHz	0.59 %			
						5 kHz to 6 kHz	0.68 %			
						6 kHz to 7 kHz	0.77 %			
						7 kHz to 8 kHz	0.85 %			
						8 kHz to 9 kHz	0.94 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						9 kHz to 10 kHz	1.1 %			
						10 kHz to 15 kHz	1.5 %			
						15 kHz to 20 kHz	1.9 %			
						20 kHz to 25 kHz	2.4 %			
						25 kHz to 30 kHz	2.8 %			
						30 kHz to 40 kHz	1.9 %			
						40 kHz to 50 kHz	2.4 %			
						50 kHz to 60 kHz	2.8 %			
						60 kHz to 70 kHz	3.2 %			
						70 kHz to 80 kHz	3.7 %			
						80 kHz to 90 kHz	4.1 %			
						90 kHz to 100 kHz	4.6 %			
		20 μF	to	50 μF		20 Hz to 50 Hz	0.34 %			
						50 Hz to 500 Hz	0.20 %			
						500 Hz to 700 Hz	0.22 %			
						700 Hz to 1 kHz	0.25 %			
						1 kHz to 2 kHz	0.33 %			
						2 kHz to 3 kHz	0.42 %			
						3 kHz to 4 kHz	0.51 %			
						4 kHz to 5 kHz	0.59 %			
						5 kHz to 6 kHz	0.68 %			
						6 kHz to 7 kHz	0.77 %			
						7 kHz to 8 kHz	0.85 %			
						8 kHz to 9 kHz	0.94 %			
						9 kHz to 10 kHz	1.1 %			
						10 kHz to 15 kHz	1.5 %			
						15 kHz to 20 kHz	1.9 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						20 kHz to 25 kHz	2.4 %			
						25 kHz to 30 kHz	2.8 %			
						30 kHz to 40 kHz	1.9 %			
						40 kHz to 50 kHz	2.4 %			
		50 μF	to	100 μF		20 Hz to 50 Hz	0.34 %			
						50 Hz to 500 Hz	0.20 %			
						500 Hz to 700 Hz	0.22 %			
						700 Hz to 1 kHz	0.25 %			
						1 kHz to 2 kHz	0.33 %			
						2 kHz to 3 kHz	0.42 %			
						3 kHz to 4 kHz	0.51 %			
						4 kHz to 5 kHz	0.59 %			
						5 kHz to 6 kHz	0.68 %			
						6 kHz to 7 kHz	0.77 %			
						7 kHz to 8 kHz	0.85 %			
						8 kHz to 9 kHz	0.94 %			
						9 kHz to 10 kHz	1.1 %			
						10 kHz to 15 kHz	1.5 %			
						15 kHz to 20 kHz	1.9 %			
		100 μF	to	500 μF		20 Hz to 50 Hz	0.44 %			
						50 Hz to 70 Hz	0.29 %			
						70 Hz to 200 Hz	0.33 %			
						200 Hz to 300 Hz	0.42 %			
						300 Hz to 400 Hz	0.51 %			
						400 Hz to 500 Hz	0.59 %			
						500 Hz to 600 Hz	0.68 %			
						600 Hz to 700 Hz	0.77 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						700 Hz to 800 Hz 800 Hz to 900 Hz 900 Hz to 2 kHz 2 kHz to 3 kHz 3 kHz to 4 kHz 4 kHz to 5 kHz	0.85 % 0.94 % 1.1 % 1.5 % 1.9 % 2.4 %			
		500 μF	to	1 mF		20 Hz to 50 Hz 50 Hz to 70 Hz 70 Hz to 200 Hz 200 Hz to 300 Hz 300 Hz to 400 Hz 400 Hz to 500 Hz 500 Hz to 600 Hz 600 Hz to 700 Hz 700 Hz to 800 Hz 800 Hz to 900 Hz 900 Hz to 1 kHz 1 kHz to 2 kHz	0.44 % 0.29 % 0.33 % 0.42 % 0.51 % 0.59 % 0.68 % 0.77 % 0.85 % 0.94 % 1.1 % 1.9 %			
		1 mF	to	5 mF		20 Hz to 30 Hz 30 Hz to 40 Hz 40 Hz to 80 Hz 80 Hz to 90 Hz 90 Hz to 160 Hz 160 Hz to 300 Hz 300 Hz to 400 Hz 400 Hz to 500 Hz 500 Hz to 600 Hz	0.70 % 0.78 % 0.89 % 0.96 % 1.1 % 1.7 % 2.1 % 2.6 % 3.0 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						600 Hz to 700 Hz 700 Hz to 800 Hz 800 Hz to 900 Hz 900 Hz to 1 kHz	3.4 % 3.9 % 4.3 % 4.7 %			
		5 mF	to	10 mF		50 Hz to 60 Hz 60 Hz to 160 Hz 160 Hz to 220 Hz 220 Hz to 300 Hz	1.4 % 1.9 % 2.3 % 3.0 %			
	Capacity / Electrical capacity meters					1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz	0.010 % 0.004 % 0.009 % 0.003 % 0.003 % 0.006 % 0.019 % 0.053 % 0.081 %	Direct measurement of a capacity standard	TP7, TP21, TP24	
						1 pF 100 Hz to 1 kHz 1 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz	0.019 % 0.022 % 0.023 % 0.10 %			
						10 pF 100 Hz to 1 kHz 1 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz	0.006 % 0.012 % 0.050 % 0.025 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory

U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit					
		100 pF		100 Hz to 1 kHz	0.011 %			
				1 kHz to 100 kHz	0.011 %			
				100 kHz to 1 MHz	0.021 %			
				1 MHz to 10 MHz	0.31 %			
		1 nF		100 Hz to 1 kHz	0.006 %			
				1 kHz to 100 kHz	0.010 %			
				100 kHz to 1 MHz	0.011 %			
				1 MHz to 10 MHz	0.64 %			
		10 nF		100 Hz to 1 kHz	0.005 %			
				1 kHz to 10 kHz	0.006 %			
				10 kHz to 100 kHz	0.010 %			
		100 nF		100 Hz to 1 kHz	0.015 %			
				1 kHz to 100 kHz	0.013 %			
		1 μF		100 Hz to 1 kHz	0.024 %			
				1 kHz to 10 kHz	0.025 %			
				10 kHz to 100 kHz	0.042 %			
		10 μF		100 Hz to 1 kHz	0.054 %			
				1 kHz to 10 kHz	0.060 %			
				10 kHz to 100 kHz	0.26 %			
		100 μF		100 Hz to 1 kHz	0.25 %			
				1 kHz to 10 kHz	0.24 %			
		220 pF	to	1.1 nF	10 Hz to 10 kHz	0.5 % + 0.01 nF		
		1.1 nF	to	3.3 nF	10 Hz to 3 kHz	0.5 % + 0.01 nF		
		3.3 nF	to	11 nF	10 Hz to 1 kHz	0.25 % + 0.01 nF		
		11 nF	to	110 nF	10 Hz to 1 kHz	0.25 % + 0.1 nF		
		110 nF	to	330 nF	10 Hz to 1 kHz	0.25 % + 0.3 nF		
		0.33 μF	to	1.1 μF	10 Hz to 600 Hz	0.25 % + 1 nF		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		1.1 μF	to	3.3 μF		10 Hz to 300 kHz	0.25 % + 3 nF			
		3.3 μF	to	11 μF		10 Hz to 150 Hz	0.25 % + 10 nF			
		11 μF	to	33 μF		10 Hz to 120 Hz	0.40 % + 30 nF			
		33 μF	to	110 μF		10 Hz to 80 Hz	0.45 % + 100 nF			
		110 μF	to	330 μF		0 Hz to 50 Hz	0.45 % + 300 nF			
		0.33 mF	to	1.1 mF		0 Hz to 20 Hz	0.45 % + 1 μF			
		1.1 mF	to	3.3 mF		0 Hz to 6 Hz	0.45 % + 3 μF			
		3.3 mF	to	11 mF		0 Hz to 2 Hz	0.45 % + 10 μF			
		11 mF	to	33 mF		0 Hz to 0.6 Hz	0.75 % + 30 μF			
		33 mF	to	110 mF		0 Hz to 0.2 Hz	1 % + 100 μF			
8*	Loss factor D / Loss factor meters	-0.001	to	0.001		10 pF, 100 pF, 1 nF 10 nF 100 nF	0.00001 (abs.) 0.000011 (abs.) 0.00006 (abs.)	Direct measurement of reference loss factor at f = 1 kHz	TP30, TP24	
		0.001	to	0.01		10 pF, 100 pF, 1 nF 100 nF 1 μF	0.00002 (abs.) 0.00006 (abs.) 0.0003 (abs.)			
		0.01	to	0.1		10 pF, 100 pF, 10 nF 1 nF 1 μF	0.00006 (abs.) 0.0011 (abs.) 0.0003 (abs.)			
		0.1	to	1		10 pF, 100 pF, 1 nF, 10 nF 100 nF	0.0006 (abs.) 0.0003 (abs.)			
9*	Inductance / Inductance standards	1 μH	to	5 μH		10 kHz to 15 kHz 15 kHz to 20 kHz	2.4 % 1.7 %	Direct measurement by a RLC meter	TP8, TP24	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						20 kHz to 25 kHz	1.3 %			
						25 kHz to 30 kHz	1.1 %			
						30 kHz to 40 kHz	0.89 %			
						40 kHz to 50 kHz	0.71 %			
						50 kHz to 60 kHz	0.60 %			
						60 kHz to 70 kHz	0.53 %			
						70 kHz to 80 kHz	0.47 %			
						80 kHz to 90 kHz	0.43 %			
						90 kHz to 100 kHz	0.40 %			
						100 kHz to 150 kHz	0.56 %			
						150 kHz to 200 kHz	0.45 %			
						200 kHz to 250 kHz	0.36 %			
						250 kHz to 300 kHz	0.32 %			
						300 kHz to 400 kHz	0.36 %			
						400 kHz to 500 kHz	0.31 %			
						500 kHz to 700 kHz	0.28 %			
						700 kHz to 1 MHz	0.24 %			
		5 μH	to	10 μH		2 kHz to 3 kHz	2.4 %			
						3 kHz to 5 kHz	1.7 %			
						5 kHz to 7 kHz	1.1 %			
						7 kHz to 10 kHz	0.79 %			
						10 kHz to 15 kHz	0.60 %			
						15 kHz to 20 kHz	0.45 %			
						20 kHz to 25 kHz	0.38 %			
						25 kHz to 30 kHz	0.34 %			
						30 kHz to 40 kHz	0.31 %			
						40 kHz to 50 kHz	0.27 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						50 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 150 kHz 150 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz	0.25 % 0.23 % 0.24 % 0.22 % 0.20 % 0.13 %			
		10 μH	to	50 μH		1 kHz to 2 kHz 2 kHz to 3 kHz 3 kHz to 5 kHz 5 kHz to 7 kHz 7 kHz to 10 kHz 10 kHz to 15 kHz 15 kHz to 20 kHz 20 kHz to 25 kHz 25 kHz to 30 kHz 30 kHz to 50 kHz 50 kHz to 70 kHz 70 kHz to 100 kHz 100 kHz to 250 kHz 250 kHz to 300 kHz 300 kHz to 1 MHz	2.4 % 1.3 % 0.89 % 0.60 % 0.47 % 0.38 % 0.31 % 0.27 % 0.25 % 0.23 % 0.21 % 0.19 % 0.20 % 0.12 % 0.13 %			
		50 μH	to	100 μH		200 Hz to 300 Hz 300 Hz to 400 Hz 400 Hz to 500 Hz 500 Hz to 600 Hz 600 Hz to 700 Hz 700 Hz to 800 Hz	2.6 % 1.8 % 1.3 % 1.1 % 0.89 % 0.79 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						800 Hz to 900 Hz	0.71 %			
						900 Hz to 1 kHz	0.65 %			
						1 kHz to 2 kHz	0.60 %			
						2 kHz to 3 kHz	0.38 %			
						3 kHz to 5 kHz	0.31 %			
						5 kHz to 10 kHz	0.25 %			
						10 kHz to 15 kHz	0.21 %			
						15 kHz to 30 kHz	0.19 %			
						30 kHz to 50 kHz	0.18 %			
						50 kHz to 100 kHz	0.12 %			
						100 kHz to 250 kHz	0.20 %			
						250 kHz to 300 kHz	0.12 %			
						300 kHz to 1 MHz	0.13 %			
		100 μH	to	500 μH		80 Hz to 90 Hz	6.2 %			
						90 Hz to 100 Hz	5.4 %			
						100 Hz to 200 Hz	2.6 %			
						200 Hz to 300 Hz	1.3 %			
						300 Hz to 400 Hz	0.89 %			
						400 Hz to 500 Hz	0.71 %			
						500 Hz to 600 Hz	0.60 %			
						600 Hz to 700 Hz	0.53 %			
						700 Hz to 800 Hz	0.47 %			
						800 Hz to 900 Hz	0.43 %			
						900 Hz to 1 kHz	0.40 %			
						1 kHz to 2 kHz	0.38 %			
						2 kHz to 3 kHz	0.27 %			
						3 kHz to 5 kHz	0.23 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						5 kHz to 10 kHz	0.21 %			
						10 kHz to 25 kHz	0.18 %			
						25 kHz to 30 kHz	0.12 %			
						30 kHz to 100 kHz	0.12 %			
						100 kHz to 300 kHz	0.11 %			
						300 kHz to 1 MHz	0.17 %			
		500 μH	to	1 mH		50 Hz to 60 Hz	2.3 %			
						60 Hz to 70 Hz	1.9 %			
						70 Hz to 80 Hz	1.6 %			
						80 Hz to 100 Hz	1.4 %			
						100 Hz to 200 Hz	0.60 %			
						200 Hz to 300 Hz	0.38 %			
						300 Hz to 400 Hz	0.31 %			
						400 Hz to 500 Hz	0.27 %			
						500 Hz to 600 Hz	0.25 %			
						600 Hz to 900 Hz	0.23 %			
						900 Hz to 1 kHz	0.21 %			
						1 kHz to 2 kHz	0.38 %			
						2 kHz to 3 kHz	0.27 %			
						3 kHz to 5 kHz	0.23 %			
						5 kHz to 10 kHz	0.21 %			
						10 kHz to 25 kHz	0.18 %			
						25 kHz to 100 kHz	0.12 %			
						100 kHz to 300 kHz	0.11 %			
						300 kHz to 1 MHz	0.17 %			
		1 mH	to	5 mH		20 Hz to 30 Hz	3.9 %			
						30 Hz to 40 Hz	2.5 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						40 Hz to 50 Hz	1.8 %			
						50 Hz to 60 Hz	1.3 %			
						60 Hz to 80 Hz	1.0 %			
						80 Hz to 90 Hz	0.74 %			
						90 Hz to 100 Hz	0.66 %			
						100 Hz to 200 Hz	0.38 %			
						200 Hz to 300 Hz	0.27 %			
						300 Hz to 500 Hz	0.23 %			
						500 Hz to 700 Hz	0.21 %			
						700 Hz to 1 kHz	0.19 %			
						1 kHz to 3 kHz	0.18 %			
						3 kHz to 100 kHz	0.12 %			
						100 kHz to 300 kHz	0.17 %			
						300 kHz to 600 kHz	0.21 %			
						600 kHz to 1 MHz	0.24 %			
		5 mH	to	10 mH		20 Hz to 30 Hz	1.1 %			
						30 Hz to 40 Hz	0.75 %			
						40 Hz to 50 Hz	0.61 %			
						50 Hz to 60 Hz	0.37 %			
						60 Hz to 70 Hz	0.33 %			
						70 Hz to 80 Hz	0.30 %			
						80 Hz to 90 Hz	0.28 %			
						90 Hz to 100 Hz	0.26 %			
						100 Hz to 200 Hz	0.38 %			
						200 Hz to 300 Hz	0.27 %			
						300 Hz to 500 Hz	0.23 %			
						500 Hz to 700 Hz	0.21 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						700 Hz to 1 kHz 0.19 %				
						1 kHz to 3 kHz 0.18 %				
						3 kHz to 100 kHz 0.12 %				
						100 kHz to 300 kHz 0.17 %				
						300 kHz to 600 kHz 0.21 %				
						600 kHz to 1 MHz 0.24 %				
		10 mH	to	50 mH		20 Hz to 30 Hz 0.69 %				
						30 Hz to 40 Hz 0.54 %				
						40 Hz to 50 Hz 0.47 %				
						50 Hz to 60 Hz 0.27 %				
						60 Hz to 80 Hz 0.24 %				
						80 Hz to 100 Hz 0.22 %				
						100 Hz to 250 Hz 0.18 %				
						250 Hz to 1 kHz 0.12 %				
						1 kHz to 30 kHz 0.11 %				
						30 kHz to 100 kHz 0.17 %				
						100 kHz to 200 kHz 0.21 %				
						200 kHz to 300 kHz 0.23 %				
						300 kHz to 400 kHz 0.31 %				
						400 kHz to 500 kHz 0.35 %				
						500 kHz to 700 kHz 0.43 %				
						700 kHz to 1 MHz 0.55 %				
		50 mH	to	100 mH		20 Hz to 30 Hz 0.40 %				
						30 Hz to 50 Hz 0.37 %				
						50 Hz to 60 Hz 0.27 %				
						60 Hz to 80 Hz 0.24 %				
						80 Hz to 100 Hz 0.22 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
 CAB number 2273, Calibration Laboratory
 U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						100 Hz to 250 Hz	0.18 %			
						250 Hz to 1 kHz	0.12 %			
						1 kHz to 30 kHz	0.11 %			
						30 kHz to 100 kHz	0.17 %			
						100 kHz to 200 kHz	0.21 %			
						200 kHz to 300 kHz	0.23 %			
						300 kHz to 400 kHz	0.47 %			
						400 kHz to 500 kHz	0.55 %			
						500 kHz to 600 kHz	0.62 %			
						600 kHz to 700 kHz	0.70 %			
						700 kHz to 800 kHz	0.78 %			
						800 kHz to 1 MHz	0.93 %			
		100 mH	to	500 mH		20 Hz to 30 Hz	0.27 %			
						30 Hz to 50 Hz	0.35 %			
						50 Hz to 100 Hz	0.18 %			
						100 Hz to 1 kHz	0.12 %			
						1 kHz to 30 kHz	0.13 %			
						30 kHz to 60 kHz	0.21 %			
						60 kHz to 100 kHz	0.24 %			
						100 kHz to 150 kHz	0.33 %			
						150 kHz to 200 kHz	0.39 %			
						200 kHz to 250 kHz	0.45 %			
						250 kHz to 300 kHz	0.51 %			
						300 kHz to 400 kHz	1.8 %			
						400 kHz to 500 kHz	2.1 %			
						500 kHz to 600 kHz	2.5 %			
						600 kHz to 700 kHz	2.9 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						700 kHz to 800 kHz	3.3 %			
						800 kHz to 900 kHz	3.7 %			
						900 kHz to 1 MHz	4.1 %			
		500 mH	to	1 H		20 Hz to 30 Hz	0.27 %			
						30 Hz to 50 Hz	0.35 %			
						50 Hz to 100 Hz	0.18 %			
						100 Hz to 1 kHz	0.12 %			
						1 kHz to 30 kHz	0.13 %			
						30 kHz to 60 kHz	0.21 %			
						60 kHz to 100 kHz	0.24 %			
						100 kHz to 150 kHz	0.51 %			
						150 kHz to 200 kHz	0.62 %			
						200 kHz to 250 kHz	0.74 %			
						250 kHz to 300 kHz	0.86 %			
						300 kHz to 400 kHz	3.3 %			
						400 kHz to 500 kHz	4.1 %			
						500 kHz to 600 kHz	4.9 %			
						600 kHz to 700 kHz	5.6 %			
						700 kHz to 800 kHz	6.4 %			
						800 kHz to 900 kHz	7.2 %			
						900 kHz to 1 MHz	8.0 %			
		1 H	to	2 H		20 Hz to 50 Hz	0.34 %			
						50 Hz to 100 Hz	0.18 %			
						100 Hz to 1 kHz	0.12 %			
						1 kHz to 5 kHz	0.15 %			
						5 kHz to 10 kHz	0.24 %			
						10 kHz to 15 kHz	0.28 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						15 kHz to 20 kHz	0.31 %			
						20 kHz to 25 kHz	0.35 %			
						25 kHz to 30 kHz	0.38 %			
						30 kHz to 50 kHz	0.35 %			
						50 kHz to 70 kHz	0.43 %			
						70 kHz to 100 kHz	0.55 %			
						100 kHz to 150 kHz	1.9 %			
						150 kHz to 200 kHz	2.5 %			
						200 kHz to 250 kHz	3.1 %			
						250 kHz to 300 kHz	3.7 %			
						300 kHz to 400 kHz	6.4 %			
						400 kHz to 500 kHz	8.0 %			
		2 H	to	5 H		20 Hz to 50 Hz	0.34 %			
						50 Hz to 100 Hz	0.18 %			
						100 Hz to 1 kHz	0.12 %			
						1 kHz to 5 kHz	0.15 %			
						5 kHz to 10 kHz	0.24 %			
						10 kHz to 15 kHz	0.28 %			
						15 kHz to 20 kHz	0.31 %			
						20 kHz to 25 kHz	0.35 %			
						25 kHz to 30 kHz	0.38 %			
						30 kHz to 50 kHz	0.35 %			
						50 kHz to 70 kHz	0.43 %			
						70 kHz to 100 kHz	0.55 %			
						100 kHz to 150 kHz	1.9 %			
						150 kHz to 200 kHz	2.5 %			
						200 kHz to 250 kHz	3.1 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						250 kHz to 300 kHz	3.7 %			
		5 H	to	10 H		50 Hz to 100 Hz	0.18 %			
						100 Hz to 1 kHz	0.12 %			
						1 kHz to 5 kHz	0.15 %			
						5 kHz to 10 kHz	0.24 %			
						10 kHz to 15 kHz	0.28 %			
						15 kHz to 20 kHz	0.31 %			
						20 kHz to 25 kHz	0.35 %			
						25 kHz to 30 kHz	0.38 %			
						30 kHz to 40 kHz	0.47 %			
						40 kHz to 50 kHz	0.55 %			
						50 kHz to 60 kHz	0.62 %			
						60 kHz to 70 kHz	0.70 %			
						70 kHz to 80 kHz	0.78 %			
						80 kHz to 90 kHz	0.86 %			
						90 kHz to 100 kHz	0.93 %			
						100 kHz to 150 kHz	3.7 %			
						150 kHz to 200 kHz	4.9 %			
						200 kHz to 250 kHz	6.0 %			
						250 kHz to 300 kHz	7.2 %			
		10 H	to	50 H		50 Hz to 100 Hz	0.18 %			
						100 Hz to 500 Hz	0.15 %			
						500 Hz to 800 Hz	0.22 %			
						800 Hz to 2 kHz	0.24 %			
						2 kHz to 3 kHz	0.28 %			
						3 kHz to 5 kHz	0.35 %			
						5 kHz to 7 kHz	0.43 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						7 kHz to 10 kHz	0.55 %			
						10 kHz to 15 kHz	0.74 %			
						15 kHz to 20 kHz	0.93 %			
						20 kHz to 25 kHz	1.2 %			
						25 kHz to 30 kHz	1.4 %			
						30 kHz to 40 kHz	1.8 %			
						40 kHz to 50 kHz	2.1 %			
						50 kHz to 60 kHz	2.5 %			
						60 kHz to 70 kHz	2.9 %			
						70 kHz to 80 kHz	3.3 %			
						80 kHz to 90 kHz	3.7 %			
						90 kHz to 100 kHz	4.1 %			
		50 H	to	100 H		50 Hz to 100 Hz	0.18 %			
						100 Hz to 500 Hz	0.15 %			
						500 Hz to 800 Hz	0.22 %			
						800 Hz to 1 kHz	0.24 %			
						1 kHz to 2 kHz	0.31 %			
						2 kHz to 3 kHz	0.39 %			
						3 kHz to 5 kHz	0.55 %			
						5 kHz to 7 kHz	0.70 %			
						7 kHz to 10 kHz	0.93 %			
						10 kHz to 15 kHz	1.4 %			
						15 kHz to 20 kHz	1.8 %			
						20 kHz to 25 kHz	2.1 %			
						25 kHz to 30 kHz	2.5 %			
		100 H	to	500 H		20 Hz to 35 Hz	0.38 %			
						35 Hz to 50 Hz	0.40 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						50 Hz to 70 Hz	0.28 %			
						70 Hz to 200 Hz	0.31 %			
						200 Hz to 300 Hz	0.39 %			
						300 Hz to 400 Hz	0.47 %			
						400 Hz to 500 Hz	0.55 %			
						500 Hz to 600 Hz	0.62 %			
						600 Hz to 700 Hz	0.70 %			
						700 Hz to 800 Hz	0.78 %			
						800 Hz to 900 Hz	0.86 %			
						900 Hz to 2 kHz	0.93 %			
						2 kHz to 3 kHz	1.4 %			
						3 kHz to 5 kHz	2.1 %			
						5 kHz to 7 kHz	2.9 %			
						7 kHz to 8 kHz	3.3 %			
						8 kHz to 9 kHz	3.7 %			
						9 kHz to 10 kHz	4.1 %			
		500 H	to	1 kH		20 Hz to 35 Hz	0.38 %			
						35 Hz to 50 Hz	0.40 %			
						50 Hz to 70 Hz	0.28 %			
						70 Hz to 100 Hz	0.31 %			
						100 Hz to 200 Hz	0.31 %			
						200 Hz to 300 Hz	0.39 %			
						300 Hz to 400 Hz	0.47 %			
						400 Hz to 500 Hz	0.55 %			
						500 Hz to 600 Hz	0.62 %			
						600 Hz to 700 Hz	0.70 %			
						700 Hz to 800 Hz	0.78 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						800 Hz to 900 Hz 900 Hz to 1 kHz	0.86 % 0.93 %			
	Inductance / Inductance meters					1 kHz 1 kHz 1 kHz 100 Hz, 1 kHz, 10 kHz 100 Hz, 1 kHz, 10 kHz 100 Hz, 1 kHz, 10 kHz 100 Hz, 1 kHz 100 Hz, 1 kHz 1 kHz 1 kHz	0.20 % 0.14 % 0.03 % 0.02 % 0.03 % 0.03 % 0.02 % 0.05 % 0.2 % 1.5 %	Direct measurement of an inductance standard	TP8, TP24	
10*	Non-linear distortion / Non-linear distortion meters	0.01 %		to	100 %	20 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.003 % 0.03 % 0.05 %	Comparison with a distortion analyzer	TP10	
	Nonlinear distortion / sine signal generators	0.01 %		to	100 %	20 Hz to 20 kHz (BW 80 kHz) 20 kHz to 50 kHz (BW 500 kHz) 50 kHz to 100 kHz (BW 500 kHz)	1 dB + 0.01 % (abs.) 2 dB + 0.04 % (abs.) 2 dB + 0.06 % (abs.)	Measurement by a distortion analyzer	TP10	
11*	HF power / HF power meters and sources and spectrum analyzers, radio	-50 dBm		to	-20 dBm	10 MHz to 30 MHz	7.8 %	Measurement by a wattmeter, 50 Ω	TP11, TP23	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
	communication testers (COM, CMS, CTS,...) and radio navigation testers, simulators and imitators (ATC, DMF, TACAN, VOR – ILS, TCAS), testing and inspection equipment and their parts					30 MHz to 18 GHz	5.4 %			
		-20 dBm	to	+17 dBm		10 MHz to 30 MHz	6.5 %			
						30 MHz to 18 GHz	4.4 %			
		+17 dBm	to	+40 dBm		10 MHz to 18 GHz	5.7 %			
12*	HF voltage, peak-to-peak value / HF voltage meters, oscilloscopes	5 mV	to	3 V		50 kHz to 1 MHz	4.7 %	Direct generation by a calibrator, 50 Ω	TP12, TP23	
						1 MHz to 10 MHz	6.6 %			
						10 MHz to 30 MHz	9.0 %			
						30 MHz to 250 MHz	6.8 %			
		3 V	to	5.5 V		50 kHz to 1 MHz	4.7 %			
						1 MHz to 10 MHz	6.6 %			
					10 MHz to 30 MHz	9.0 %				
					30 MHz to 250 MHz	6.8 %				
					250 MHz to 300 MHz	6.8 %				
	HF voltage, peak-to-peak value / HF voltage sources	20 mV	to	1.5 V		20 MHz to 30 MHz	6.9 %	Measurement by a wattmeter, 50 Ω	TP12	
					30 MHz to 100 MHz	4.7 %				
					100 MHz to 1 GHz	4.3 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		40 mV	to	100 mV		1 MHz to 10 MHz	3.2 %	Measurement by a HF voltmeter, 50 Ω	TP12	
		100 mV	to	250 mV		10 MHz to 20 MHz	3.5 %			
		250 mV	to	1 V		1 MHz to 10 MHz	3.7 %			
		1 V	to	1.5 V		10 MHz to 20 MHz	4.9 %			
		1.5 V	to	3 V		1 MHz to 10 MHz	2.5 %			
						10 MHz to 20 MHz	3.6 %			
13	HF attenuation / HF attenuator	0 dB	to	60 dB		1.2 GHz to 3.8 GHz	0.30 dB	Direct measurement with attenuation meter, comparison method	TP13	
		60 dB	to	80 dB		3.8 GHz to 8.2 GHz	0.40 dB			
						8.2 GHz to 18 GHz	0.70 dB			
						1.2 GHz to 3.8 GHz	0.80 dB			
						3.8 GHz to 8.2 GHz	1.2 dB			
						8.2 GHz to 18 GHz	1.5 dB			
13*	HF attenuation / HF attenuators and attenuation meters, radio communication testers (COM, CMS, CTS,...) and radio navigation testers, simulators and imitators (ATC, DMF, TACAN, VOR – ILS, TCAS), testing and inspection equipment and their parts	0 dB	to	30 dB		10 MHz to 30 MHz	0.50 dB	Wattmeter measurement, power method	TP13	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						30 MHz to 2 GHz 2 GHz to 18 GHz	0.15 dB 0.35 dB			
		30 dB		to	50 dB	10 MHz to 30 MHz 30 MHz to 2 GHz 2 GHz to 18 GHz	0.70 dB 0.35 dB 0.50 dB			
14*	Depth of AM / amplitude modulated signal generators, amplitude modulation meters, radio communication testers (COM, CMS, CTS,...) and radio navigation testers, simulators and imitators (ATC, DMF, TACAN, VOR – ILS, TCAS), testing and inspection equipment and their parts					f_c 150 kHz to 10 MHz	f_{mod} 20 Hz to 50 Hz 50 Hz to 10 kHz	3.2 % rel. 3.2 % rel. 2.2 % rel.	Direct measurement by AM depth meter	TP14
		5 %		to	99 %	10 MHz to 1.3 GHz	20 Hz to 50 Hz 50 Hz to 50 kHz 50 kHz to 100 kHz	3.2 % rel. 1.2 % rel. 3.2 % rel.		
15*	FM frequency deviation / frequency modulated signal generators, frequency modulation meters, radio communication testers (COM, CMS, CTS,...) and					f_c	f_{mod}		Direct measurement with a FM frequency deviation meter	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
	radio navigation testers, simulators and imitators (ATC, DMF, TACAN, VOR – ILS, TCAS), testing and inspection equipment and their parts	90 Hz	to	40 kHz	150 kHz to 10 MHz	20 Hz to 10 kHz	2.1 % + 1 Hz			
		90 Hz	to	400 kHz	10 MHz to 1.3 GHz	20 Hz to 50 Hz	5.1 % + 1 Hz			
					50 Hz to 100 kHz		1.1 % + 1 Hz			
					100 kHz to 200 kHz		5.1 % + 1 Hz			
16*	DC power / DC wattmeters	0.1 mW	to	11 kW	33 mV to 25 V	3.3 mA to 9 mA	0.04 %	Direct generation with a calibrator	TP16	
						9 mA to 33 mA	0.03 %			
						33 mA to 90 mA	0.04 %			
						90 mA to 330 mA	0.03 %			
						0.33 A to 0.9 A	0.08 %			
						0.9 A to 2.2 A	0.06 %			
						2.2 A to 4.5 A	0.12 %			
						4.5 A to 11 A	0.09 %			
					25 V to 100 V	3.3 mA to 9 mA	0.04 %			
						9 mA to 33 mA	0.03 %			
						33 mA to 50 mA	0.04 %			
						50 mA to 100 mA	0.013 %			
						100 mA to 330 mA	0.03 %			
						0.33 A to 0.9 A	0.08 %			
						0.9 A to 2.2 A	0.06 %			
						2.2 A to 4.5 A	0.12 %			
						4.5 A to 11 A	0.09 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						100 V to 1,000 V	3.3 mA to 9 mA	0.04 %		
							9 mA to 33 mA	0.03 %		
						33 mA to 90 mA	0.04 %			
							90 mA to 330 mA	0.03 %		
							0.33 A to 0.9 A	0.08 %		
							0.9 A to 2.2 A	0.06 %		
							2.2 A to 4.5 A	0.12 %		
							4.5 A to 11 A	0.09 %		
		363 mW	to	550 kW		33 mV to 1,000 V	11 A to 550 A	0.50 %	Indirect generation with a current coil calibrator	TP16
	DC power / Sources								Direct measurement with multimeters or indirect measurement with a current shunt	TP16
		1 mW	to	200 kW		1 V to 1,000 V	1 mA to 200 mA	0.009 %		
							200 mA to 200 A	0.045 %		
17*	AC power / AC wattmeters (f: 45 Hz to 65 Hz, PF = 1)	0.1 mW	to	11 kW		33 mV to 330 mV	3.3 mA to 9 mA	0.40 %	Direct generation with a calibrator	TP16
							9 mA to 33 mA	0.25 %		
							33 mA to 90 mA	0.35 %		
							90 mA to 330 mA	0.25 %		
							0.33 A to 0.9 A	0.35 %		
							0.9 A to 2.2 A	0.25 %		
							2.2 A to 4.5 A	0.35 %		
							4.5 A to 11 A	0.25 %		
						330 mV to 1,000 V	3.3 mA to 9 mA	0.25 %		
							9 mA to 33 mA	0.15 %		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						33 mA to 90 mA 90 mA to 330 mA 0.33 A to 0.9 A 0.9 A to 2.2 A 2.2 A to 4.5 A 4.5 A to 11 A	0.25 % 0.15 % 0.25 % 0.15 % 0.20 % 0.15 %			
		11 kW	to	550 kW	33 mV to 1,000 V	11 A to 550 A	0.60 %	Indirect generation with a current coil calibrator	TP16	
	AC power / AC wattmeters (f: 65 Hz to 500 Hz, PF = 1)	1.089 mW	to	11 kW	330 mV to 1,000 V	3.3 mA to 9 mA 9 mA to 33 mA 33 mA to 90 mA 90 mA to 330 mA 0.33 A to 0.9 A 0.9 A to 2.2 A 2.2 A to 4.5 A 4.5 A to 11 A	0.26 % 0.16 % 0.26 % 0.16 % 0.26 % 0.16 % 0.21 % 0.16 %	Direct generation with a calibrator	TP16	
	AC power / AC wattmeters (330 mV to 1,000 V, PF = 0.5 inductive and capacitive)	0.545 mW	to	5.5 kW	45 Hz to 65 Hz	3.3 mA to 9 mA 9 mA to 33 mA 33 mA to 90 mA 90 mA to 330 mA 0.33 A to 0.9 A 0.9 A to 2.2 A	0.55 % 0.50 % 0.55 % 0.50 % 0.55 % 0.50 %	Direct generation with a calibrator	TP16	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						2.2 A to 4.5 A	0.55 %			
						4.5 A to 11 A	0.50 %			
						65 Hz to 500 Hz	2.8 %			
18*	DC electrical work / DC electrical work meters (t: 600 s to 24 h)	66 mWs	to	950.4 MWs		33 mV to 25 V	3.3 mA to 9 mA	0.055 %	Direct generation with a calibrator	TP16
							9 mA to 33 mA	0.050 %		
							33 mA to 90 mA	0.055 %		
							90 mA to 330 mA	0.050 %		
							0.33 A to 0.9 A	0.090 %		
							0.9 A to 2.2 A	0.070 %		
							2.2 A to 4.5 A	0.13 %		
							4.5 A to 11 A	0.10 %		
					25 V to 100 V	3.3 mA to 9 mA	9 mA to 33 mA	0.055 %		
							33 mA to 50 mA	0.050 %		
							50 mA to 100 mA	0.055 %		
							100 mA to 330 mA	0.040 %		
							0.33 A to 0.9 A	0.050 %		
							0.9 A to 2.2 A	0.090 %		
							2.2 A to 4.5 A	0.070 %		
							4.5 A to 11 A	0.13 %		
							4.5 A to 11 A	0.10 %		
					100 V to 1,000 V	3.3 mA to 9 mA	9 mA to 33 mA	0.055 %		
							33 mA to 90 mA	0.050 %		
							90 mA to 330 mA	0.055 %		
							0.33 A to 0.9 A	0.050 %		
							0.33 A to 0.9 A	0.090 %		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						0.9 A to 2.2 A 2.2 A to 4.5 A 4.5 A to 11 A	0.070 % 0.13 % 0.10 %			
		217.8 Ws	to	47.52 GWs	33 mV to 1,000 V	11 A to 550 A	0.55 %	Indirect generation with a current coil calibrator	TP16	
	DC electrical work / Special electrical work meters with current inputs (t: 600 s to 24 h)	0 Ws	to	2.4 GWh	I ₁ : 0 µA to 2 A	I ₂ : 10 µA to 2 A	0.1 % + 1 Ws	Direct generation with calibrators	TP16	
19*	AC electrical work / AC electrical work meters (f: 45 Hz to 65 Hz, PF = 1, t: 600 s to 24 h)	0.66 Ws	to	950.4 MWs	330 mV to 1,000 V	3.3 mA to 9 mA 9 mA to 33 mA 33 mA to 90 mA 90 mA to 330 mA 0.33 A to 0.9 A 0.9 A to 2.2 A 2.2 A to 4.5 A 4.5 A to 11 A	0.28 % 0.18 % 0.28 % 0.18 % 0.28 % 0.18 % 0.23 % 0.18 %	Direct generation with a calibrator	TP16	
		2.178 kW	to	47.52 GWs	330 mV to 1,000 V	11 A to 550 A	0.63 %	Indirect generation with a current coil calibrator	TP16	
	AC electrical work / AC electrical work meters (f: 65 Hz to 500 Hz, PF = 1, t: 600 s to 24 h)	0.66 Ws	to	950.4 MWs	330 mV to 1,000 V	3.3 mA to 9 mA 9 mA to 33 mA	0.29 % 0.19 %	Direct generation with a calibrator	TP16	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
						33 mA to 90 mA 90 mA to 330 mA 0.33 A to 0.9 A 0.9 A to 2.2 A 2.2 A to 4.5 A 4.5 A to 11 A	0.29 % 0.19 % 0.29 % 0.19 % 0.24 % 0.19 %			
	AC electrical work / AC electrical work meters (f: 45 Hz to 65 Hz, PF = 0.5 inductive and capacitive t: 600 s to 24 h)	0.33 Ws		to	475.2 MWs	330 mV to 1,000 V 3.3 mA to 9 mA 9 mA to 33 mA 33 mA to 90 mA 90 mA to 330 mA 0.33 A to 0.9 A 0.9 A to 2.2 A 2.2 A to 4.5 A 4.5 A to 11 A	0.58 % 0.53 % 0.58 % 0.53 % 0.58 % 0.53 % 0.58 % 0.53 %			
	AC electrical work / AC electrical work meters (f: 65 Hz to 500 Hz, PF = 0.5 inductive and capacitive t: 600 s to 24 h)	0.33 Ws		to	475.2 MWs	330 mV to 1,000 V 3.3 mA to 11 A	2.9 %			
20*	Reflection coefficient / Measurement of impedance matching at 50 Ω impedance	0.00		to	0.10	10 MHz to 2 GHz	0.020	Measurement with a directional bridge, N connector, 50 Ω	TP17	
		0.10		to	0.20	10 MHz to 2 GHz	0.030			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place	
		min	unit	max	unit						
		0.20	to	0.30		10 MHz to 2 GHz	0.045				
		0.00	to	0.15		2 GHz to 18 GHz	0.070	Measurement with a directional coupling, N connector, 50 Ω	TP17		
		0.15	to	0.20		2 GHz to 18 GHz	0.080				
		0.20	to	0.30		2 GHz to 18 GHz	0.10				
21*	Voltages above 1,000 V / DC and AC high voltage sources and surge generators - peak value	1 kV	to	3 kV		0 Hz	0.21 %	Measurement with a multimeter with a resistance divider	TP18		
		3 kV	to	90 kV		0 Hz	0.20 %	High voltage probe measurement	TP18		
		90 kV	to	100 kV		0 Hz	0.25 %				
		1 kV	to	1.5 kV		up to 75 MHz	3 dB	Measurement using an oscilloscope with a HV probe	TP18		
		1 kV	to	3 kV		up to 10 MHz	3 dB				
1 kV	to	8 kV		up to 1 MHz	3 dB						
1 kV	to	10 kV		up to 500 kHz	3 dB						
1 kV	to	14 kV		rise time >10 ns	3 dB						
		1 kV	to	25 kV		50 Hz	0.3 %	Measurement using a measuring transformer	TP18		
	Voltages above 1,000 V / DC and AC high voltage meters, HV voltage/current transducers	1 kV	to	4 kV		50 Hz	0.5 %	Generation using generator and measuring transformer	TP18		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
		1 kV	to	30 kV		0 Hz	0.25 %	Generation using a generator and resistor divider	TP18	
22	Mains impedance / Instruments for the inspection of electrical installations and resistance-based impedance standards			0.16 Ω		50 Hz	0.006 Ω	Generation using a reference socket	TP6, TP27	
		0.17 Ω	to	1 Ω		50 Hz	0.5 % + 0.006 Ω			
		1 Ω	to	10 Ω		50 Hz	0.3 % + 0.006 Ω			
		10 Ω	to	2 kΩ		50 Hz	0.1 % + 0.006 Ω			
23*	Phase shift / Phase shifted voltage signal sources			360°		U ₁ = U ₂ , where U ₁ : 10 mV to 50 V U ₂ : 10 mV to 50 V or U ₁ : 10 mV to 30 V U ₂ : 1 V to 250 V	1°	Measurement by a phase shift meter	TP32	
	Phase shift / Phase shifted voltage signal meters			360°		U ₁ : 10 mV to 3 V U ₂ : 10 mV to 3 V	1°	Direct generation with a calibrator		
24*	Oscilloscope vertical deflection coefficient			55 V		1 kHz	0.05 % + 20 μV	Direct measurement of commutated voltage with a multimeter	TP2, TP23	
		12 mV	to							
		-300 V	to	300 V		0 Hz	0.005 % + 10 μV	Direct generation with a calibrator	TP2, TP23	
		12 mV	to	55 V		1 kHz	0.30 % + 100 μV			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

- ² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.
- ³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

CMC for the field of measured quantity: Time and frequency quantities

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
1*	Frequency / LF and HF counters, LF and HF generators, variable speed drives, frequency standards, frequency comparators and pulse generators	0.005 Hz	to	400 kHz			$3.3 \times 10^{-4} / f$ (note 4)	Comparison with a standard, measurement	TP9, TP22	
		400 kHz	to	1.5 GHz			1.2×10^{-9}			
		1.5 GHz	to	18 GHz			3×10^{-9}			
				1 MHz		τ : 1 s to 2,000 s τ : 2,000 s to 15,000 s τ : over 15,000 s	3×10^{-10} 1×10^{-11} 5×10^{-12}			
				5 MHz		τ : 1 s to 2,000 s τ : 2,000 s to 15,000 s τ : over 15,000 s	3×10^{-10} 1×10^{-11} 5×10^{-12}			
				10 MHz		τ : 1 s to 2,000 s τ : 2,000 s to 15,000 s τ : over 15,000 s	3×10^{-10} 1×10^{-11} 5×10^{-12}			
		1 Hz	to	18 GHz			1×10^{-11} 1×10^{-9}	Generation of reference signal	TP9, TP22	
2*	Time stamps / Oscilloscopes, transient recorders	1.8 ns	to	2.2 ns			0.005 %	Direct generation by a calibrator, generator	TP23	
		4.5 ns	to	11 ns			0.005 %			
		18 ns	to	22 ns			0.005 %			
		45 ns	to	60 s			0.005 %			
				1 s			1×10^{-9}			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.

CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
3*	Rise time / Oscilloscopes, transient recorders	0.1 ns				50 mV, 500 mV, 1 V	0.1 ns	Direct generation by a calibrator, generator	TP23	
		0.70 ns		to 0.80 ns		4.5 mV to 5.5 mV	0.3 ns			
						9 mV to 11 mV	0.2 ns			
						45 mV to 55 mV	0.2 ns			
				90 mV to 110 mV	0.2 ns					
				450 mV to 550 mV	0.2 ns					
				0.9 V to 1.1 V	0.2 ns					
		0.80 ns		to 0.90 ns		22.5 mV to 27.5 mV	0.2 ns			
						225 mV to 275 mV	0.2 ns			
						2.25 V to 2.75 V	0.2 ns			
4*	Time Interval / Stopwatches, timers, time meters and time interval sources	0.1 s		to	10 ⁵ s	Electronically switched	0.004 %	Direct measurement of time by a counter, stopwatch	TP33	
		1 s		to	90,000 s	Manually switched	0.20 s			
		100 ps		to	60 s		0.2 % + 50 ps	Direct measurement of time intervals with an oscilloscope	TP33	

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

⁴ Measured frequency in Hz

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

HES, s.r.o.
CAB number 2273, Calibration Laboratory
U Dráhy 411/11, 664 49 Ostopovice

CMC for the field of measured quantity: Physicochemical quantities

Ord. number ₁	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work place
		min	unit	max	unit					
1*	Humidity / Analog and digital hygrometers, humidity transducers and humidity measuring chains, including humidity probes	5 % RH	to	10 % RH	(10 to 50) °C	2.2 %	Comparison with a standard hygrometer in a conditioning chamber	TP45		
		10 % RH	to	50 % RH	(10 to 50) °C	1.4 %				
		50 % RH	to	70 % RH	(10 to 50) °C	1.6 %				
		70 % RH	to	90 % RH	(10 to 50) °C	1.8 %				
2*	pH / Electrical parts of pH meters and pH simulators	0 pH	to	14 pH		0.01 pH	Direct generation with a calibrator	TP1, TP21		

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).